



Notice of Simulator Solution File Review

At the request of the State Toxicologist a review of the following simulator solution records has been accomplished. The following file consists of simulator solution analyses performed and completed by the State Toxicology Laboratory for a specific batch number. The file contains the simulator solution data entry form along with a file review record and the chromatograms generated by the Toxicology Laboratory during the analyses of the solutions. This file has been reviewed by Tpr. Ken Denton and Mr. Rod Gullberg for accuracy and completeness. Where computations regarding simulator solution values have been found to be incorrect, the corrected values have been written in by Mr. Rod Gullberg along with initials and date. The corrected values were then evaluated to ensure that the solution still conformed to those standards established by the State Toxicologist.

Where computation values changed for a specific batch number, the analysts employed by the State Toxicology Laboratory were asked to review the revisions, ensure the solution complied with the criteria established by the State Toxicologist and then re-sign their affidavit. Their signature will appear on their original affidavit along with a statement regarding their review of the results.

Where a dating error occurred that analyst will have made the correction on the original data form including their initials and date and then re-signed their original affidavit.

Tpr. Ken Denton

Date

Rod G. Gullberg

Date

Washington State Toxicology Laboratory

Simulator Solution Data Entry Review Form

Reviewer KEN DENTON / ROD GULLBERG Date 10-1-07
Location TOX LAB SEATTLE Batch Number 06045

Form Review Criteria

Preparation date precedes all analysis dates: Okay Not Okay ___
Data entry corresponds to all chromatograms: Okay Not Okay ___
All signatures present: Okay Not Okay ___



Computations:

Avg. solution concentration: Correct Not Correct ___
Standard deviation: Correct Not Correct ___
Range: Correct Not Correct ___
Precision: Correct Not Correct ___
Equivalent vapor concent.: Correct Not Correct ___
External Control Information
(lot # and future date): Correct Not Correct ___

Complies with accuracy and precision requirements established by the
State Toxicologist: Yes No ___

Corrections Necessary:

Comments:

Reviewer Signature:  Date: 10-1-07
Reviewer Signature:  Date: 10/1/2007

WASHINGTON STATE TOXICOLOGY LABORATORY
FORENSIC LABORATORY SERVICES BUREAU
 WASHINGTON STATE PATROL
 2203 AIRPORT WAY S, SUITE 360
 SEATTLE, WASHINGTON 98134-2027
 (206) 262-6100 FAX (206) 262-6145




Preparation and certification of **0.08** g/210L Quality Assurance solution
 Batch number **06045** Date: 11/14/2006
 Preparation: 22.2 mL of absolute ethyl alcohol diluted to 18 Liters with water
 Concentration of ethanol (g/100mL) measured by gas chromatography:

	Anal 1	Anal 2	Anal 3	Anal 4	Anal 5	Anal 6	Anal 7	Anal 8	Anal 9	Anal 10	Anal 11	Anal 12	Anal 13	Anal 14	Anal 15	Anal 16
1	0.097	0.098	0.097													
2	0.098	0.099	0.096													
3	0.097	0.099	0.097													
4	0.099	0.099	0.097													
5	0.100	0.099	0.098													
Ctrl	0.101	0.099	0.099													

External Control:
 Lot #: A041837 Exp date: 04/2010
 Target concentration: 0.10 g/100mL

Statistics:
 Avg. solution concent.: 0.0980 g/100 mL
 SD: 0.00113
 Range (3xSD): 0.0946 to 0.1014
 Precision CV (%): 1.1570 %

Equivalent vapor concent.: 0.0797 g/210L

Analyst	Name	Signature	Date
1	Paige Long		11/14/2006
2	Naziha Nuwayhid, PhD		11/15/2006
3	Edward Formoso		11/15/2006
4			
5			
6			
7			
8			
9			
10			
11			
12			
13			
14			
15			
16			

Prepared by: Paige Long according to the approved protocol



STATE OF WASHINGTON
WASHINGTON STATE PATROL

WASHINGTON STATE TOXICOLOGY LABORATORY

2203 Airport Way South, Suite 360•Seattle, Washington 98134-2927•(206) 262-6100•FAX (206) 262-6145

DATAMASTER QUALITY ASSURANCE SOLUTION
CERTIFICATION


I, Naziha Nuwayhid, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: Bachelor and Masters degrees in Biology, Ph.D. degree in Basic Medical Science, ten years experience in clinical laboratory sciences, one year in clinical toxicology and six years in forensic toxicology. I am also board certified by the American Board of Clinical Chemistry.



The quality assurance solution, Lot Number 06045, was prepared in the Washington State Toxicology Laboratory on 11/14/2006. I examined and tested this solution. The mean concentration of the alcohol was 0.08 grams per 100ml.

Dated: 11/20/2006
Seattle, WA


Naziha Nuwayhid, Ph.D.
Forensic Toxicologist

NN/km
NNQA

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions.



STATE OF WASHINGTON

WASHINGTON STATE PATROL

WASHINGTON STATE TOXICOLOGY LABORATORY

2203 Airport Way South, Suite 360•Seattle, Washington 98134-2927•(206) 262-6100•FAX (206) 262-6145

DATAMASTER QUALITY ASSURANCE SOLUTION
CERTIFICATION


I, E. Paige Long, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Biology, and MS degree in Forensic Science.

The quality assurance solution, Lot Number 06045, was prepared in the Washington State Toxicology Laboratory on 11/14/2006. I examined and tested this solution. The mean concentration of the alcohol was 0.08 grams per 100ml.

Dated: 11/20/2006
Seattle, WA


E. Paige Long
Forensic Toxicologist

EPL/km
PLQA



STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY

2203 Airport Way South, Suite 360•Seattle, Washington 98134-2927•(206) 262-6100•FAX (206) 262-6145

DATAMASTER QUALITY ASSURANCE SOLUTION
CERTIFICATION

I, Edward J. Formoso, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Chemistry and twenty-nine years of experience in the Washington State Toxicology Laboratory.

The quality assurance solution, Lot Number 06045, was prepared in the Washington State Toxicology Laboratory on 11/14/2006. I examined and tested this solution. The mean concentration of the alcohol was 0.08 grams per 100ml.

Dated: 11/20/2006
Seattle, WA


Edward J. Formoso
Forensic Toxicologist

EJF/km
EFQA

D:\HPCHEM\1\METHODS\BLDALCO2.M
 11/14/2006 2:20:34 PM
 Instrument 5
 DB-ALC2


0:10ctl
 p long

vial # 8




#	Compound	Area	RT
1	ethanol	861	1.110
2	n-propanol	1805	1.950

Totals:



ethanol 0.101 g/100ml




n-propanol 1.000 g/100ml

D:\HPCHEM\1\METHODS\BLDALCO2.M
 11/14/2006 2:24:08 PM
 Instrument 5
 DB-ALC2


blank
 p long

vial # 9




#	Compound	Area	RT
1	ethanol	0	0.000
2	n-propanol	1699	1.949

Totals:



ethanol 0.000 g/100ml




n-propanol 1.000 g/100ml

D:\HPCHEM\1\METHODS\BLDALCO2.M
 11/14/2006 2:29:01 PM
 Instrument 5
 DB-ALC2


QA06045-1
 p long

vial # 10




#	Compound	Area	RT
1	ethanol	757	1.110
2	n-propanol	1643	1.949

Totals:



ethanol 0.097 g/100ml




n-propanol 1.000 g/100ml

D:\HPCHEM\1\METHODS\BLDALCO2.M
 11/14/2006 2:32:24 PM
 Instrument 5
 DB-ALC2


QA06045-2
 p long

vial # 11




#	Compound	Area	RT
1	ethanol	754	1.110
2	n-propanol	1630	1.949

Totals:



ethanol 0.098 g/100ml




n-propanol 1.000 g/100ml

D:\HPCHEM\1\METHODS\BLDALCO2.M
 11/14/2006 2:36:03 PM
 Instrument 5
 DB-ALC2


QA06045-3
 p long

vial # 12




#	Compound	Area	RT
1	ethanol	749	1.110
2	n-propanol	1625	1.948

Totals:



ethanol 0.097 g/100ml




n-propanol 1.000 g/100ml

D:\HPCHEM\1\METHODS\BLDALCO2.M
 11/14/2006 2:40:59 PM
 Instrument 5
 DB-ALC2


QA06045-4
 p long

vial # 13




#	Compound	Area	RT
1	ethanol	845	1.111
2	n-propanol	1802	1.950

Totals:



ethanol 0.099 g/100ml




n-propanol 1.000 g/100ml

D:\HPCHEM\1\METHODS\BLDALCO2.M
 11/14/2006 2:44:21 PM
 Instrument 5
 DB-ALC2


QA06045-5
 p long

vial # 14




#	Compound	Area	RT
1	ethanol	836	1.110
2	n-propanol	1775	1.950

Totals:



ethanol 0.100 g/100ml




n-propanol 1.000 g/100ml

C:\HPCHEM\1\METHODS\BLDALCO.M
 11/15/2006 2:38:07 PM
 Instrument 1
 DB ALC 1


06045 QA-1
 N Nuwayhid, PhD

vial # 34




#	Compound	Area	RT
1	Ethanol	1263	1.076
2	n-Propanol	2734	1.769

Tot



Ethanol

0.098 g/100ml




n-Propanol

1.000 g/100ml

C:\HPCHEM\1\METHODS\BLDALCO.M
 11/15/2006 2:41:11 PM
 Instrument 1
 DB ALC 1


06045 QA-2
 N Nuwayhid, PhD

vial # 35




#	Compound	Area	RT
1	Ethanol	1245	1.076
2	n-Propanol	2683	1.769

Tot



Ethanol

0.099 g/100ml




n-Propanol

1.000 g/100ml

C:\HPCHEM\1\METHODS\BLDALCO.M
 11/15/2006 2:44:16 PM
 Instrument 1
 DB ALC 1


06045 QA-3
 N Nuwayhid, PhD

vial # 36




#	Compound	Area	RT
1	Ethanol	1234	1.075
2	n-Propanol	2665	1.768

Tot



Ethanol

0.099 g/100ml




n-Propanol

1.000 g/100ml

C:\HPCHEM\1\METHODS\BLDALCO.M
 11/15/2006 2:47:21 PM
 Instrument 1
 DB ALC 1


06045 QA-4
 N Nuwayhid, PhD

vial # 37




#	Compound	Area	RT
1	Ethanol	1227	1.076
2	n-Propanol	2642	1.769

Tot



Ethanol

0.099 g/100ml




n-Propanol

1.000 g/100ml

C:\HPCHEM\1\METHODS\BLDALCO.M
 11/15/2006 2:50:26 PM
 Instrument 1
 DB ALC 1


06045 QA-5
 N Nuwayhid, PhD

vial # 38




#	Compound	Area	RT
1	Ethanol	1244	1.076
2	n-Propanol	2688	1.769

Tot



Ethanol

0.099 g/100ml



n-Propanol


1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
 11/15/2006 2:53:31 PM
 Instrument 1
 DB ALC 1


0.10 Ctrl-NN
 N Nuwayhid, PhD

vial # 39




#	Compound	Area	RT
1	Ethanol	1271	1.076
2	n-Propanol	2726	1.769

Tot



Ethanol

0.099 g/100ml




n-Propanol

1.000 g/100ml


C:\HPCHEM\1\METHODS\BLDALCO.M
 11/15/2006 2:56:35 PM
 Instrument 1
 DB ALC 1

Blank
 N Nuwayhid, PhD
 vial # 40




#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2743	1.770

Tot



Ethanol 0.000 g/100ml




n-Propanol 1.000 g/100ml

D:\HPCHEM\1\METHODS\BLDALCO.M
 11/15/2006 5:06:37 PM
 Instrument 4
 DB-ALC1


06045
 ED FORMOSO

vial # 13




#	Compound	Area	RT
1	Ethanol	615	1.007
2	n-Propanol	1414	1.662

Totals:



Ethanol 0.097 g/100ml




n-Propanol 1.000 g/100ml

D:\HPCHEM\1\METHODS\BLDALCO.M
 11/15/2006 5:09:52 PM
 Instrument 4
 DB-ALC1


06045
 ED FORMOSO

vial # 14




#	Compound	Area	RT
1	Ethanol	674	1.008
2	n-Propanol	1566	1.663

Totals:



Ethanol 0.096 g/100ml




n-Propanol 1.000 g/100ml

D:\HPCHEM\1\METHODS\BLDALCO.M
 11/15/2006 5:13:06 PM
 Instrument 4
 DB-ALC1


06045
 ED FORMOSO

vial # 15




#	Compound	Area	RT
1	Ethanol	600	1.007
2	n-Propanol	1380	1.662

Totals:



Ethanol 0.097 g/100ml




n-Propanol 1.000 g/100ml

D:\HPCHEM\1\METHODS\BLDALCO.M
 11/15/2006 5:16:19 PM
 Instrument 4
 DB-ALC1


06045
 ED FORMOSO

vial # 16




#	Compound	Area	RT
1	Ethanol	607	1.007
2	n-Propanol	1396	1.662

Totals:



Ethanol 0.097 g/100ml




n-Propanol 1.000 g/100ml

D:\HPCHEM\1\METHODS\BLDALCO.M
 11/15/2006 5:19:39 PM
 Instrument 4
 DB-ALC1


06045
 ED FORMOSO

vial # 17




#	Compound	Area	RT
1	Ethanol	623	1.007
2	n-Propanol	1420	1.662

Totals:



Ethanol 0.098 g/100ml




n-Propanol 1.000 g/100ml

D:\HPCHEM\1\METHODS\BLDALCO.M
 11/15/2006 5:22:48 PM
 Instrument 4
 DB-ALC1


0.10 CONTROL
 ED FORMOSO

vial # 18




#	Compound	Area	RT
1	Ethanol	620	1.007
2	n-Propanol	1401	1.662

Totals:



Ethanol 0.099 g/100ml




n-Propanol 1.000 g/100ml

D:\HPCHEM\1\METHODS\BLDALCO.M
 11/15/2006 5:26:07 PM
 Instrument 4
 DB-ALC1


BLANK
 ED FORMOSO

vial # 19




#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1416	1.662

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml