Notice of Simulator Solution File Review

At the request of the State Toxicologist a review of the following simulator solution records has been accomplished. The following file consists of simulator solution analyses performed and completed by the State Toxicology Laboratory for a specific batch number. The file contains the simulator solution data entry form along with a file review record and the chromatograms generated by the Toxicology Laboratory during the analyses of the solutions. This file has been reviewed by Tpr. Ken Denton and Mr. Rod Gullberg for accuracy and completeness. Where computations regarding simulator solution values have been found to be incorrect, the corrected values have been written in by Mr. Rod Gullberg along with initials and date. The corrected values were then evaluated to ensure that the solution still conformed to those standards established by the State Toxicologist.

Where computation values changed for a specific batch number, the analysts employed by the State Toxicology Laboratory were asked to review the revisions, ensure the solution complied with the criteria established by the State Toxicologist and then re-sign their affidavit. Their signature will appear on their original affidavit along with a statement regarding their review of the results.

Where a dating error occurred that analyst will have made the correction on the original data form including their initials and date and then re-signed their original affidavit.

K.L.D. 10/1/2007

Tpr. Ken Denton Date

R.G. Gullberg 10-1-07

Rod G. Gullberg Date

Washington State Toxicology Laboratory

Simulator Solution Data Entry Review Form

Reviewer KEN BENTON / ROB GULLBERG Date 9-27-07

Location TOX LAB SEATTLE Batch Number 07015

Form Review Criteria

Preparation date precedes all analysis dates: Okay X Not Okay ___

Data entry corresponds to all chromatograms: Okay X Not Okay ___

All signatures present: Okay X Not Okay ___

Computations:

Avg. solution concentration: Correct ___ Not Correct X

Standard deviation: Correct ___ Not Correct X

Range: Correct ___ Not Correct X

Precision: Correct ___ Not Correct X RAB 9-27-07

Equivalent vapor concent.: Correct ~~X~~ Not Correct X

External Control Information
(lot # and future date): Correct X Not Correct ___

Complies with accuracy and precision requirements established by the State Toxicologist: Yes X No ___

Corrections Necessary:

Comments:

Reviewer Signature: RB Gullberg Date: 9-27-07

Reviewer Signature: K.L. Benton Date: 9-27-~~02~~07

9/26/2007

WASHINGTON STATE TOXICOLOGY LABORATORY
FORENSIC LABORATORY SERVICES BUREAU
WASHINGTON STATE PATROL
2203 AIRPORT WAY S, SUITE 360
SEATTLE, WASHINGTON 98134-2027
(206) 262-6100 FAX (206) 262-6145

Preparation and certification of **0.08** g/210L Simulator External Standard solution

Batch number **07015** Date: 5/8/2007

Preparation: 66.5 mL of absolute ethyl alcohol diluted to 52 Liters with water

Concentration of ethanol (g/100mL) measured by gas chromatography:

	Anal 1	Anal 2	Anal 3	Anal 4	Anal 5	Anal 6	Anal 7	Anal 8	Anal 9	Anal10	Anal 11	Anal 12	Anal 13	Anal14	Anal 15	Anal 16
1	0.102	0.100	0.100	0.101	0.100	0.102	0.100	0.101	0.100	0.102	0.099	0.101	0.102	0.101	0.101	
2	0.101	0.102	0.101	0.102	0.100	0.102	0.102	0.101	0.100	0.101	0.099	0.101	0.102	0.101	0.101	
3	0.102	0.101	0.100	0.101	0.101	0.102	0.100	0.101	0.100	0.102	0.100	0.101	0.102	0.102	0.101	
4	0.101	0.100	0.102	0.101	0.101	0.101	0.101	0.101	0.100	0.102	0.099	0.101	0.102	0.101	0.101	
5	0.102	0.100	0.101	0.101	0.101	0.102	0.102	0.101	0.100	0.102	0.099	0.102	0.102	0.101	0.102	
Ctrl	0.103	0.099	0.101	0.100	0.100	0.102	0.102	0.098	0.097	0.100	0.098	0.100	0.099	0.098	0.100	

External Control:
Lot #: A048730 Exp date: 3/2011
Target concentration: 0.10 g/100mL

Equivalent vapor concent.: ~~0.0820~~ 0.0821 g/210L

Statistics:
Avg. solution concent.: ~~0.1009~~ 0.1010 g/100 mL
SD: ~~0.00090~~ 0.00086 (0.0984)
Range (3xSD): ~~0.0982~~ to 0.1036
Precision CV (%): ~~0.8959~~ 0.8535 %

Analyst	Name	Signature	Date
1	Estuardo J. Miranda		05/10/2007
2	Justin L Knoy		05/08/2007
3	Sarah M Swenson		05/09/2007
4	Brian Capron		05/10/2007
5	Brianne Akins		05/14/2007
6	Kelly Gross		05/14/2007
7	Melissa Pemberton		05/15/2007
8	Rebecca Flaherty		05/16/2007
9	Edward Formoso		05/17/2007
10	Brittany Ball		05/18/2007
11	Lisa Piquette		05/18/2007
12	Christopher S Johnston		05/22/2007
13	Naziha Nuwayhid, PhD		05/22/2007
14	Asa Louis		05/15/2007
15	Amanda Black		05/23/2007
16			

Prepared by: Estuardo J. Miranda according to the approved protocol

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Estuardo J. Miranda, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: Bachelor of Science in Chemistry, Master of Science in Zoology and nine years experience in Forensic Toxicology.

The simulator solution, Lot Number 07015, was prepared in the Washington State Toxicology Laboratory on 5/8/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/8/2008.

Dated: 5/29/2007
Seattle, WA

Estuardo J. Miranda
Forensic Toxicologist

EM/jr
EMSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions. 9-20-07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Justin L. Knoy, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Biology, and MS degree in Forensic Science.

The simulator solution, Lot Number 07015, was prepared in the Washington State Toxicology Laboratory on 5/8/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/8/2008.

Dated: 5/29/2007
Seattle, WA

Justin L. Knoy
Forensic Toxicologist

JLK/jr
JKSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions.

10/1/07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Sarah M Swenson, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Chemistry and over three years of experience in forensic toxicology.

The simulator solution, Lot Number 07015, was prepared in the Washington State Toxicology Laboratory on 5/8/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/8/2008.

Dated: 5/29/2007
Seattle, WA

Sarah M Swenson
Forensic Toxicologist

SS/jr
SSSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions.

10/1/07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Brian Capron, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Biology and nine years of experience in forensic toxicology.

The simulator solution, Lot Number 07015, was prepared in the Washington State Toxicology Laboratory on 5/8/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/8/2008.

Dated: 5/29/2007
Seattle, WA

Brian Capron
Brian Capron
Forensic Toxicologist

BC/jr
BCSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions.

Brian Capron 10.1.07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Brianne E. Akins, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Biology.

The simulator solution, Lot Number 07015, was prepared in the Washington State Toxicology Laboratory on 5/8/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/8/2008.

Dated: 5/29/2007
Seattle, WA

Brianne E. Akins

Brianne E. Akins
Forensic Toxicologist

BEA/jr
BLSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions.

Brianne E. Akins 10·1·07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Kelly D. Gross, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: B.S. degree in Chemistry and fifteen years of forensic laboratory experience.

The simulator solution, Lot Number 07015, was prepared in the Washington State Toxicology Laboratory on 5/8/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/8/2008.

Dated: 5/29/2007
Seattle, WA

Kelly D. Gross
Forensic Toxicologist

KDG/jr
KDGSIMSOL

~~A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions.~~

RDG 10-1-07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Melissa L. Pemberton, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: Bachelors degree in Microbiology and over fifteen years of experience as a forensic toxicologist.

The simulator solution, Lot Number 07015, was prepared in the Washington State Toxicology Laboratory on 5/8/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/8/2008.

Dated: 5/29/2007
Seattle, WA

Melissa L. Pemberton
Forensic Toxicologist

MP/jr
MPSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions.

10-1-07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Rebecca Flaherty, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degrees in Biochemistry and Psychobiology and MS degree in Forensic Science.

The simulator solution, Lot Number 07015, was prepared in the Washington State Toxicology Laboratory on 5/8/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/8/2008.

Dated: 5/29/2007
Seattle, WA

Rebecca Flaherty
Forensic Toxicologist

RF/jr
RFSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions. 09/28/07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Edward J. Formoso, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Chemistry and twenty-nine years of experience in the Washington State Toxicology Laboratory.

The simulator solution, Lot Number 07015, was prepared in the Washington State Toxicology Laboratory on 5/8/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/8/2008.

Dated: 5/29/2007
Seattle, WA

Edward J. Formoso
Forensic Toxicologist

EJF/jr
EFSIMSOL

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Brittany Ball, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Biology

The simulator solution, Lot Number 07015, was prepared in the Washington State Toxicology Laboratory on 5/8/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/8/2008.

Dated: 5/29/2007
Seattle, WA

Brittany Ball
Brittany Ball
Forensic Toxicologist

BB/jr
BBSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions. Brittany Ball 10/1/07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Lisa Piquette, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Biochemistry, and two years laboratory experience in formulation chemistry.

The simulator solution, Lot Number 07015, was prepared in the Washington State Toxicology Laboratory on 5/8/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/8/2008.

Dated: 5/29/2007
Seattle, WA

Lisa Piquette
Forensic Toxicologist

LRP/jr
LPSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions. Lisa Noble 9/28/07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Christopher S. Johnston, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Biochemistry.

The simulator solution, Lot Number 07015, was prepared in the Washington State Toxicology Laboratory on 5/8/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/8/2008.

Dated: 5/29/2007
Seattle, WA

Christopher S. Johnston
Forensic Toxicologist

CSJ/jr
CJSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions. 9.28.07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Naziha Nuwayhid, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: Bachelor and Masters degrees in Biology, Ph.D. degree in Basic Medical Science, ten years experience in clinical laboratory sciences, one year in clinical toxicology and six years in forensic toxicology. I am also board certified by the American Board of Clinical Chemistry.

The simulator solution, Lot Number 07015, was prepared in the Washington State Toxicology Laboratory on 5/8/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/8/2008.

Dated: 5/29/2007
Seattle, WA

Naziha Nuwayhid, Ph.D.
Forensic Toxicologist

NN/jr
NNSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions.

9-28-07
NN

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Asa J. Louis, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: B.S. degree in Biochemistry and eight years of toxicology experience.

The simulator solution, Lot Number 07015, was prepared in the Washington State Toxicology Laboratory on 5/8/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/8/2008.

Dated: 5/29/2007
Seattle, WA

Asa J. Louis
Forensic Toxicologist

AJL/jr
AJLSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions.

2007 OCT 01

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Amanda M. Black, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degrees in Chemistry and Veterinary Science.

The simulator solution, Lot Number 07015, was prepared in the Washington State Toxicology Laboratory on 5/8/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/8/2008.

Dated: 5/29/2007
Seattle, WA

Amanda M. Black
Forensic Toxicologist

AB/jr
ABSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions. 9-28-07

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/10/2007 5:40:08 PM
Instrument 1
DB ALC 1

Sim Sol 07015-1
Estuardo J. Miranda

vial # 33



#	Compound	Area	RT
1	Ethanol	1311	1.076
2	n-Propanol	2814	1.767
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/10/2007 5:43:12 PM
Instrument 1
DB ALC 1

Sim Sol 07015-2
Estuardo J. Miranda

vial # 34



#	Compound	Area	RT
1	Ethanol	1308	1.077
2	n-Propanol	2817	1.768
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/10/2007 5:46:17 PM
Instrument 1
DB ALC 1

Sim Sol 07015-3
Estuardo J. Miranda

vial # 35



#	Compound	Area	RT
1	Ethanol	1323	1.077
2	n-Propanol	2830	1.768
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/10/2007 5:49:22 PM
Instrument 1
DB ALC 1

Sim Sol 07015-4
Estuardo J. Miranda

vial # 36



#	Compound	Area	RT
1	Ethanol	1315	1.077
2	n-Propanol	2841	1.767
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/10/2007 5:52:27 PM
Instrument 1
DB ALC 1

Sim Sol 07015-5
Estuardo J. Miranda

vial # 37



#	Compound	Area	RT
1	Ethanol	1334	1.076
2	n-Propanol	2855	1.768
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/10/2007 5:55:32 PM
Instrument 1
DB ALC 1

0.100 Control EM
Estuardo J. Miranda

vial # 38



#	Compound	Area	RT
1	Ethanol	1439	1.076
2	n-Propanol	3031	1.780
Tot			



Ethanol 0.103 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/10/2007 5:58:36 PM
Instrument 1
DB ALC 1

Blank
Estuardo J. Miranda

vial # 39



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2822	1.771
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/8/2007 6:30:11 PM
Instrument 5
DB-ALC2

SIM 07015
Justin Knoy

vial # 54



#	Compound	Area	RT
1	ethanol	761	1.108
2	n-propanol	1533	1.945

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/8/2007 6:42:28 PM
Instrument 5
DB-ALC2

SIM 07015
Justin Knoy

vial # 55



#	Compound	Area	RT
1	ethanol	769	1.107
2	n-propanol	1514	1.945

Totals:



ethanol 0.102 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/8/2007 6:46:02 PM
Instrument 5
DB-ALC2

SIM 07015
Justin Knoy

vial # 56



#	Compound	Area	RT
1	ethanol	729	1.107
2	n-propanol	1452	1.945

Totals:



ethanol 0.101 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/8/2007 6:49:34 PM
Instrument 5
DB-ALC2

SIM 07015
Justin Knoy

vial # 57



#	Compound	Area	RT
1	ethanol	726	1.107
2	n-propanol	1452	1.944

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/8/2007 6:53:08 PM
Instrument 5
DB-ALC2

SIM 07015
Justin Knoy

vial # 58



#	Compound	Area	RT
1	ethanol	739	1.107
2	n-propanol	1481	1.944

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/8/2007 6:56:42 PM
Instrument 5
DB-ALC2

0.10 CTRL JK
Justin Knoy

vial # 59



#	Compound	Area	RT
1	ethanol	753	1.107
2	n-propanol	1536	1.945

Totals:



ethanol 0.099 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/8/2007 7:00:14 PM
Instrument 5
DB-ALC2

blank
Justin Knoy

vial # 60



#	Compound	Area	RT
1	ethanol	0	0.000
2	n-propanol	1547	1.944

Totals:



ethanol 0.000 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/9/2007 1:23:21 PM
Instrument 5
DB-ALC2

07015-1
SARAH SWENSON

vial # 56



#	Compound	Area	RT
1	ethanol	730	1.105
2	n-propanol	1499	1.941

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/9/2007 1:26:55 PM
Instrument 5
DB-ALC2

07015-2
SARAH SWENSON

vial # 57



#	Compound	Area	RT
1	ethanol	724	1.105
2	n-propanol	1481	1.941

Totals:



ethanol 0.101 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/9/2007 1:30:28 PM
Instrument 5
DB-ALC2

07015-3
SARAH SWENSON

vial # 58



#	Compound	Area	RT
1	ethanol	731	1.105
2	n-propanol	1499	1.941

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/9/2007 1:34:03 PM
Instrument 5
DB-ALC2

07015-4
SARAH SWENSON

vial # 59



#	Compound	Area	RT
1	ethanol	797	1.106
2	n-propanol	1603	1.942

Totals:



ethanol 0.102 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/9/2007 1:37:35 PM
Instrument 5
DB-ALC2

07015-5
SARAH SWENSON

vial # 60



#	Compound	Area	RT
1	ethanol	769	1.105
2	n-propanol	1570	1.941

Totals:



ethanol 0.101 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/9/2007 1:50:59 PM
Instrument 5
DB-ALC2

0.10 CTL-SS
SARAH SWENSON

vial # 61



#	Compound	Area	RT
1	ethanol	777	1.106
2	n-propanol	1589	1.942

Totals:



ethanol 0.101 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/9/2007 1:54:28 PM
Instrument 5
DB-ALC2

BLANK
SARAH SWENSON

vial # 62



#	Compound	Area	RT
1	ethanol	0	0.000
2	n-propanol	1558	1.941

Totals:



ethanol 0.000 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/10/2007 10:28:51 AM
Instrument 4
DB-ALC1

07015
bcapron

vial # 14



#	Compound	Area	RT
1	Ethanol	699	1.004
2	n-Propanol	1533	1.657

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/10/2007 10:32:06 AM
Instrument 4
DB-ALC1

07015
bcapron

vial # 15



#	Compound	Area	RT
1	Ethanol	690	1.004
2	n-Propanol	1487	1.657

Totals:



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/10/2007 10:35:25 AM
Instrument 4
DB-ALC1

07015
bcapron

vial # 16



#	Compound	Area	RT
1	Ethanol	682	1.004
2	n-Propanol	1486	1.657

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/10/2007 10:38:41 AM
Instrument 4
DB-ALC1

07015
bcapron

vial # 17



#	Compound	Area	RT
1	Ethanol	681	1.004
2	n-Propanol	1482	1.657

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/10/2007 10:42:00 AM
Instrument 4
DB-ALC1

07015
bcapron

vial # 18



#	Compound	Area	RT
1	Ethanol	682	1.004
2	n-Propanol	1485	1.656

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/10/2007 10:45:15 AM
Instrument 4
DB-ALC1

0.10 control bc
bcapron

vial # 19



#	Compound	Area	RT
1	Ethanol	708	1.004
2	n-Propanol	1565	1.657

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/10/2007 10:48:29 AM
Instrument 4
DB-ALC1

blank
bcapron

vial # 20



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	742	1.656

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/14/2007 10:10:29 AM
Instrument 5
DB-ALC2

sim 07015 a
Brianne E. Akins

vial # 13



#	Compound	Area	RT
1	ethanol	849	1.116
2	n-propanol	1703	1.952

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/14/2007 10:14:02 AM
Instrument 5
DB-ALC2

sim 07015 b
Brianne E. Akins

vial # 14



#	Compound	Area	RT
1	ethanol	960	1.116
2	n-propanol	1920	1.953

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/14/2007 10:17:32 AM
Instrument 5
DB-ALC2

sim 07015 c
Brianne E. Akins

vial # 15



#	Compound	Area	RT
1	ethanol	843	1.116
2	n-propanol	1671	1.953

Totals:



ethanol 0.101 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/14/2007 10:20:59 AM
Instrument 5
DB-ALC2

sim 07015 d
Brianne E. Akins

vial # 16



#	Compound	Area	RT
1	ethanol	839	1.117
2	n-propanol	1673	1.953

Totals:



ethanol 0.101 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/14/2007 10:24:23 AM
Instrument 5
DB-ALC2

sim 07015 e
Brianne E. Akins

vial # 17



#	Compound	Area	RT
1	ethanol	855	1.116
2	n-propanol	1708	1.953

Totals:



ethanol 0.101 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/14/2007 10:27:48 AM
Instrument 5
DB-ALC2

0.10 CONTROL-BA
Brianne E. Akins

vial # 18



#	Compound	Area	RT
1	ethanol	882	1.117
2	n-propanol	1772	1.954

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/14/2007 10:39:37 AM
Instrument 5
DB-ALC2

BLANK
Brianne E. Akins

vial # 19



#	Compound	Area	RT
1	ethanol	0	0.000
2	n-propanol	1753	1.954

Totals:



ethanol 0.000 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/14/2007 6:05:31 PM
Instrument 3
db-alc2

07015 sim
kgross

vial # 12



#	Compound	Area	RT
1	ETHANOL	706	1.062
2	n-PROPANOL	1493	1.822

Totals:



ETHANOL 0.102 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/14/2007 6:08:38 PM
Instrument 3
db-alc2

07015 sim
kgross

vial # 13



#	Compound	Area	RT
1	ETHANOL	734	1.061
2	n-PROPANOL	1548	1.821

Totals:



ETHANOL 0.102 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/14/2007 6:11:46 PM
Instrument 3
db-alc2

07015 sim
kgross

vial # 14



#	Compound	Area	RT
1	ETHANOL	729	1.062
2	n-PROPANOL	1535	1.822

Totals:



ETHANOL 0.102 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/14/2007 6:14:53 PM
Instrument 3
db-alc2

07015 sim
kgross

vial # 15



#	Compound	Area	RT
1	ETHANOL	707	1.061
2	n-PROPANOL	1501	1.821

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/14/2007 6:18:00 PM
Instrument 3
db-alc2

07015 sim
kgross

vial # 16



#	Compound	Area	RT
1	ETHANOL	741	1.061
2	n-PROPANOL	1567	1.821

Totals:



ETHANOL 0.102 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/14/2007 5:59:17 PM
Instrument 3
db-alc2

0.10 control kdg
kgross

vial # 10



#	Compound	Area	RT
1	ETHANOL	698	1.061
2	n-PROPANOL	1477	1.821

Totals:



ETHANOL 0.102 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/14/2007 6:02:24 PM
Instrument 3
db-alc2

blank
kgross

vial # 11



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1471	1.820

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/15/2007 2:50:06 PM
Instrument 5
DB-ALC2

07015 sim soln
m pemberton

vial # 13



#	Compound	Area	RT
1	ethanol	820	1.114
2	n-propanol	1660	1.951

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/15/2007 2:53:38 PM
Instrument 5
DB-ALC2

07015 sim soln
m pemberton

vial # 14



#	Compound	Area	RT
1	ethanol	871	1.116
2	n-propanol	1732	1.953

Totals:



ethanol 0.102 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/15/2007 2:57:08 PM
Instrument 5
DB-ALC2

07015 sim soln
m pemberton

vial # 15



#	Compound	Area	RT
1	ethanol	826	1.115
2	n-propanol	1681	1.952

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/15/2007 3:00:35 PM
Instrument 5
DB-ALC2

07015 sim soln
m pemberton

vial # 16



#	Compound	Area	RT
1	ethanol	847	1.116
2	n-propanol	1695	1.954

Totals:



ethanol 0.101 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/15/2007 3:04:00 PM
Instrument 5
DB-ALC2

07015 sim soln
m pemberton

vial # 17



#	Compound	Area	RT
1	ethanol	907	1.117
2	n-propanol	1800	1.955

Totals:



ethanol 0.102 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/15/2007 3:07:25 PM
Instrument 5
DB-ALC2

0.10 controlmp
m pemberton

vial # 18



#	Compound	Area	RT
1	ethanol	908	1.118
2	n-propanol	1813	1.956

Totals:



ethanol 0.102 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/15/2007 3:19:39 PM
Instrument 5
DB-ALC2

blk
m pemberton

vial # 19



#	Compound	Area	RT
1	ethanol	0	0.000
2	n-propanol	1715	1.952

Totals:



ethanol 0.000 g/100ml



n-propanol 1.000 g/100ml

Sequence Parameters:

Operator:	RFLAHERTY
Data File Naming:	Prefix/Counter
Signal 1 Prefix:	SIG1
Counter:	0001
Signal 2 Prefix:	SIG2
Counter:	0001
Data Directory:	C:\HPCHEM\2\DATA\
Data Subdirectory:	070516RF
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	BLDALCO3	1	Sample		
2	Vial 2	0.079 CAL	BLDALCO3	1	Calib		
3	Vial 3	0.158 CAL	BLDALCO3	1	Calib		
4	Vial 4	0.316 CAL	BLDALCO3	1	Calib		
5	Vial 5	BLANK	BLDALCO3	1	Sample		
6	Vial 6	0.02 CONTROL-RF	BLDALCO3	1	Ctrl Samp		
7	Vial 7	0.04 CONTROL-RF	BLDALCO3	1	Ctrl Samp		
8	Vial 8	0.10 CONTROL-RF	BLDALCO3	1	Ctrl Samp		
9	Vial 9	0.20 CONTROL-RF	BLDALCO3	1	Ctrl Samp		
10	Vial 10	BLANK	BLDALCO3	1	Sample		
11	Vial 11	07015-RF	BLDALCO3	1	Sample		
12	Vial 12	07015-RF	BLDALCO3	1	Sample		
13	Vial 13	07015-RF	BLDALCO3	1	Sample		
14	Vial 14	07015-RF	BLDALCO3	1	Sample		
15	Vial 15	07015-RF	BLDALCO3	1	Sample		
16	Vial 16	0.10 CONTROL-AG	BLDALCO3	1	Ctrl Samp		
17	Vial 17	BLANK	BLDALCO3	1	Sample		

Sequence Table (Back Injector):

No entries - empty table!

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/16/2007 1:52:45 PM
Instrument 3
db-alc2

BLANK
RFLAHERTY

vial # 1



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1464	1.821

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/16/2007 1:55:52 PM
Instrument 3
db-alc2

0.079 CAL
RFLAHERTY

vial # 2



#	Compound	Area	RT
1	ETHANOL	545	1.061
2	n-PROPANOL	1455	1.821

Totals:



ETHANOL 0.080 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/16/2007 1:58:59 PM
Instrument 3
db-alc2

0.158 CAL
RFLAHERTY

vial # 3



#	Compound	Area	RT
1	ETHANOL	1085	1.062
2	n-PROPANOL	1439	1.822

Totals:



ETHANOL 0.161 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/16/2007 2:02:06 PM
Instrument 3
db-alc2

0.316 CAL
RFLAHERTY

vial # 4



#	Compound	Area	RT
1	ETHANOL	2175	1.062
2	n-PROPANOL	1446	1.821

Totals:



ETHANOL 0.316 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/16/2007 2:05:13 PM
Instrument 3
db-alc2

BLANK
RFLAHERTY

vial # 5



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1453	1.820

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/16/2007 2:08:21 PM
Instrument 3
db-alc2

0.02 CONTROL-RF
RFLAHERTY

vial # 6



#	Compound	Area	RT
1	ETHANOL	135	1.061
2	n-PROPANOL	1444	1.820

Totals:



ETHANOL 0.020 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/16/2007 2:11:28 PM
Instrument 3
db-alc2

0.04 CONTROL-RF
RFLAHERTY

vial # 7



#	Compound	Area	RT
1	ETHANOL	270	1.061
2	n-PROPANOL	1456	1.820

Totals:



ETHANOL 0.039 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/16/2007 2:14:35 PM
Instrument 3
db-alc2

0.10 CONTROL-RF
RFLAHERTY

vial # 8



#	Compound	Area	RT
1	ETHANOL	666	1.061
2	n-PROPANOL	1429	1.820

Totals:



ETHANOL 0.098 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/16/2007 2:17:42 PM
Instrument 3
db-alc2

0.20 CONTROL-RF
RFLAHERTY

vial # 9



#	Compound	Area	RT
1	ETHANOL	1338	1.061
2	n-PROPANOL	1432	1.820

Totals:



ETHANOL 0.196 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/16/2007 2:20:49 PM
Instrument 3
db-alc2

BLANK
RFLAHERTY

vial # 10



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1474	1.822

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/16/2007 2:23:56 PM
Instrument 3
db-alc2

07015-RF
RFLAHERTY

vial # 11



#	Compound	Area	RT
1	ETHANOL	690	1.061
2	n-PROPANOL	1438	1.821

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/16/2007 2:27:03 PM
Instrument 3
db-alc2

07015-RF
RFLAHERTY

vial # 12



#	Compound	Area	RT
1	ETHANOL	690	1.061
2	n-PROPANOL	1438	1.820

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/16/2007 2:30:10 PM
Instrument 3
db-alc2

07015-RF
RFLAHERTY

vial # 13



#	Compound	Area	RT
1	ETHANOL	694	1.061
2	n-PROPANOL	1445	1.820

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/16/2007 2:33:17 PM
Instrument 3
db-alc2

07015-RF
RFLAHERTY

vial # 14



#	Compound	Area	RT
1	ETHANOL	698	1.061
2	n-PROPANOL	1451	1.820

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/16/2007 2:36:25 PM
Instrument 3
db-alc2

07015-RF
RFLAHERTY

vial # 15



#	Compound	Area	RT
1	ETHANOL	701	1.061
2	n-PROPANOL	1459	1.820

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/16/2007 2:39:32 PM
Instrument 3
db-alc2

0.10 CONTROL-~~AG~~ RF
RFLAHERTY

5/16/07
RF

vial # 16



#	Compound	Area	RT
1	ETHANOL	693	1.061
2	n-PROPANOL	1480	1.820

Totals:



ETHANOL 0.098 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/16/2007 2:42:39 PM
Instrument 3
db-alc2

BLANK
RFLAHERTY

vial # 17



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1508	1.820

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/17/2007 12:40:14 PM
Instrument 4
DB-ALC1

07015
ED FORMOSO

vial # 21



#	Compound	Area	RT
1	Ethanol	720	1.004
2	n-Propanol	1581	1.658

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/17/2007 12:48:53 PM
Instrument 4
DB-ALC1

07015
ED FORMOSO

vial # 22



#	Compound	Area	RT
1	Ethanol	753	1.004
2	n-Propanol	1642	1.658

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/17/2007 12:52:09 PM
Instrument 4
DB-ALC1

07015
ED FORMOSO

vial # 23



#	Compound	Area	RT
1	Ethanol	745	1.004
2	n-Propanol	1623	1.658

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/17/2007 12:55:42 PM
Instrument 4
DB-ALC1

07015
ED FORMOSO

vial # 24



#	Compound	Area	RT
1	Ethanol	723	1.004
2	n-Propanol	1581	1.658

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/17/2007 12:59:03 PM
Instrument 4
DB-ALC1

07015
ED FORMOSO

vial # 25



#	Compound	Area	RT
1	Ethanol	725	1.004
2	n-Propanol	1589	1.658

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/17/2007 1:02:23 PM
Instrument 4
DB-ALC1

0.10 CONTROL
ED FORMOSO

vial # 26



#	Compound	Area	RT
1	Ethanol	724	1.004
2	n-Propanol	1640	1.658

Totals:



Ethanol 0.097 g/100ml



n-Propanol 1.000 g/100ml

Sequence Parameters:

Operator:	BB
Data File Naming:	Prefix/Counter
Signal 1 Prefix:	SIG1
Counter:	0001
Signal 2 Prefix:	SIG2
Counter:	0001
Data Directory:	C:\HPCHEM\2\DATA\
Data Subdirectory:	070518BB
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	BLDALCO3	1	Sample		
2	Vial 2	0.079 CAL	BLDALCO3	1	Calib		
3	Vial 3	0.158 CAL	BLDALCO3	1	Calib		
4	Vial 4	0.316 CAL	BLDALCO3	1	Calib		
5	Vial 5	BLANK	BLDALCO3	1	Sample		
6	Vial 6	0.04 CONTROL-BB	BLDALCO3	1	Ctrl Samp		
7	Vial 7	0.10 CONTROL-BB	BLDALCO3	1	Ctrl Samp		
8	Vial 8	0.20 CONTROL-BB	BLDALCO3	1	Ctrl Samp		
9	Vial 9	SIM015-A	BLDALCO3	1	Sample		
10	Vial 10	SIM015-B	BLDALCO3	1	Sample		
11	Vial 11	SIM015-C	BLDALCO3	1	Sample		
12	Vial 12	SIM015-D	BLDALCO3	1	Sample		
13	Vial 13	SIM015-E	BLDALCO3	1	Sample		
14	Vial 14	0.10 CONTROL-BB	BLDALCO3	1	Sample		
15	Vial 15	BLANK	BLDALCO3	1	Sample		

Sequence Table (Back Injector):

No entries - empty table!

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/18/2007 8:19:05 AM
Instrument 3
db-alc2

BLANK
BB

vial # 1



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1459	1.821

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/18/2007 8:22:13 AM 0.079 CAL
Instrument 3 BB
db-alc2

vial # 2



#	Compound	Area	RT
1	ETHANOL	554	1.061
2	n-PROPANOL	1460	1.821

Totals:



ETHANOL 0.079 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/18/2007 8:25:20 AM 0.158 CAL
Instrument 3 BB
db-alc2

vial # 3



#	Compound	Area	RT
1	ETHANOL	1093	1.062
2	n-PROPANOL	1441	1.821

Totals:



ETHANOL 0.159 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/18/2007 8:28:27 AM 0.316 CAL
Instrument 3 BB
db-alc2

vial # 4



#	Compound	Area	RT
1	ETHANOL	2174	1.062
2	n-PROPANOL	1449	1.820

Totals:



ETHANOL 0.315 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/18/2007 8:31:34 AM
Instrument 3
db-alc2

BLANK
BB

vial # 5



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1295	1.820

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/18/2007 8:34:42 AM
Instrument 3
db-alc2

0.04 CONTROL-BB
BB

vial # 6



#	Compound	Area	RT
1	ETHANOL	283	1.061
2	n-PROPANOL	1462	1.821

Totals:



ETHANOL 0.040 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/18/2007 8:37:49 AM 0.10 CONTROL-BB
Instrument 3 BB
db-alc2

vial # 7



#	Compound	Area	RT
1	ETHANOL	689	1.061
2	n-PROPANOL	1444	1.821

Totals:



ETHANOL 0.100 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/18/2007 8:40:56 AM 0.20 CONTROL-BB
Instrument 3 BB
db-alc2

vial # 8



#	Compound	Area	RT
1	ETHANOL	1376	1.062
2	n-PROPANOL	1440	1.821

Totals:



ETHANOL 0.201 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/18/2007 8:44:03 AM SIM015-A
Instrument 3 BB
db-alc2

vial # 9



#	Compound	Area	RT
1	ETHANOL	705	1.062
2	n-PROPANOL	1449	1.822

Totals:



ETHANOL 0.102 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/18/2007 8:47:11 AM SIM015-B
Instrument 3 BB
db-alc2

vial # 10



#	Compound	Area	RT
1	ETHANOL	697	1.062
2	n-PROPANOL	1441	1.821
Totals:			



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/18/2007 8:50:18 AM
Instrument 3
db-alc2

SIM015-C
BB

vial # 11



#	Compound	Area	RT
1	ETHANOL	706	1.062
2	n-PROPANOL	1454	1.822

Totals:



ETHANOL 0.102 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/18/2007 8:53:25 AM
Instrument 3
db-alc2

SIM015-D
BB

vial # 12



#	Compound	Area	RT
1	ETHANOL	712	1.062
2	n-PROPANOL	1460	1.822

Totals:



ETHANOL 0.102 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/18/2007 8:56:32 AM
Instrument 3
db-alc2

SIM015-E
BB

vial # 13



#	Compound	Area	RT
1	ETHANOL	706	1.061
2	n-PROPANOL	1455	1.821

Totals:



ETHANOL 0.102 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/18/2007 8:59:40 AM
Instrument 3
db-alc2

0.10 CONTROL-BB
BB

vial # 14



#	Compound	Area	RT
1	ETHANOL	706	1.062
2	n-PROPANOL	1481	1.821

Totals:



ETHANOL 0.100 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/18/2007 9:02:47 AM
Instrument 3
db-alc2

BLANK
BB

vial # 15



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1479	1.820

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/18/2007 1:38:46 PM
Instrument 4
DB-ALC1

07015-1
Lisa Piquette

vial # 64



#	Compound	Area	RT
1	Ethanol	663	1.003
2	n-Propanol	1463	1.655

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/18/2007 1:41:59 PM
Instrument 4
DB-ALC1

07015-2
Lisa Piquette

vial # 65



#	Compound	Area	RT
1	Ethanol	682	1.003
2	n-Propanol	1509	1.655

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/18/2007 1:45:11 PM
Instrument 4
DB-ALC1

07015-3
Lisa Piquette

vial # 66



#	Compound	Area	RT
1	Ethanol	700	1.003
2	n-Propanol	1539	1.655

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/18/2007 1:48:23 PM
Instrument 4
DB-ALC1

07015-4
Lisa Piquette

vial # 67



#	Compound	Area	RT
1	Ethanol	675	1.002
2	n-Propanol	1492	1.655

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/18/2007 1:54:01 PM
Instrument 4
DB-ALC1

07015-5
Lisa Piquette

vial # 68



#	Compound	Area	RT
1	Ethanol	683	1.003
2	n-Propanol	1506	1.655

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/18/2007 1:57:20 PM
Instrument 4
DB-ALC1

BLANK
Lisa Piquette

vial # 69



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1565	1.655

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/18/2007 2:00:38 PM
Instrument 4
DB-ALC1

0.10 CONTROL LP
Lisa Piquette

vial # 70



#	Compound	Area	RT
1	Ethanol	694	1.002
2	n-Propanol	1558	1.655

Totals:



Ethanol 0.098 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/22/2007 2:54:42 PM
Instrument 1
DB ALC 1

BLANK
Chris Johnston

vial # 1



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2733	1.767
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/22/2007 2:57:47 PM
Instrument 1
DB ALC 1

SIM07015
Chris Johnston

vial # 2



#	Compound	Area	RT
1	Ethanol	1265	1.076
2	n-Propanol	2760	1.767
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/22/2007 3:00:52 PM
Instrument 1
DB ALC 1

SIM07015
Chris Johnston

vial # 3



#	Compound	Area	RT
1	Ethanol	1262	1.077
2	n-Propanol	2745	1.767
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/22/2007 3:03:56 PM
Instrument 1
DB ALC 1

SIM07015
Chris Johnston

vial # 4



#	Compound	Area	RT
1	Ethanol	1273	1.076
2	n-Propanol	2785	1.767
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/22/2007 3:07:01 PM
Instrument 1
DB ALC 1

SIM07015
Chris Johnston

vial # 5



#	Compound	Area	RT
1	Ethanol	1265	1.076
2	n-Propanol	2757	1.767
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/22/2007 3:10:06 PM
Instrument 1
DB ALC 1

SIM07015
Chris Johnston

vial # 6



#	Compound	Area	RT
1	Ethanol	1290	1.077
2	n-Propanol	2805	1.767
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/22/2007 3:13:11 PM
Instrument 1
DB ALC 1

0.10 CONTROL-CJ
Chris Johnston

vial # 7



#	Compound	Area	RT
1	Ethanol	1278	1.077
2	n-Propanol	2828	1.768
Tot			



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/22/2007 3:16:16 PM
Instrument 1
DB ALC 1

BLANK
Chris Johnston

vial # 8



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2868	1.767
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/22/2007 2:01:39 PM
Instrument 1
DB ALC 1

07015 Sim-1
N Nuwayhid, PhD

vial # 33



#	Compound	Area	RT
1	Ethanol	1294	1.076
2	n-Propanol	2806	1.767
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/22/2007 2:04:44 PM
Instrument 1
DB ALC 1

07015 Sim-2
N Nuwayhid, PhD

vial # 34



#	Compound	Area	RT
1	Ethanol	1303	1.076
2	n-Propanol	2818	1.767
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/22/2007 2:07:49 PM
Instrument 1
DB ALC 1

07015 Sim-3
N Nuwayhid, PhD

vial # 35



#	Compound	Area	RT
1	Ethanol	1303	1.076
2	n-Propanol	2826	1.767
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/22/2007 2:10:54 PM
Instrument 1
DB ALC 1

07015 Sim-4
N Nuwayhid, PhD

vial # 36



#	Compound	Area	RT
1	Ethanol	1318	1.076
2	n-Propanol	2851	1.767
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/22/2007 2:13:59 PM
Instrument 1
DB ALC 1

07015 Sim-5
N Nuwayhid, PhD

vial # 37



#	Compound	Area	RT
1	Ethanol	1340	1.076
2	n-Propanol	2900	1.767
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/22/2007 2:17:03 PM
Instrument 1
DB ALC 1

0.10 Ctrl-NN
N Nuwayhid, PhD

vial # 38



#	Compound	Area	RT
1	Ethanol	1364	1.076
2	n-Propanol	3050	1.767
Tot			



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/22/2007 2:20:08 PM
Instrument 1
DB ALC 1

Blank
N Nuwayhid, PhD

vial # 39



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2963	1.767
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

Sequence Parameters:

Operator:	alouis
Data File Naming:	Prefix/Counter
Signal 1 Prefix:	SIG1
Counter:	0001
Signal 2 Prefix:	SIG2
Counter:	0001
Data Directory:	C:\HPCHEM\1\DATA\
Data Subdirectory:	070515A1
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	blank	BLDALCO	1	Sample		
2	Vial 2	0.079 std	BLDALCO	1	Calib		
3	Vial 3	0.158 std	BLDALCO	1	Calib		
4	Vial 4	0.316 std	BLDALCO	1	Calib		
5	Vial 5	blank	BLDALCO	1	Ctrl Samp		
6	Vial 8	0.02 std	BLDALCO	1	Sample		
7	Vial 9	0.04 con al	BLDALCO	1	Ctrl Samp		
8	Vial 10	0.10 con al	BLDALCO	1	Ctrl Samp		
9	Vial 11	0.20 con al	BLDALCO	1	Ctrl Samp		
10	Vial 12	blank	BLDALCO	1	Sample		
11	Vial 13	07015a	BLDALCO	1	Sample		
12	Vial 14	07015b	BLDALCO	1	Sample		
13	Vial 15	07015c	BLDALCO	1	Sample		
14	Vial 16	07015d	BLDALCO	1	Sample		
15	Vial 17	07015e	BLDALCO	1	Sample		

Sequence Table (Back Injector):

No entries - empty table!

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/15/2007 2:18:03 PM
Instrument 1
DB ALC 1

blank
alouis

vial # 1



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2801	1.766
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/15/2007 2:21:07 PM
Instrument 1
DB ALC 1

0.079 std
alouis

vial # 2



#	Compound	Area	RT
1	Ethanol	993	1.076
2	n-Propanol	2766	1.766
Tot			



Ethanol 0.080 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/15/2007 2:24:12 PM
Instrument 1
DB ALC 1

0.158 std
alouis

vial # 3



#	Compound	Area	RT
1	Ethanol	1990	1.076
2	n-Propanol	2767	1.766
Tot			



Ethanol 0.161 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/15/2007 2:27:17 PM
Instrument 1
DB ALC 1

0.316 std
alouis

vial # 4



#	Compound	Area	RT
1	Ethanol	4028	1.077
2	n-Propanol	2850	1.767
Tot			



Ethanol 0.315 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/15/2007 2:30:22 PM
Instrument 1
DB ALC 1

blank
alouis

vial # 5



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2774	1.767
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/15/2007 2:33:27 PM
Instrument 1
DB ALC 1

0.02 std
alouis

vial # 8



#	Compound	Area	RT
1	Ethanol	255	1.076
2	n-Propanol	2810	1.767
Tot			



Ethanol 0.019 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/15/2007 2:36:31 PM
Instrument 1
DB ALC 1

0.04 con al
alouis

vial # 9



#	Compound	Area	RT
1	Ethanol	507	1.076
2	n-Propanol	2829	1.767
Tot			



Ethanol 0.039 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/15/2007 2:39:36 PM
Instrument 1
DB ALC 1

0.10 con al
alouis

vial # 10



#	Compound	Area	RT
1	Ethanol	1438	1.076
2	n-Propanol	3256	1.766
Tot			



Ethanol 0.098 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/15/2007 2:42:41 PM
Instrument 1
DB ALC 1

0.20 con al
alouis

vial # 11



#	Compound	Area	RT
1	Ethanol	2596	1.077
2	n-Propanol	2895	1.767
Tot			



Ethanol 0.200 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/15/2007 2:45:46 PM
Instrument 1
DB ALC 1

blank
alouis

vial # 12



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2850	1.767
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/15/2007 2:48:51 PM
Instrument 1
DB ALC 1

07015a
alouis

vial # 13



#	Compound	Area	RT
1	Ethanol	1283	1.076
2	n-Propanol	2804	1.767
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/15/2007 2:51:55 PM 07015b
Instrument 1 alouis
DB ALC 1

vial # 14



#	Compound	Area	RT
1	Ethanol	1304	1.076
2	n-Propanol	2854	1.767
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/15/2007 2:55:00 PM
Instrument 1
DB ALC 1

07015c
alouis

vial # 15



#	Compound	Area	RT
1	Ethanol	1328	1.076
2	n-Propanol	2891	1.767
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/15/2007 2:58:05 PM
Instrument 1
DB ALC 1

07015d
alouis

vial # 16



#	Compound	Area	RT
1	Ethanol	1321	1.077
2	n-Propanol	2899	1.767
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/15/2007 3:01:10 PM
Instrument 1
DB ALC 1

07015e
alouis

vial # 17



#	Compound	Area	RT
1	Ethanol	1362	1.076
2	n-Propanol	2977	1.767
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

Sequence Parameters:

Operator:	Amanda Black
Data File Naming:	Prefix/Counter
Signal 1 Prefix:	SIG1
Counter:	0001
Signal 2 Prefix:	SIG2
Counter:	0001
Data Directory:	C:\HPCHEM\1\DATA\
Data Subdirectory:	070523AB
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	BLDALCO	1	Sample		
2	Vial 2	0.079 CAL	BLDALCO	1	Calib		
3	Vial 3	0.158 CAL	BLDALCO	1	Calib		
4	Vial 4	0.316 CAL	BLDALCO	1	Calib		
5	Vial 5	BLANK	BLDALCO	1	Sample		
6	Vial 6	0.04 MIX	VOL	1	Calib		
7	Vial 7	0.08 MIX	VOL	1	Calib		
8	Vial 8	0.02 STD	BLDALCO	1	Sample		
9	Vial 9	0.04 CONTROL-AB	BLDALCO	1	Ctrl Samp		
10	Vial 10	0.10 CONTROL-AB	BLDALCO	1	Ctrl Samp		
11	Vial 11	0.20 CONTROL-AB	BLDALCO	1	Ctrl Samp		
12	Vial 12	BLANK	BLDALCO	1	Sample		
13	Vial 13	0.10 CONTROL-AB	BLDALCO	1	Ctrl Samp		
14	Vial 14	BLANK	BLDALCO	1	Sample		
15	Vial 15	SIM 07015-1	BLDALCO	1	Sample		
16	Vial 16	SIM 07015-2	BLDALCO	1	Sample		
17	Vial 17	SIM 07015-3	BLDALCO	1	Sample		
18	Vial 18	SIM 07015-4	BLDALCO	1	Sample		
19	Vial 19	SIM 07015-5	BLDALCO	1	Sample		
20	Vial 20	0.10 CONTROL-AB	BLDALCO	1	Sample		
21	Vial 21	BLANK	BLDALCO	1	Sample		

Sequence Table (Back Injector):

No entries - empty table!

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 11:39:57 AM
Instrument 1
DB ALC 1

BLANK
Amanda Black

vial # 1



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2810	1.768
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 11:43:02 AM
Instrument 1
DB ALC 1

0.079 CAL
Amanda Black

vial # 2



#	Compound	Area	RT
1	Ethanol	1012	1.077
2	n-Propanol	2815	1.768
Tot			



Ethanol 0.079 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 11:46:07 AM
Instrument 1
DB ALC 1

0.158 CAL
Amanda Black

vial # 3



#	Compound	Area	RT
1	Ethanol	1989	1.077
2	n-Propanol	2770	1.768
Tot			



Ethanol 0.159 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 11:49:11 AM
Instrument 1
DB ALC 1

0.316 CAL
Amanda Black

vial # 4



#	Compound	Area	RT
1	Ethanol	3980	1.077
2	n-Propanol	2800	1.768
Tot			



Ethanol 0.315 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 11:52:16 AM
Instrument 1
DB ALC 1

BLANK
Amanda Black

vial # 5



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2787	1.768
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\VOL.M
5/23/2007 11:55:28 AM
Instrument 1
DB ALC 1

0.04 MIX
Amanda Black

vial # 6



#	Compound	Area	RT
1	METHANOL	256	0.855
2	ETHANOL	522	1.075
3	ISOPROPANOL	972	1.311
4	ACETONE	2056	1.579
5	n-PROPANOL	2891	1.767
Tot			



METHANOL 40.700 mg/dL



ETHANOL 40.593 mg/dL



ISOPROPANOL 40.324 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\VOL.M



ACETONE 40.597 mg/dL



n-PROPANOL 1.000 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\VOL.M
5/23/2007 11:58:33 AM
Instrument 1
DB ALC 1

0.08 MIX
Amanda Black

vial # 7



#	Compound	Area	RT
1	METHANOL	509	0.855
2	ETHANOL	1042	1.075
3	ISOPROPANOL	1960	1.311
4	ACETONE	4130	1.579
5	n-PROPANOL	2944	1.767
Tot			



METHANOL 79.713 mg/dL



ETHANOL 79.737 mg/dL



ISOPROPANOL 79.863 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\VOL.M



ACETONE 79.818 mg/dL



n-PROPANOL 1.000 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 12:01:46 PM
Instrument 1
DB ALC 1

0.02 STD
Amanda Black

vial # 8



#	Compound	Area	RT
1	Ethanol	261	1.077
2	n-Propanol	2826	1.768
Tot			



Ethanol 0.020 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 12:04:51 PM
Instrument 1
DB ALC 1

0.04 CONTROL-AB
Amanda Black

vial # 9



#	Compound	Area	RT
1	Ethanol	507	1.077
2	n-Propanol	2814	1.768
Tot			



Ethanol 0.039 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 12:07:55 PM
Instrument 1
DB ALC 1

0.10 CONTROL-AB
Amanda Black

vial # 10



#	Compound	Area	RT
1	Ethanol	1262	1.077
2	n-Propanol	2809	1.768
Tot			



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 12:11:00 PM
Instrument 1
DB ALC 1

0.20 CONTROL-AB
Amanda Black

vial # 11



#	Compound	Area	RT
1	Ethanol	2542	1.077
2	n-Propanol	2823	1.768
Tot			



Ethanol 0.200 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 12:14:05 PM
Instrument 1
DB ALC 1

BLANK
Amanda Black

vial # 12



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2924	1.768
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 12:17:10 PM
Instrument 1
DB ALC 1

0.10 CONTROL-AB
Amanda Black

vial # 13



#	Compound	Area	RT
1	Ethanol	1259	1.077
2	n-Propanol	2790	1.768
Tot			



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 12:20:15 PM
Instrument 1
DB ALC 1

BLANK
Amanda Black

vial # 14



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2802	1.768
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 12:23:19 PM
Instrument 1
DB ALC 1

SIM 07015-1
Amanda Black

vial # 15



#	Compound	Area	RT
1	Ethanol	1309	1.077
2	n-Propanol	2873	1.767
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 12:26:24 PM
Instrument 1
DB ALC 1

SIM 07015-2
Amanda Black

vial # 16



#	Compound	Area	RT
1	Ethanol	1294	1.077
2	n-Propanol	2831	1.768
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 12:29:29 PM
Instrument 1
DB ALC 1

SIM 07015-3
Amanda Black

vial # 17



#	Compound	Area	RT
1	Ethanol	1271	1.077
2	n-Propanol	2768	1.768
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 12:32:34 PM
Instrument 1
DB ALC 1

SIM 07015-4
Amanda Black

vial # 18



#	Compound	Area	RT
1	Ethanol	1288	1.077
2	n-Propanol	2820	1.768
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 12:35:39 PM
Instrument 1
DB ALC 1

SIM 07015-5
Amanda Black

vial # 19



#	Compound	Area	RT
1	Ethanol	1319	1.077
2	n-Propanol	2864	1.768
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 12:38:43 PM
Instrument 1
DB ALC 1

0.10 CONTROL-AB
Amanda Black

vial # 20



#	Compound	Area	RT
1	Ethanol	1302	1.077
2	n-Propanol	2886	1.768
Tot			



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
5/23/2007 12:41:48 PM
Instrument 1
DB ALC 1

BLANK
Amanda Black

vial # 21



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2972	1.768
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml