Notice of Simulator Solution File Review

At the request of the State Toxicologist a review of the following simulator solution records has been accomplished. The following file consists of simulator solution analyses performed and completed by the State Toxicology Laboratory for a specific batch number. The file contains the simulator solution data entry form along with a file review record and the chromatograms generated by the Toxicology Laboratory during the analyses of the solutions. This file has been reviewed by Tpr. Ken Denton and Mr. Rod Gullberg for accuracy and completeness. Where computations regarding simulator solution values have been found to be incorrect, the corrected values have been written in by Mr. Rod Gullberg along with initials and date. The corrected values were then evaluated to ensure that the solution still conformed to those standards established by the State Toxicologist.

Where computation values changed for a specific batch number, the analysts employed by the State Toxicology Laboratory were asked to review the revisions, ensure the solution complied with the criteria established by the State Toxicologist and then re-sign their affidavit. Their signature will appear on their original affidavit along with a statement regarding their review of the results.

Where a dating error occurred that analyst will have made the correction on the original data form including their initials and date and then re-signed their original affidavit.

K.L.D. 10/1/2007

Tpr. Ken Denton Date

R.G. Gullberg 10-1-07

Rod G. Gullberg Date

Washington State Toxicology Laboratory

Simulator Solution Data Entry Review Form

Reviewer KEN DENTON / ROD GULLBERG Date 9-27-07

Location TOX LAB SEATTLE Batch Number 07016

Form Review Criteria

Preparation date precedes all analysis dates: Okay ___ Not Okay X

Data entry corresponds to all chromatograms: Okay X Not Okay ___

All signatures present: Okay X Not Okay ___

Computations:

Avg. solution concentration: Correct ___ Not Correct X

Standard deviation: Correct ___ Not Correct X

Range: Correct ___ Not Correct X

Precision: Correct ___ Not Correct X

Equivalent vapor concent.: Correct ___ Not Correct X

External Control Information
(lot # and future date): Correct X Not Correct ___

Complies with accuracy and precision requirements established by the State Toxicologist: Yes X No ___

Corrections Necessary:
DATE OF ANALYSIS DOES NOT CORRESPOND TO CHROMATOGRAMS FOR BRITTANY BALL AND REBECCA F.

Comments:

Reviewer Signature: R G Gullberg Date: 9-27-07

Reviewer Signature: K.L. D Date: 9/27/2007

9/26/2007

WASHINGTON STATE TOXICOLOGY LABORATORY
FORENSIC LABORATORY SERVICES BUREAU
WASHINGTON STATE PATROL
2203 AIRPORT WAY S, SUITE 360
SEATTLE, WASHINGTON 98134-2027
(206) 262-6100 FAX (206) 262-6145

Preparation and certification of **0.08** g/210L Simulator External Standard solution

Batch number **07016** Date: 5/30/2007

Preparation: 66.5 mL of absolute ethyl alcohol diluted to 52 Liters with water

Concentration of ethanol (g/100mL) measured by gas chromatography:

	Anal 1	Anal 2	Anal 3	Anal 4	Anal 5	Anal 6	Anal 7	Anal 8	Anal 9	Anal10	Anal 11	Anal 12	Anal 13	Anal14	Anal 15	Anal 16
1	0.100	0.099	0.100	0.100	0.103	0.101	0.099	0.102	0.101	0.101	0.101	0.104	0.098	0.098	0.101	0.099
2	0.099	0.100	0.100	0.102	0.100	0.101	0.099	0.102	0.101	0.104	0.100	0.102	0.100	0.101	0.101	0.100
3	0.100	0.102	0.100	0.100	0.101	0.101	0.099	0.102	0.101	0.099	0.099	0.103	0.099	0.101	0.101	0.100
4	0.101	0.102	0.100	0.101	0.101	0.101	0.100	0.102	0.101	0.100	0.099	0.103	0.099	0.101	0.102	0.100
5	0.099	0.101	0.100	0.101	0.100	0.101	0.101	0.102	0.101	0.100	0.099	0.102	0.099	0.101	0.101	0.100
Ctrl	0.098	0.097	0.098	0.098	0.099	0.098	0.097	0.100	0.099	0.099	0.097	0.100	0.097	0.099	0.099	0.098

RJB 9-27-07

External Control:
Lot #: A048730 Exp date: 3/2011
Target concentration: 0.10 g/100mL

RJB 9-27-07

Equivalent vapor concent.: ~~0.0820~~ 0.0818 g/210L

Statistics:
Avg. solution concent.: ~~0.1008~~ 0.1006 g/100 mL
SD: ~~0.00126~~ 0.00125 (0.0969)
Range (3xSD): ~~0.0970~~ to ~~0.1046~~ 0.1043
Precision CV (%): ~~1.2540~~ 1.2413 %

Analyst	Name	Signature	Date
1	Christopher S Johnston	Christopher S. Johnston	06/01/2007
2	Justin L Knoy	Justin L Knoy	05/30/2007
3	Sarah M Swenson		05/30/2007
4	Estuardo J. Miranda		05/31/2007
5	Asa Louis		06/01/2007
6	Brittany Ball	Brittany Ball	~~06/01/2007~~ 06/04/2007 BB 10/1/07
7	Brian Capron		06/04/2007
8	Edward Formoso		06/05/2007
9	Kelly Gross		06/05/2007
10	Melissa Pemberton	Melissa L Pemberton	06/06/2007
11	Naziha Nuwayhid, PhD		06/05/2007
12	Rebecca Flaherty		~~06/07/2007~~ 06/08/2007 RF 09/28/07
13	Lisa Noble		06/08/2007
14	Amanda Black		06/12/2007
15	Brianna Peterson	Brianna Peterson	06/12/2007
16	Brianne Akins	Brianne Akins	06/13/2007

Prepared by: Christopher S Johnston according to the approved protocol

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Christopher S. Johnston, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Biochemistry.

The simulator solution, Lot Number 07016, was prepared in the Washington State Toxicology Laboratory on 5/30/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/30/2008.

Dated: 7/12/2007
Seattle, WA

Christopher S. Johnston
Forensic Toxicologist

CSJ/jr
CJSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions. 9.28.07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Justin L. Knoy, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Biology, and MS degree in Forensic Science.

The simulator solution, Lot Number 07016, was prepared in the Washington State Toxicology Laboratory on 5/30/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/30/2008.

Dated: 7/12/2007
Seattle, WA

Justin L. Knoy
Forensic Toxicologist

JLK/jr
JKSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions.

10/1/07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Sarah M Swenson, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Chemistry and over three years of experience in forensic toxicology.

The simulator solution, Lot Number 07016, was prepared in the Washington State Toxicology Laboratory on 5/30/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/30/2008.

Dated: 7/12/2007
Seattle, WA

Sarah M Swenson
Forensic Toxicologist

SS/jr
SSSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions.

10/1/07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Estuardo J. Miranda, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: Bachelor of Science in Chemistry, Master of Science in Zoology and nine years experience in Forensic Toxicology.

The simulator solution, Lot Number 07016, was prepared in the Washington State Toxicology Laboratory on 5/30/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/30/2008.

Dated: 7/12/2007
Seattle, WA

Estuardo J. Miranda
Forensic Toxicologist

EM/jr
EMSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions. 9-28-07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Asa J. Louis, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: B.S. degree in Biochemistry and eight years of toxicology experience.

The simulator solution, Lot Number 07016, was prepared in the Washington State Toxicology Laboratory on 5/30/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/30/2008.

Dated: 7/12/2007
Seattle, WA

Asa J. Louis
Forensic Toxicologist

AJL/jr
AJLSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions.

2007 OCT 01

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Brittany Ball, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Biology.

The simulator solution, Lot Number 07016, was prepared in the Washington State Toxicology Laboratory on 5/30/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/30/2008.

Dated: 7/12/2007
Seattle, WA

Brittany Ball
Brittany Ball
Forensic Toxicologist

BB/jr
BBSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions. Brittany Ball 10/1/07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Brian Capron, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Biology and nine years of experience in forensic toxicology.

The simulator solution, Lot Number 07016, was prepared in the Washington State Toxicology Laboratory on 5/30/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/30/2008.

Dated: 7/12/2007
Seattle, WA

Brian Capron
Forensic Toxicologist

BC/jr
BCSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions.

Brian Capron 10.1.07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Edward J. Formoso, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Chemistry and twenty-nine years of experience in the Washington State Toxicology Laboratory.

The simulator solution, Lot Number 07016, was prepared in the Washington State Toxicology Laboratory on 5/30/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/30/2008.

Dated: 7/12/2007
Seattle, WA

Edward J. Formoso
Forensic Toxicologist

EJF/jr
EFSIMSOL

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Kelly D. Gross, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: B.S. degree in Chemistry and fifteen years of forensic laboratory experience.

The simulator solution, Lot Number 07016, was prepared in the Washington State Toxicology Laboratory on 5/30/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/30/2008.

Dated: 7/12/2007
Seattle, WA

Kelly D. Gross
Forensic Toxicologist

KDG/jr
KDGSIMSOL

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Melissa L. Pemberton, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: Bachelors degree in Microbiology and over fifteen years of experience as a forensic toxicologist.

The simulator solution, Lot Number 07016, was prepared in the Washington State Toxicology Laboratory on 5/30/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/30/2008.

Dated: 7/12/2007
Seattle, WA

Melissa L. Pemberton
Forensic Toxicologist

MP/jr
MPSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions.

10-01-08

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Naziha Nuwayhid, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: Bachelor and Masters Degrees in Biology, Ph.D. degree in Basic Medical Science, ten years experience in clinical laboratory sciences, one year in clinical toxicology and six years in forensic toxicology. I am also board certified by the American Board of Clinical Chemistry.

The simulator solution, Lot Number 07016, was prepared in the Washington State Toxicology Laboratory on 5/30/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/30/2008.

Dated: 7/12/2007
Seattle, WA

Naziha Nuwayhid, Ph.D.
Forensic Toxicologist

NN/jr
NNSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions.

9-28-07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Rebecca Flaherty, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degrees in Biochemistry and Psychobiology and MS degree in Forensic Science.

The simulator solution, Lot Number 07016, was prepared in the Washington State Toxicology Laboratory on 5/30/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/30/2008.

Dated: 7/12/2007
Seattle, WA

Rebecca Flaherty
Forensic Toxicologist

RF/jr
RFSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions. 09/28/07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Lisa Noble, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Biochemistry and two years laboratory experience in formulation chemistry.

The simulator solution, Lot Number 07016, was prepared in the Washington State Toxicology Laboratory on 5/30/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/30/2008.

Dated: 7/12/2007
Seattle, WA

Lisa Noble
Forensic Toxicologist

LRN/jr
LPSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions. Lisa Noble 9/28/07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Amanda M. Black, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degrees in Chemistry and Veterinary Science.

The simulator solution, Lot Number 07016, was prepared in the Washington State Toxicology Laboratory on 5/30/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/30/2008.

Dated: 7/12/2007
Seattle, WA

Amanda M. Black
Forensic Toxicologist

AB/jr
ABSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions. 9-28-07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Brianna Peterson, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Chemistry, MS degree in Forensic Science, Ph.D. degree in Toxicology, and two years of experience in forensic toxicology.

The simulator solution, Lot Number 07016, was prepared in the Washington State Toxicology Laboratory on 5/30/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/30/2008.

Dated: 7/12/2007
Seattle, WA

Brianna Peterson
Brianna Peterson
Forensic Toxicologist

BP/jr
BPSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions.

Brianna Peterson 10/1/07

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION

I, Brianne E. Akins, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Biology.

The simulator solution, Lot Number 07016, was prepared in the Washington State Toxicology Laboratory on 5/30/2007. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/30/2008.

Dated: 7/12/2007
Seattle, WA

Brianne E. Akins
Brianne E. Akins
Forensic Toxicologist

BEA/jr
BLSIMSOL

A review of solution batch records was recently completed. After this review, I checked the file for this solution and reviewed all changes that were made. I found that the solution still conformed to those standards established by the State Toxicologist for the certification of simulator solutions.
Brianne E. Akins 10/1/07

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/1/2007 4:55:20 PM
Instrument 4
DB-ALC1

SIM 07016
Chris Johnston

vial # 1



#	Compound	Area	RT
1	Ethanol	687	1.007
2	n-Propanol	1624	1.663

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/1/2007 4:58:37 PM
Instrument 4
DB-ALC1

SIM 07016
Chris Johnston

vial # 2



#	Compound	Area	RT
1	Ethanol	673	1.007
2	n-Propanol	1604	1.662

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/1/2007 5:01:52 PM
Instrument 4
DB-ALC1

SIM 07016
Chris Johnston

vial # 3



#	Compound	Area	RT
1	Ethanol	689	1.007
2	n-Propanol	1630	1.663

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/1/2007 5:05:05 PM
Instrument 4
DB-ALC1

SIM 07016
Chris Johnston

vial # 4



#	Compound	Area	RT
1	Ethanol	700	1.007
2	n-Propanol	1646	1.663

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/1/2007 5:08:16 PM
Instrument 4
DB-ALC1

SIM 07016
Chris Johnston

vial # 5



#	Compound	Area	RT
1	Ethanol	679	1.007
2	n-Propanol	1619	1.662

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/1/2007 5:11:36 PM
Instrument 4
DB-ALC1

0.10ContolCJ
Chris Johnston

vial # 6



#	Compound	Area	RT
1	Ethanol	673	1.007
2	n-Propanol	1629	1.662

Totals:



Ethanol 0.098 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/1/2007 5:17:00 PM
Instrument 4
DB-ALC1

Blank
Chris Johnston

vial # 7



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1718	1.662

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/30/2007 3:20:08 PM
Instrument 5
DB-ALC2

SIM 07016
Justin Knoy

vial # 56



#	Compound	Area	RT
1	ethanol	777	1.111
2	n-propanol	1630	1.948

Totals:



ethanol 0.099 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/30/2007 3:23:42 PM
Instrument 5
DB-ALC2

SIM 07016
Justin Knoy

vial # 57



#	Compound	Area	RT
1	ethanol	796	1.111
2	n-propanol	1647	1.948

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/30/2007 3:27:15 PM
Instrument 5
DB-ALC2

SIM 07016
Justin Knoy

vial # 58



#	Compound	Area	RT
1	ethanol	831	1.111
2	n-propanol	1688	1.948

Totals:



ethanol 0.102 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/30/2007 3:30:49 PM
Instrument 5
DB-ALC2

SIM 07016
Justin Knoy

vial # 59



#	Compound	Area	RT
1	ethanol	852	1.111
2	n-propanol	1726	1.949

Totals:



ethanol 0.102 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/30/2007 3:34:23 PM
Instrument 5
DB-ALC2

SIM 07016
Justin Knoy

vial # 60



#	Compound	Area	RT
1	ethanol	854	1.111
2	n-propanol	1752	1.949

Totals:



ethanol 0.101 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/30/2007 3:47:14 PM
Instrument 5
DB-ALC2

0.10 CTRL JK
Justin Knoy

vial # 61



#	Compound	Area	RT
1	ethanol	806	1.111
2	n-propanol	1722	1.948

Totals:



ethanol 0.097 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
5/30/2007 3:50:42 PM
Instrument 5
DB-ALC2

blank
Justin Knoy

vial # 62



#	Compound	Area	RT
1	ethanol	0	0.000
2	n-propanol	1740	1.948

Totals:



ethanol 0.000 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/30/2007 6:46:40 PM
Instrument 4
DB-ALC1

07016-1
Sarah Swenson

vial # 53



#	Compound	Area	RT
1	Ethanol	691	1.004
2	n-Propanol	1477	1.658

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/30/2007 6:50:01 PM
Instrument 4
DB-ALC1

07016-2
Sarah Swenson

vial # 54



#	Compound	Area	RT
1	Ethanol	704	1.004
2	n-Propanol	1506	1.658

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/30/2007 6:53:19 PM
Instrument 4
DB-ALC1

07016-3
Sarah Swenson

vial # 55



#	Compound	Area	RT
1	Ethanol	706	1.004
2	n-Propanol	1516	1.658

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/30/2007 6:56:37 PM
Instrument 4
DB-ALC1

07016-4
Sarah Swenson

vial # 56



#	Compound	Area	RT
1	Ethanol	714	1.004
2	n-Propanol	1525	1.658

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/30/2007 6:59:55 PM
Instrument 4
DB-ALC1

07016-5
Sarah Swenson

vial # 57



#	Compound	Area	RT
1	Ethanol	700	1.005
2	n-Propanol	1501	1.658

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/30/2007 7:05:21 PM
Instrument 4
DB-ALC1

0.10 CTL-SS
Sarah Swenson

vial # 58



#	Compound	Area	RT
1	Ethanol	707	1.004
2	n-Propanol	1550	1.658

Totals:



Ethanol 0.098 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/30/2007 7:08:43 PM
Instrument 4
DB-ALC1

BLANK
Sarah Swenson

vial # 59



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1618	1.658

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/31/2007 3:35:46 PM
Instrument 4
DB-ALC1

Sim Sol 07016-1
Estuardo J. Miranda

vial # 53



#	Compound	Area	RT
1	Ethanol	702	1.007
2	n-Propanol	1580	1.662

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/31/2007 3:39:07 PM
Instrument 4
DB-ALC1

Sim Sol 07016-2
Estuardo J. Miranda

vial # 54



#	Compound	Area	RT
1	Ethanol	720	1.007
2	n-Propanol	1596	1.663

Totals:



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/31/2007 3:44:21 PM
Instrument 4
DB-ALC1

Sim Sol 07016-3
Estuardo J. Miranda

vial # 55



#	Compound	Area	RT
1	Ethanol	749	1.007
2	n-Propanol	1688	1.663

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/31/2007 3:47:46 PM
Instrument 4
DB-ALC1

Sim Sol 07016-4
Estuardo J. Miranda

vial # 56



#	Compound	Area	RT
1	Ethanol	704	1.007
2	n-Propanol	1566	1.663

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/31/2007 3:51:13 PM
Instrument 4
DB-ALC1

Sim Sol 07016-5
Estuardo J. Miranda

vial # 57



#	Compound	Area	RT
1	Ethanol	718	1.007
2	n-Propanol	1599	1.662

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/31/2007 3:54:25 PM
Instrument 4
DB-ALC1

0.100 EM Control
Estuardo J. Miranda

vial # 58



#	Compound	Area	RT
1	Ethanol	695	1.006
2	n-Propanol	1601	1.662

Totals:



Ethanol 0.098 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
5/31/2007 3:57:46 PM
Instrument 4
DB-ALC1

BLANK
Estuardo J. Miranda

vial # 59



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1628	1.662

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/1/2007 3:23:08 PM
Instrument 4
DB-ALC1

07016a
alouis

vial # 51



#	Compound	Area	RT
1	Ethanol	715	1.008
2	n-Propanol	1636	1.664

Totals:



Ethanol 0.103 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/1/2007 3:26:25 PM 07016b
Instrument 4 alouis
DB-ALC1

vial # 52



#	Compound	Area	RT
1	Ethanol	690	1.008
2	n-Propanol	1630	1.664

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/1/2007 3:29:43 PM
Instrument 4
DB-ALC1

07016c
alouis

vial # 53



#	Compound	Area	RT
1	Ethanol	713	1.008
2	n-Propanol	1674	1.665

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/1/2007 3:33:00 PM
Instrument 4
DB-ALC1

07016d
alouis

vial # 54



#	Compound	Area	RT
1	Ethanol	731	1.008
2	n-Propanol	1720	1.664

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/1/2007 3:36:15 PM
Instrument 4
DB-ALC1

07016e
alouis

vial # 55



#	Compound	Area	RT
1	Ethanol	733	1.008
2	n-Propanol	1735	1.664

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/1/2007 12:44:43 PM
Instrument 4
DB-ALC1

0.10 con al
alouis

vial # 10



#	Compound	Area	RT
1	Ethanol	698	1.009
2	n-Propanol	1664	1.666

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

Sequence Parameters:

Operator:	BBALL
Data File Naming:	Prefix/Counter
Signal 1 Prefix:	SIG1
Counter:	0001
Signal 2 Prefix:	SIG2
Counter:	0001
Data Directory:	C:\HPCHEM\2\DATA\
Data Subdirectory:	070604BB
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	BLDALCO3	1	Sample		
2	Vial 2	0.079 CAL	BLDALCO3	1	Calib		
3	Vial 3	0.158 CAL	BLDALCO3	1	Calib		
4	Vial 4	0.316 CAL	BLDALCO3	1	Calib		
5	Vial 5	BLANK	BLDALCO3	1	Sample		
6	Vial 6	0.04 CONTROL-BB	BLDALCO3	1	Ctrl Samp		
7	Vial 7	0.10 CONTROL-BB	BLDALCO3	1	Ctrl Samp		
8	Vial 8	0.20 CONTROL-BB	BLDALCO3	1	Ctrl Samp		
9	Vial 9	SIM 016A	BLDALCO3	1	Sample		
10	Vial 10	SIM 016B	BLDALCO3	1	Sample		
11	Vial 11	SIM 016C	BLDALCO3	1	Sample		
12	Vial 12	SIM 016D	BLDALCO3	1	Sample		
13	Vial 13	SIM 016E	BLDALCO3	1	Sample		
14	Vial 14	.10CONTROL-BB	BLDALCO3	1	Sample		
15	Vial 15	BLANK	BLDALCO3	1	Sample		

Sequence Table (Back Injector):

No entries - empty table!

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/4/2007 1:09:44 PM
Instrument 3
db-alc2

BLANK
BBALL

vial # 1



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1601	1.821

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/4/2007 1:12:52 PM
Instrument 3
db-alc2

0.079 CAL
BBALL

vial # 2



#	Compound	Area	RT
1	ETHANOL	568	1.061
2	n-PROPANOL	1613	1.822

Totals:



ETHANOL 0.079 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/4/2007 1:15:59 PM
Instrument 3
db-alc2

0.158 CAL
BBALL

vial # 3



#	Compound	Area	RT
1	ETHANOL	1117	1.061
2	n-PROPANOL	1590	1.821

Totals:



ETHANOL 0.157 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/4/2007 1:19:06 PM
Instrument 3
db-alc2

0.316 CAL
BBALL

vial # 4



#	Compound	Area	RT
1	ETHANOL	2281	1.062
2	n-PROPANOL	1598	1.822

Totals:



ETHANOL 0.317 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/4/2007 1:22:13 PM
Instrument 3
db-alc2

BLANK
BBALL

vial # 5



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1586	1.821

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/4/2007 1:25:20 PM
Instrument 3
db-alc2

0.04 CONTROL-BB
BBALL

vial # 6



#	Compound	Area	RT
1	ETHANOL	279	1.061
2	n-PROPANOL	1580	1.821

Totals:



ETHANOL 0.040 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/4/2007 1:28:27 PM
Instrument 3
db-alc2

0.10 CONTROL-BB
BBALL

vial # 7



#	Compound	Area	RT
1	ETHANOL	695	1.062
2	n-PROPANOL	1578	1.822

Totals:



ETHANOL 0.098 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/4/2007 1:31:34 PM
Instrument 3
db-alc2

0.20 CONTROL-BB
BBALL

vial # 8



#	Compound	Area	RT
1	ETHANOL	1388	1.062
2	n-PROPANOL	1584	1.821

Totals:



ETHANOL 0.195 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/4/2007 1:34:41 PM
Instrument 3
db-alc2

SIM 016A
BBALL

vial # 9



#	Compound	Area	RT
1	ETHANOL	717	1.061
2	n-PROPANOL	1583	1.821

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/4/2007 1:37:49 PM
Instrument 3
db-alc2

SIM 016B
BBALL

vial # 10



#	Compound	Area	RT
1	ETHANOL	713	1.062
2	n-PROPANOL	1580	1.822

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/4/2007 1:40:56 PM
Instrument 3
db-alc2

SIM 016C
BBALL

vial # 11



#	Compound	Area	RT
1	ETHANOL	716	1.061
2	n-PROPANOL	1587	1.820

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/4/2007 1:44:03 PM
Instrument 3
db-alc2

SIM 016D
BBALL

vial # 12



#	Compound	Area	RT
1	ETHANOL	717	1.062
2	n-PROPANOL	1585	1.822

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/4/2007 1:47:10 PM
Instrument 3
db-alc2

SIM 016E
BBALL

vial # 13



#	Compound	Area	RT
1	ETHANOL	717	1.061
2	n-PROPANOL	1593	1.821

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/4/2007 1:50:17 PM
Instrument 3
db-alc2

.10CONTROL-BB
BBALL

vial # 14



#	Compound	Area	RT
1	ETHANOL	723	1.062
2	n-PROPANOL	1646	1.822

Totals:



ETHANOL 0.098 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/4/2007 1:53:24 PM
Instrument 3
db-alc2

BLANK
BBALL

vial # 15



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1579	1.821

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/4/2007 2:00:24 PM
Instrument 4
DB-ALC1

0.10 control bc
bcapron

vial # 48



#	Compound	Area	RT
1	Ethanol	727	1.010
2	n-Propanol	1755	1.668

Totals:



Ethanol 0.097 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/4/2007 2:03:45 PM
Instrument 4
DB-ALC1

blank
bcapron

vial # 49



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1677	1.668

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/4/2007 2:07:05 PM
Instrument 4
DB-ALC1

07016
bcapron

vial # 50



#	Compound	Area	RT
1	Ethanol	702	1.010
2	n-Propanol	1664	1.668

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/4/2007 2:10:26 PM
Instrument 4
DB-ALC1

07016
bcapron

vial # 51



#	Compound	Area	RT
1	Ethanol	718	1.010
2	n-Propanol	1701	1.668

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/4/2007 2:13:44 PM
Instrument 4
DB-ALC1

07016
bcapron

vial # 52



#	Compound	Area	RT
1	Ethanol	735	1.010
2	n-Propanol	1753	1.668

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/4/2007 2:17:02 PM
Instrument 4
DB-ALC1

07016
bcapron

vial # 53



#	Compound	Area	RT
1	Ethanol	718	1.010
2	n-Propanol	1693	1.668

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/4/2007 2:25:34 PM
Instrument 4
DB-ALC1

07016
bcapron

vial # 54



#	Compound	Area	RT
1	Ethanol	792	1.010
2	n-Propanol	1852	1.668

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 10:30:30 AM
Instrument 1
DB ALC 1

07016
ED FORMOSO

vial # 15



#	Compound	Area	RT
1	Ethanol	1323	1.076
2	n-Propanol	3084	1.767
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 10:33:35 AM
Instrument 1
DB ALC 1

07016
ED FORMOSO

vial # 16



#	Compound	Area	RT
1	Ethanol	1324	1.076
2	n-Propanol	3096	1.767
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 10:36:40 AM
Instrument 1
DB ALC 1

07016
ED FORMOSO

vial # 17



#	Compound	Area	RT
1	Ethanol	1325	1.076
2	n-Propanol	3094	1.766
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 10:39:45 AM
Instrument 1
DB ALC 1

07016
ED FORMOSO

vial # 18



#	Compound	Area	RT
1	Ethanol	1338	1.076
2	n-Propanol	3136	1.767
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 10:42:50 AM
Instrument 1
DB ALC 1

07016
ED FORMOSO

vial # 19



#	Compound	Area	RT
1	Ethanol	1315	1.076
2	n-Propanol	3078	1.767
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 10:45:54 AM
Instrument 1
DB ALC 1

0.10 CONTROL
ED FORMOSO

vial # 20



#	Compound	Area	RT
1	Ethanol	1284	1.076
2	n-Propanol	3070	1.766
Tot			



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 3:51:01 PM
Instrument 1
DB ALC 1

0.10 ctl kdg
kgross

vial # 20



#	Compound	Area	RT
1	Ethanol	1250	1.076
2	n-Propanol	2960	1.766
Tot			



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 3:54:05 PM
Instrument 1
DB ALC 1

blank
kgross

vial # 21



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2928	1.766
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 3:57:10 PM
Instrument 1
DB ALC 1

07016 sim
kgross

vial # 22



#	Compound	Area	RT
1	Ethanol	1298	1.076
2	n-Propanol	3001	1.766
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 4:00:15 PM
Instrument 1
DB ALC 1

07016 sim
kgross

vial # 23



#	Compound	Area	RT
1	Ethanol	1283	1.076
2	n-Propanol	2951	1.766
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 4:03:20 PM
Instrument 1
DB ALC 1

07016 sim
kgross

vial # 24



#	Compound	Area	RT
1	Ethanol	1295	1.076
2	n-Propanol	2987	1.766
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 4:06:25 PM
Instrument 1
DB ALC 1

07016 sim
kgross

vial # 25



#	Compound	Area	RT
1	Ethanol	1328	1.075
2	n-Propanol	3070	1.766
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 4:09:29 PM
Instrument 1
DB ALC 1

07016 sim
kgross

vial # 26



#	Compound	Area	RT
1	Ethanol	1326	1.076
2	n-Propanol	3054	1.767
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/6/2007 10:20:42 AM
Instrument 5
DB-ALC2

07016 sim soln
m pemberton

vial # 13



#	Compound	Area	RT
1	ethanol	868	1.113
2	n-propanol	1940	1.951

Totals:



ethanol 0.101 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/6/2007 10:24:14 AM
Instrument 5
DB-ALC2

07016 sim soln
m pemberton

vial # 14



#	Compound	Area	RT
1	ethanol	895	1.118
2	n-propanol	1944	1.957

Totals:



ethanol 0.104 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/6/2007 10:27:45 AM
Instrument 5
DB-ALC2

07016 sim soln
m pemberton

vial # 15



#	Compound	Area	RT
1	ethanol	816	1.112
2	n-propanol	1854	1.951

Totals:



ethanol 0.099 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/6/2007 10:31:11 AM
Instrument 5
DB-ALC2

07016 sim soln
m pemberton

vial # 16



#	Compound	Area	RT
1	ethanol	808	1.113
2	n-propanol	1823	1.951

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/6/2007 10:34:35 AM
Instrument 5
DB-ALC2

07016 sim soln
m pemberton

vial # 17



#	Compound	Area	RT
1	ethanol	832	1.113
2	n-propanol	1871	1.952

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/6/2007 10:38:09 AM
Instrument 5
DB-ALC2

0.10 controlmp
m pemberton

vial # 18



#	Compound	Area	RT
1	ethanol	848	1.115
2	n-propanol	1931	1.953

Totals:



ethanol 0.099 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/6/2007 10:49:53 AM
Instrument 5
DB-ALC2

blk
m pemberton

vial # 19



#	Compound	Area	RT
1	ethanol	0	0.000
2	n-propanol	2013	1.953

Totals:



ethanol 0.000 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 3:29:27 PM
Instrument 1
DB ALC 1

0.10 ctl nn
kgross

vial # 13



#	Compound	Area	RT
1	Ethanol	1249	1.076
2	n-Propanol	2999	1.767
Tot			



Ethanol 0.097 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 3:32:32 PM
Instrument 1
DB ALC 1

blank
kgross

vial # 14



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	3080	1.766
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 3:35:37 PM
Instrument 1
DB ALC 1

07016 sim nn
kgross

vial # 15



#	Compound	Area	RT
1	Ethanol	1326	1.076
2	n-Propanol	3061	1.766
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 3:38:42 PM
Instrument 1
DB ALC 1

07016 sim nn
kgross

vial # 16



#	Compound	Area	RT
1	Ethanol	1292	1.076
2	n-Propanol	3025	1.766
Tot			



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 3:41:47 PM
Instrument 1
DB ALC 1

07016 sim nn
kgross

vial # 17



#	Compound	Area	RT
1	Ethanol	1284	1.076
2	n-Propanol	3020	1.766
Tot			



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 3:44:51 PM
Instrument 1
DB ALC 1

07016 sim nn
kgross

vial # 18



#	Compound	Area	RT
1	Ethanol	1293	1.076
2	n-Propanol	3032	1.767
Tot			



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/5/2007 3:47:56 PM
Instrument 1
DB ALC 1

07016 sim nn
kgross

vial # 19



#	Compound	Area	RT
1	Ethanol	1288	1.076
2	n-Propanol	3021	1.767
Tot			



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/8/2007 8:14:11 AM
Instrument 1
DB ALC 1

07016-1 rf
R Flaherty

vial # 24



#	Compound	Area	RT
1	Ethanol	1412	1.086
2	n-Propanol	3091	1.779
Tot			



Ethanol 0.104 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/8/2007 8:17:15 AM
Instrument 1
DB ALC 1

07016-2 rf
R Flaherty

vial # 25



#	Compound	Area	RT
1	Ethanol	1354	1.081
2	n-Propanol	3022	1.774
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/8/2007 8:20:20 AM
Instrument 1
DB ALC 1

07016-3 rf
R Flaherty

vial # 26



#	Compound	Area	RT
1	Ethanol	1322	1.079
2	n-Propanol	2924	1.771
Tot			



Ethanol 0.103 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/8/2007 8:23:25 AM
Instrument 1
DB ALC 1

07016-4 rf
R Flaherty

vial # 27



#	Compound	Area	RT
1	Ethanol	1394	1.079
2	n-Propanol	3086	1.772
Tot			



Ethanol 0.103 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/8/2007 8:26:30 AM
Instrument 1
DB ALC 1

07016-5 rf
R Flaherty

vial # 28



#	Compound	Area	RT
1	Ethanol	1396	1.077
2	n-Propanol	3130	1.768
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/8/2007 8:29:35 AM
Instrument 1
DB ALC 1

0.1 con RF
R Flaherty

vial # 29



#	Compound	Area	RT
1	Ethanol	1389	1.077
2	n-Propanol	3183	1.768
Tot			



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/8/2007 8:32:39 AM
Instrument 1
DB ALC 1

blank
R Flaherty

vial # 30



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	3090	1.770
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/8/2007 10:02:22 AM
Instrument 5
DB-ALC2

07016-1
Lisa Noble

vial # 14



#	Compound	Area	RT
1	ethanol	764	1.112
2	n-propanol	1740	1.951

Totals:



ethanol 0.098 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/8/2007 10:05:53 AM
Instrument 5
DB-ALC2

07016-2
Lisa Noble

vial # 15



#	Compound	Area	RT
1	ethanol	785	1.113
2	n-propanol	1742	1.952

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/8/2007 10:09:20 AM
Instrument 5
DB-ALC2

07016-3
Lisa Noble

vial # 16



#	Compound	Area	RT
1	ethanol	785	1.112
2	n-propanol	1765	1.951

Totals:



ethanol 0.099 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/8/2007 10:12:47 AM
Instrument 5
DB-ALC2

07016-4
Lisa Noble

vial # 17



#	Compound	Area	RT
1	ethanol	803	1.113
2	n-propanol	1800	1.952

Totals:



ethanol 0.099 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/8/2007 10:16:12 AM
Instrument 5
DB-ALC2

07016-5
Lisa Noble

vial # 18



#	Compound	Area	RT
1	ethanol	806	1.113
2	n-propanol	1814	1.952

Totals:



ethanol 0.099 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/8/2007 10:28:24 AM
Instrument 5
DB-ALC2

BLANK
Lisa Noble

vial # 19



#	Compound	Area	RT
1	ethanol	0	0.000
2	n-propanol	1844	1.952

Totals:



ethanol 0.000 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/8/2007 10:31:48 AM
Instrument 5
DB-ALC2

0.10 CONTROL LN
Lisa Noble

vial # 20



#	Compound	Area	RT
1	ethanol	803	1.113
2	n-propanol	1846	1.952

Totals:



ethanol 0.097 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 11:37:29 AM
Instrument 4
DB-ALC1

SIM 07016-1
Amanda Black

vial # 43



#	Compound	Area	RT
1	Ethanol	677	1.005
2	n-Propanol	1551	1.660

Totals:



Ethanol 0.098 g/100ml



n-Propanol 1.000 g/100ml

CALIBRATION DATA FILED
WITH ST 0704411

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 11:40:48 AM
Instrument 4
DB-ALC1

SIM 07016-2
Amanda Black

vial # 44



#	Compound	Area	RT
1	Ethanol	692	1.005
2	n-Propanol	1539	1.660

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 11:44:05 AM
Instrument 4
DB-ALC1

SIM 07016-3
Amanda Black

vial # 45



#	Compound	Area	RT
1	Ethanol	701	1.005
2	n-Propanol	1554	1.660

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 11:47:22 AM
Instrument 4
DB-ALC1

SIM 07016-4
Amanda Black

vial # 46



#	Compound	Area	RT
1	Ethanol	715	1.005
2	n-Propanol	1580	1.660

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 11:50:38 AM
Instrument 4
DB-ALC1

SIM 07016-5
Amanda Black

vial # 47



#	Compound	Area	RT
1	Ethanol	700	1.005
2	n-Propanol	1545	1.660

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 11:56:05 AM
Instrument 4
DB-ALC1

0.10 AB Control
Amanda Black

vial # 48



#	Compound	Area	RT
1	Ethanol	714	1.005
2	n-Propanol	1617	1.660

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 11:59:24 AM
Instrument 4
DB-ALC1

BLANK
Amanda Black

vial # 49



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1643	1.660

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

Sequence: C:\HPCHEM\1\SEQUENCE\BPSIM.S

Sequence Parameters:

Operator:	Brianna Peterson
Data File Naming:	Prefix/Counter
Signal 1 Prefix:	SIG1
Counter:	0001
Signal 2 Prefix:	SIG2
Counter:	0001
Data Directory:	C:\HPCHEM\1\DATA\
Data Subdirectory:	070612BP
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	BLDALCO	1	Sample		
2	Vial 2	0.079 CAL	BLDALCO	1	Calib		
3	Vial 3	0.158 CAL	BLDALCO	1	Calib		
4	Vial 4	0.316 CAL	BLDALCO	1	Calib		
5	Vial 5	BLANK	BLDALCO	1	Sample		
6	Vial 6	0.02 STD	BLDALCO	1	Sample		
7	Vial 7	0.04 CTRL BP	BLDALCO	1	Sample		
8	Vial 8	0.20 CTRL BP	BLDALCO	1	Sample		
9	Vial 9	BLANK	BLDALCO	1	Sample		
10	Vial 10	SIM 07016-1	BLDALCO	1	Sample		
11	Vial 11	SIM 07016-2	BLDALCO	1	Sample		
12	Vial 12	SIM 07016-3	BLDALCO	1	Sample		
13	Vial 13	SIM 07016-4	BLDALCO	1	Sample		
14	Vial 14	SIM 07016-5	BLDALCO	1	Sample		
15	Vial 15	0.10 CTRL BP	BLDALCO	1	Sample		
16	Vial 33		BLDALCO	1	Sample		
17	Vial 34		BLDALCO	1	Sample		
18	Vial 35		BLDALCO	1	Sample		
19	Vial 36		BLDALCO	1	Sample		
20	Vial 37		BLDALCO	1	Sample		
21	Vial 38		BLDALCO	1	Sample		
22	Vial 39		BLDALCO	1	Sample		
23	Vial 40		BLDALCO	1	Sample		
24	Vial 41		BLDALCO	1	Sample		
25	Vial 42		BLDALCO	1	Sample		
26	Vial 43		BLDALCO	1	Sample		
27	Vial 44		BLDALCO	1	Sample		

Sequence Table (Back Injector):

No entries - empty table!

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 10:16:48 AM
Instrument 1
DB ALC 1

BLANK
Brianna Peterson

vial # 1



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2790	1.768
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 10:19:49 AM
Instrument 1
DB ALC 1

0.079 CAL
Brianna Peterson

vial # 2



#	Compound	Area	RT
1	Ethanol	1019	1.078
2	n-Propanol	2774	1.769
Tot			



Ethanol 0.079 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 10:22:54 AM
Instrument 1
DB ALC 1

0.158 CAL
Brianna Peterson

vial # 3



#	Compound	Area	RT
1	Ethanol	1995	1.078
2	n-Propanol	2727	1.769
Tot			



Ethanol 0.156 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 10:25:59 AM
Instrument 1
DB ALC 1

0.316 CAL
Brianna Peterson

vial # 4



#	Compound	Area	RT
1	Ethanol	4036	1.078
2	n-Propanol	2760	1.769
Tot			



Ethanol 0.316 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 10:29:03 AM
Instrument 1
DB ALC 1

BLANK
Brianna Peterson

vial # 5



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2721	1.769
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 10:32:08 AM
Instrument 1
DB ALC 1

0.02 STD
Brianna Peterson
vial # 6



#	Compound	Area	RT
1	Ethanol	253	1.078
2	n-Propanol	2722	1.768
Tot			



Ethanol 0.020 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 10:35:13 AM
Instrument 1
DB ALC 1

0.04 CTRL BP
Brianna Peterson

vial # 7



#	Compound	Area	RT
1	Ethanol	493	1.078
2	n-Propanol	2707	1.769
Tot			



Ethanol 0.039 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 10:38:18 AM
Instrument 1
DB ALC 1

0.20 CTRL BP
Brianna Peterson

vial # 8



#	Compound	Area	RT
1	Ethanol	2483	1.077
2	n-Propanol	2732	1.768
Tot			



Ethanol 0.196 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 10:41:23 AM
Instrument 1
DB ALC 1

BLANK
Brianna Peterson

vial # 9



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2764	1.769
Tot			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 10:44:28 AM
Instrument 1
DB ALC 1

SIM 07016-1
Brianna Peterson

vial # 10



#	Compound	Area	RT
1	Ethanol	1289	1.077
2	n-Propanol	2753	1.769
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 10:47:32 AM
Instrument 1
DB ALC 1

SIM 07016-2
Brianna Peterson

vial # 11



#	Compound	Area	RT
1	Ethanol	1289	1.077
2	n-Propanol	2749	1.769
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 10:50:37 AM
Instrument 1
DB ALC 1

SIM 07016-3
Brianna Peterson

vial # 12



#	Compound	Area	RT
1	Ethanol	1279	1.077
2	n-Propanol	2730	1.768
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 10:53:42 AM
Instrument 1
DB ALC 1

SIM 07016-4
Brianna Peterson

vial # 13



#	Compound	Area	RT
1	Ethanol	1300	1.078
2	n-Propanol	2764	1.769
Tot			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 10:56:46 AM
Instrument 1
DB ALC 1

SIM 07016-5
Brianna Peterson

vial # 14



#	Compound	Area	RT
1	Ethanol	1318	1.078
2	n-Propanol	2817	1.769
Tot			



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\1\METHODS\BLDALCO.M
6/12/2007 10:59:51 AM
Instrument 1
DB ALC 1

0.10 CTRL BP
Brianna Peterson

vial # 15



#	Compound	Area	RT
1	Ethanol	1290	1.077
2	n-Propanol	2808	1.769
Tot			



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/13/2007 12:21:23 PM
Instrument 5
DB-ALC2

SIM 07016 A
Brianne E. Akins

vial # 45



#	Compound	Area	RT
1	ethanol	815	1.112
2	n-propanol	1771	1.951

Totals:



ethanol 0.099 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/13/2007 12:24:56 PM
Instrument 5
DB-ALC2

SIM 07016 B
Brianne E. Akins

vial # 46



#	Compound	Area	RT
1	ethanol	847	1.112
2	n-propanol	1834	1.951

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/13/2007 12:28:26 PM
Instrument 5
DB-ALC2

SIM 07016 C
Brianne E. Akins

vial # 47



#	Compound	Area	RT
1	ethanol	865	1.112
2	n-propanol	1870	1.951

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/13/2007 12:31:55 PM
Instrument 5
DB-ALC2

SIM 07016 D
Brianne E. Akins

vial # 48



#	Compound	Area	RT
1	ethanol	885	1.113
2	n-propanol	1907	1.951

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/13/2007 12:35:23 PM
Instrument 5
DB-ALC2

SIM 07016 E
Brianne E. Akins

vial # 49



#	Compound	Area	RT
1	ethanol	885	1.113
2	n-propanol	1909	1.951

Totals:



ethanol 0.100 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/13/2007 12:48:15 PM
Instrument 5
DB-ALC2

0.10 CONTROLL-BA
Brianne E. Akins

vial # 50



#	Compound	Area	RT
1	ethanol	879	1.113
2	n-propanol	1936	1.952

Totals:



ethanol 0.098 g/100ml



n-propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO2.M
6/13/2007 12:51:51 PM
Instrument 5
DB-ALC2

BLANK
Brianne E. Akins

vial # 51



#	Compound	Area	RT
1	ethanol	0	0.000
2	n-propanol	1906	1.952

Totals:



ethanol 0.000 g/100ml



n-propanol 1.000 g/100ml