WASHINGTON STATE TOXICOLOGY LABORATORY
SIMULATOR SOLUTION DATA ENTRY REVIEW

Reviewer/ s: KEN DENTON / ROD GOULDBERG Date: 2-14-2008

Location: TOX LAB SEATTLE Solution Batch Number: 08006

	YES	NO	N/A
Preparation date precedes all analysis dates:	☒	☐	☐
Declarations signed and properly dated:	☒	☐	☐
Data entry corresponds to all chromatograms:	☒	☐	☐
All signatures present on Analysis sheet:	☒	☐	☐
Avg. solution concentration correct?:	☒	☐	☐
Standard deviation correct:	☒	☐	☐
Range correct if applicable:	☒	☐	☐
Equivalent vapor concentration correct?	☒	☐	☐
Blank Chromatograms included in file:	☒	☐	☐
External Control information correct: (lot # present and future date)	☒	☐	☐
Complies with accuracy and precision requirements established by the State Toxicologist:	☒	☐	☐
CV% Correct?	☒	☐	☐
Reviewed for outliers per policy and none found?	☒	☐	☐

Comments:

Reviewer Signature: [signature] Date: 2-14-08

Reviewer Signature: [signature] Date: 2/14/2008

Rev. Date: 2/14/2008

WASHINGTON STATE TOXICOLOGY LABORATORY
FORENSIC LABORATORY SERVICES BUREAU
WASHINGTON STATE PATROL
2203 AIRPORT WAY S, SUITE 360
SEATTLE, WASHINGTON 98134-2027
(206) 262-6100 FAX (206) 262-6145

SIMULATOR EXTERNAL STANDARD SOLUTION DATABASE

Preparation and Certification of 0.08 g/210L Simulator External Standard Solution

Batch number **08006** Date prepared: 01/23/2008

Preparation: **66.5** mL of absolute ethyl alcohol diluted to **52** Liters with water

Concentration of ethanol (g/100mL) measured by gas chromatography:

	Analyst 1	Analyst 2	Analyst 3	Analyst 4	Analyst 5	Analyst 6	Analyst 7	Analyst 8	Analyst 9	Analyst 10
1	0.102	0.101	0.101	0.102	0.102	0.107	0.101	0.104		
2	0.100	0.102	0.101	0.103	0.105	0.104	0.101	0.104		
3	0.102	0.102	0.102	0.102	0.103	0.104	0.101	0.105		
4	0.102	0.102	0.102	0.102	0.103	0.104	0.104	0.104		
5	0.101	0.101	0.102	0.103	0.103	0.104	0.101	0.105		
Ctrl	0.099	0.099	0.099	0.100	0.100	0.101	0.101	0.103		

Statistics:

Avg. solution concent.: 0.1026 g/100 mL

SD: 0.00150

Range (3.3XSD): 0.0977 to 0.1075

Precision CV (%): 1.4607 %

External Control:

Lot #: A050528 Exp date: MM 07 / YYYY 2011

Target concentration: 0.10 g/100mL

Equivalent vapor concent.: 0.0834 g/210L

Analyst	Name	Signature	Date Tested
1	Brittany Ball	Brittany Ball	01/24/2008
2	Rebecca Flaherty	[illegible]	01/25/2008
3	Asa Louis	[illegible]	01/30/2008
4	Justin L Knoy	[illegible]	02/07/2008
5	Estuardo J. Miranda	[illegible]	02/07/2008
6	Brianna Peterson	Brianna Peterson	02/08/2008
7	Christie Mitchell	Christie Mitchell	02/11/2008
8	Amanda Black	[illegible]	02/06/2008
9			
10			

Prepared by: Brittany Ball according to the approved protocol.

Final review by: [illegible]

ver 12/19/2007

Batch Worksheet Check Off

Please check the data entered into the worksheet is correct and that the date to the right of your name is the date that you tested the solution and then sign the worksheet.

Please initial below to affirm that you have:

1 – Initialed your chromatograms
2 – Checked your data
3 – Checked the date to the right of your name on the worksheet
4 – Signed the worksheet.

	Initials	Date
Amanda Black	OB	2-12-08
Asa Louis	[illegible]	20080212
Brian Capron		
Brianna Peterson	BP	2/12/08
Brianne Akins		
Brittany Ball	BB	2/12/08
Christie Mitchell	CM	2/12/2008
Christopher Johnston		
Erin Kolbrich		
Estuardo Miranda	EM	2-12-08
Gwynyth Scherperel		
Justin Knoy	[illegible]	2-12-08
Lisa Noble		
Melissa Pemberton		
Naziha Nuwayhid		
Rebecca Flaherty	RF	2/12/2008
Sarah Swenson		

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION FOR LOT 08006

I, Brittany Ball, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Biology.

The simulator solution, Lot Number 08006, was prepared in the Washington State Toxicology Laboratory on 1/23/2008. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 1/23/2009

Seattle, WA

Brittany Ball 2/12/08
Brittany Ball Date
Forensic Toxicologist

BB/jr
BBSIMSOL

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION FOR LOT 08006

I, Rebecca Flaherty, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degrees in Biochemistry and Psychobiology and MS degree in Forensic Science.

The simulator solution, Lot Number 08006, was prepared in the Washington State Toxicology Laboratory on 1/23/2008. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 1/23/2009.

Seattle, WA

02/12/08

Rebecca Flaherty Date
Forensic Toxicologist

RF/jr
RFSIMSOL

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION FOR LOT 08006

I, Asa J. Louis, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: B.S. degree in Biochemistry and eight years of toxicology experience.

The simulator solution, Lot Number 08006, was prepared in the Washington State Toxicology Laboratory on 1/23/2008. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 1/23/2009.

Seattle, WA

20080212

Asa J. Louis Date
Forensic Toxicologist

AJL/jr
AJLSIMSOL

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION FOR LOT 08006

I, Justin L. Knoy, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Biology, and MS degree in Forensic Science.

The simulator solution, Lot Number 08006, was prepared in the Washington State Toxicology Laboratory on 1/23/2008. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 1/23/2009.

Seattle, WA

2-12-08

Justin L. Knoy Date
Forensic Toxicologist

JLK/jr
JKSIMSOL

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION FOR LOT 08006

I, Estuardo J. Miranda, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: Bachelor of Science in Chemistry, Master of Science in Zoology and nine years experience in Forensic Toxicology.

The simulator solution, Lot Number 08006, was prepared in the Washington State Toxicology Laboratory on 1/23/2008. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 1/23/2009.

Seattle, WA

2-12-2008

Estuardo J. Miranda Date
Forensic Toxicologist

EM/jr
EMSIMSOL

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION FOR LOT 08006

I, Brianna Peterson, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Chemistry, MS degree in Forensic Science, Ph.D. degree in Toxicology, and two years of experience in forensic toxicology.

The simulator solution, Lot Number 08006, was prepared in the Washington State Toxicology Laboratory on 1/23/2008. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 1/23/2009.

Seattle, WA

Brianna Peterson 2/12/08
Brianna Peterson, Ph.D. Date
Forensic Toxicologist

BP/jr
BPSIMSOL

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION FOR LOT 08006

I, Christie Mitchell, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BA degree in Chemistry and MFS degree in Forensic Science.

The simulator solution, Lot Number 08006, was prepared in the Washington State Toxicology Laboratory on 1/23/2008. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 1/23/2009.

Seattle, WA

Christie Mitchell 3/12/2008
Christie Mitchell Date
Forensic Toxicologist

CM/jr
CMSIMSOL

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION FOR LOT 08006

I, Amanda M. Black, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degrees in Chemistry and Veterinary Science.

The simulator solution, Lot Number 08006, was prepared in the Washington State Toxicology Laboratory on 1/23/2008. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 1/23/2009.

Seattle, WA

2-12-08

Amanda M. Black Date
Forensic Toxicologist

AB/jr
ABSIMSOL

Sequence Parameters:

Operator:	BRITTANY BALL
Data File Naming:	Auto
Data Directory:	D:\HPCHEM\1\DATA\
Data Subdirectory:	080124BB
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	BLDALCO	1	Sample		
2	Vial 2	0.079 CAL	BLDALCO	1	Calib		
3	Vial 3	0.158 CAL	BLDALCO	1	Calib		
4	Vial 4	0.316 CAL	BLDALCO	1	Calib		
5	Vial 5	BLANK	BLDALCO	1	Ctrl Samp		
6	Vial 6	0.04 MIX	VOL	1	Calib		
7	Vial 7	0.08 MIX	VOL	1	Calib		
8	Vial 8	0.02 STD	BLDALCO	1	Sample		
9	Vial 9	BLANK	BLDALCO	1	Sample		
10	Vial 10	0.04 CONTROL-BB	BLDALCO	1	Ctrl Samp	A050530	Exp 7/2011
11	Vial 11	0.10 CONTROL-BB	BLDALCO	1	Ctrl Samp	A050528	Exp 7/2011
12	Vial 12	0.20 CONTROL-BB	BLDALCO	1	Ctrl Samp	A050527	Exp 7/2011
13	Vial 13	BLANK	BLDALCO	1	Sample		
14	Vial 14	SIM 08006-1 BB	BLDALCO	1	Sample		
15	Vial 15	SIM 08006-2 BB	BLDALCO	1	Sample		
16	Vial 16	SIM 08006-3 BB	BLDALCO	1	Sample		
17	Vial 17	SIM 08006-4 BB	BLDALCO	1	Sample		
18	Vial 18	SIM 08006-5 BB	BLDALCO	1	Sample		
19	Vial 19	0.10 CONTROL-BB	BLDALCO	1	Ctrl Samp		
20	Vial 20	BLANK	BLDALCO	1	Sample		

Sequence Table (Back Injector):

No entries - empty table!

BB 1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 3:11:18 PM
Instrument 4
DB-ALC1

BLANK
BRITTANY BALL

vial # 1



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1575	1.644

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

BB 1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 3:14:35 PM
Instrument 4
DB-ALC1

0.079 CAL
BRITTANY BALL

vial # 2



#	Compound	Area	RT
1	Ethanol	595	0.996
2	n-Propanol	1769	1.644

Totals:



Ethanol 0.081 g/100ml



n-Propanol 1.000 g/100ml

BB
1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 3:17:47 PM
Instrument 4
DB-ALC1

0.158 CAL
BRITTANY BALL

vial # 3



#	Compound	Area	RT
1	Ethanol	1144	0.996
2	n-Propanol	1714	1.644
Totals:			



Ethanol 0.160 g/100ml



n-Propanol 1.000 g/100ml

BB 1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 3:21:08 PM
Instrument 4
DB-ALC1

0.316 CAL
BRITTANY BALL

vial # 4



#	Compound	Area	RT
1	Ethanol	2613	0.996
2	n-Propanol	1951	1.644

Totals:



Ethanol 0.316 g/100ml



n-Propanol 1.000 g/100ml

BBB
1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 3:24:27 PM
Instrument 4
DB-ALC1

BLANK
BRITTANY BALL

vial # 5



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	2026	1.644

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

BB 1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M
1/24/2008 3:37:03 PM
Instrument 4
DB-ALC1

0.04 MIX
BRITTANY BALL

vial # 6



#	Compound	Area	RT
1	Methanol	161	0.796
2	Ethanol	341	0.998
3	Isopropanol	650	1.218
4	Acetone	1328	1.466
5	n-Propanol	2035	1.645

Totals:



Methanol 40.651 mg/dL



Ethanol 40.407 mg/dL



Isopropanol 39.955 mg/dL

BBB
1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M



Acetone 38.966 mg/dL



n-Propanol 1.000 mg/dL

RBB
1/24/06

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M
1/24/2008 3:40:18 PM
Instrument 4
DB-ALC1

0.08 MIX
BRITTANY BALL

vial # 7



#	Compound	Area	RT
1	Methanol	343	0.796
2	Ethanol	728	0.998
3	Isopropanol	1368	1.218
4	Acetone	2739	1.466
5	n-Propanol	2195	1.645
6		20	2.146

Totals:



Methanol 79.854 mg/dL



Ethanol 79.863 mg/dL



Isopropanol 79.651 mg/dL

BB
1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M



Acetone 79.386 mg/dL



n-Propanol 1.000 mg/dL

RB
1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 3:52:41 PM
Instrument 4
DB-ALC1

0.02 STD
BRITTANY BALL

vial # 8



#	Compound	Area	RT
1	Ethanol	145	0.996
2	n-Propanol	1694	1.644

Totals:



Ethanol 0.020 g/100ml



n-Propanol 1.000 g/100ml

BBB
1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 3:55:59 PM
Instrument 4
DB-ALC1

BLANK
BRITTANY BALL

vial # 9



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1725	1.644

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

BB
1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 3:59:15 PM
Instrument 4
DB-ALC1

0.04 CONTROL-BB
BRITTANY BALL

vial # 10



#	Compound	Area	RT
1	Ethanol	305	0.996
2	n-Propanol	1829	1.644

Totals:



Ethanol 0.039 g/100ml



n-Propanol 1.000 g/100ml

BB
1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 4:02:28 PM
Instrument 4
DB-ALC1

0.10 CONTROL-BB
BRITTANY BALL

vial # 11



#	Compound	Area	RT
1	Ethanol	765	0.996
2	n-Propanol	1820	1.644

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

BB
1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 4:05:50 PM
Instrument 4
DB-ALC1

0.20 CONTROL-BB
BRITTANY BALL

vial # 12



#	Compound	Area	RT
1	Ethanol	1552	0.996
2	n-Propanol	1850	1.644

Totals:



Ethanol 0.198 g/100ml



n-Propanol 1.000 g/100ml

BB
1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 4:09:02 PM
Instrument 4
DB-ALC1

BLANK
BRITTANY BALL

vial # 13



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1833	1.644

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

BB 1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 4:14:21 PM
Instrument 4
DB-ALC1

SIM 08006-1 BB
BRITTANY BALL

vial # 14



#	Compound	Area	RT
1	Ethanol	824	0.996
2	n-Propanol	1906	1.644

Totals:



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

BB 1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 4:17:45 PM
Instrument 4
DB-ALC1

SIM 08006-2 BB
BRITTANY BALL

vial # 15



#	Compound	Area	RT
1	Ethanol	684	0.996
2	n-Propanol	1611	1.644

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

BB 1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 4:21:03 PM
Instrument 4
DB-ALC1

SIM 08006-3 BB
BRITTANY BALL

vial # 16



#	Compound	Area	RT
1	Ethanol	752	0.996
2	n-Propanol	1744	1.644

Totals:



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

BB 1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 4:24:21 PM
Instrument 4
DB-ALC1

SIM 08006-4 BB
BRITTANY BALL

vial # 17



#	Compound	Area	RT
1	Ethanol	782	0.996
2	n-Propanol	1815	1.644

Totals:



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

BB
1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 4:27:39 PM
Instrument 4
DB-ALC1

SIM 08006-5 BB
BRITTANY BALL

vial # 18



#	Compound	Area	RT
1	Ethanol	753	0.996
2	n-Propanol	1752	1.644

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

BB 1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 4:30:55 PM
Instrument 4
DB-ALC1

0.10 CONTROL-BB
BRITTANY BALL

vial # 19



#	Compound	Area	RT
1	Ethanol	742	0.997
2	n-Propanol	1775	1.644

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

BB
1/24/08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/24/2008 4:34:08 PM
Instrument 4
DB-ALC1

BLANK
BRITTANY BALL

vial # 20



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1893	1.643

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

VBB
1/24/08

Sequence Parameters:

Operator:	Rebecca Flaherty
Data File Naming:	Auto
Data Directory:	D:\HPCHEM\1\DATA\
Data Subdirectory:	080125RF
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	BLDALCO	1	Sample		
2	Vial 2	0.079 CAL	BLDALCO	1	Calib		
3	Vial 3	0.158 CAL	BLDALCO	1	Calib		
4	Vial 4	0.316 CAL	BLDALCO	1	Calib		
5	Vial 5	BLANK	BLDALCO	1	Ctrl Samp		
6	Vial 6	0.04 MIX	BLDALCO	1	Sample		
7	Vial 7	0.08 MIX	BLDALCO	1	Sample		
8	Vial 8	0.02 STD	BLDALCO	1	Sample		
9	Vial 9	0.04 CONTROL-RF	BLDALCO	1	Ctrl Samp	LOT# A050530 exp 07/2011	
10	Vial 10	0.10 CONTROL-RF	BLDALCO	1	Ctrl Samp	LOT# A050528 exp 07/2011	
11	Vial 11	0.20 CONTROL-RF	BLDALCO	1	Ctrl Samp	LOT# A050527 exp 07/2011	
12	Vial 12	BLANK	BLDALCO	1	Sample		
13	Vial 13	SIM 08006-1 RF	BLDALCO	1	Sample		
14	Vial 14	SIM 08006-2 RF	BLDALCO	1	Sample		
15	Vial 15	SIM 08006-3 RF	BLDALCO	1	Sample		
16	Vial 16	SIM 08006-4 RF	BLDALCO	1	Sample		
17	Vial 17	SIM 08006-5 RF	BLDALCO	1	Sample		
18	Vial 18	0.10 CONTROL-RF	BLDALCO	1	Ctrl Samp		
19	Vial 19	BLANK	BLDALCO	1	Sample		

Calibration Part:

Line	Location	SampleName	Method	CalLev	Update RF	Update RT	Interval
2	Vial 2	0.079 CAL	BLDALCO	1	Replace	Average	
3	Vial 3	0.158 CAL	BLDALCO	2	Replace	Average	
4	Vial 4	0.316 CAL	BLDALCO	3	Replace	Average	

Sequence Table (Back Injector):

No entries - empty table!

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 10:43:24 AM
Instrument 4
DB-ALC1

BLANK
Rebecca Flaherty

vial # 1



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1623	1.645

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 10:46:41 AM
Instrument 4
DB-ALC1

0.079 CAL
Rebecca Flaherty

vial # 2



#	Compound	Area	RT
1	Ethanol	540	0.997
2	n-Propanol	1626	1.645

Totals:



Ethanol 0.079 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 10:49:56 AM
Instrument 4
DB-ALC1

0.158 CAL
Rebecca Flaherty

vial # 3



#	Compound	Area	RT
1	Ethanol	1026	0.997
2	n-Propanol	1571	1.645

Totals:



Ethanol 0.155 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 10:53:07 AM
Instrument 4
DB-ALC1

0.316 CAL
Rebecca Flaherty

vial # 4



#	Compound	Area	RT
1	Ethanol	2181	0.997
2	n-Propanol	1648	1.645

Totals:



Ethanol 0.317 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 10:56:30 AM
Instrument 4
DB-ALC1

BLANK
Rebecca Flaherty

vial # 5



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1668	1.645

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 10:59:42 AM
Instrument 4
DB-ALC1

0.04 MIX
Rebecca Flaherty

vial # 6



#	Compound	Area	RT
1		121	0.794
2	Ethanol	261	0.997
3		510	1.217
4		1125	1.466
5	n-Propanol	1588	1.645

Totals:



Ethanol 0.040 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 11:05:00 AM
Instrument 4
DB-ALC1

0.08 MIX
Rebecca Flaherty

vial # 7



#	Compound	Area	RT
1		257	0.794
2	Ethanol	547	0.997
3		1044	1.217
4		2255	1.466
5	n-Propanol	1644	1.645

Totals:



Ethanol 0.080 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 11:08:24 AM
Instrument 4
DB-ALC1

0.02 STD
Rebecca Flaherty

vial # 8



#	Compound	Area	RT
1	Ethanol	132	0.997
2	n-Propanol	1583	1.645

Totals:



Ethanol 0.020 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 11:11:43 AM
Instrument 4
DB-ALC1

0.04 CONTROL-RF
Rebecca Flaherty

vial # 9



#	Compound	Area	RT
1	Ethanol	259	0.997
2	n-Propanol	1577	1.645

Totals:



Ethanol 0.040 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 11:15:05 AM
Instrument 4
DB-ALC1

0.10 CONTROL-RF
Rebecca Flaherty

vial # 10



#	Compound	Area	RT
1	Ethanol	663	0.997
2	n-Propanol	1612	1.645

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 11:18:27 AM
Instrument 4
DB-ALC1

0.20 CONTROL-RF
Rebecca Flaherty

vial # 11



#	Compound	Area	RT
1	Ethanol	1303	0.997
2	n-Propanol	1584	1.645

Totals:



Ethanol 0.197 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 11:21:42 AM
Instrument 4
DB-ALC1

BLANK
Rebecca Flaherty

vial # 12



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1598	1.645

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 11:25:00 AM
Instrument 4
DB-ALC1

SIM 08006-1 RF
Rebecca Flaherty

vial # 13



#	Compound	Area	RT
1	Ethanol	721	0.997
2	n-Propanol	1707	1.645

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 11:28:17 AM
Instrument 4
DB-ALC1

SIM 08006-2 RF
Rebecca Flaherty

vial # 14



#	Compound	Area	RT
1	Ethanol	723	0.997
2	n-Propanol	1695	1.645

Totals:



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 11:33:45 AM
Instrument 4
DB-ALC1

SIM 08006-3 RF
Rebecca Flaherty

vial # 15



#	Compound	Area	RT
1	Ethanol	697	0.997
2	n-Propanol	1645	1.645

Totals:



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 11:36:58 AM
Instrument 4
DB-ALC1

SIM 08006-4 RF
Rebecca Flaherty

vial # 16



#	Compound	Area	RT
1	Ethanol	726	0.997
2	n-Propanol	1704	1.645

Totals:



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 11:40:19 AM
Instrument 4
DB-ALC1

SIM 08006-5 RF
Rebecca Flaherty

vial # 17



#	Compound	Area	RT
1	Ethanol	703	0.997
2	n-Propanol	1666	1.645

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 11:43:38 AM
Instrument 4
DB-ALC1

0.10 CONTROL-RF
Rebecca Flaherty

vial # 18



#	Compound	Area	RT
1	Ethanol	674	0.997
2	n-Propanol	1639	1.645

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/25/2008 11:46:56 AM
Instrument 4
DB-ALC1

BLANK
Rebecca Flaherty

vial # 19



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1724	1.645

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

RF

Sequence Parameters:

Operator:	alouis
Data File Naming:	Auto
Data Directory:	D:\HPCHEM\1\DATA\
Data Subdirectory:	080130AL
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	0.04 con a050530 0.10 con a050528 0.20 con a050527

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	blank	BLDALCO	1	Sample		
2	Vial 2	0.079 std	BLDALCO	1	Calib		
3	Vial 3	0.158 std	BLDALCO	1	Calib		
4	Vial 4	0.316 std	BLDALCO	1	Calib		
5	Vial 5	blank	BLDALCO	1	Ctrl Samp		
6	Vial 6	0.04 vols	VOL	1	Calib		
7	Vial 7	0.08 vols	VOL	1	Calib		
8	Vial 8	0.02 std	BLDALCO	1	Sample		
9	Vial 9	blank	BLDALCO	1	Sample		
10	Vial 10	0.04 con al	BLDALCO	1	Ctrl Samp		
11	Vial 11	0.10 con al	BLDALCO	1	Ctrl Samp		
12	Vial 12	0.20 con al	BLDALCO	1	Ctrl Samp		
13	Vial 13	blank	BLDALCO	1	Sample		
14	Vial 14	08006a al	BLDALCO	1	Sample		
15	Vial 15	08006b al	BLDALCO	1	Sample		
16	Vial 16	08006c al	BLDALCO	1	Sample		
17	Vial 17	08006d al	BLDALCO	1	Sample		
18	Vial 18	08006e al	BLDALCO	1	Sample		
19	Vial 19	0.10 con al	BLDALCO	1	Ctrl Samp		
20	Vial 20	blank	BLDALCO	1	Sample		

AL
20080204

Sequence Table (Back Injector):

No entries - empty table!

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 2:16:13 PM
Instrument 4
DB-ALC1

blank
alouis

vial # 1



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1485	1.654

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

AR
20080204

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 2:19:30 PM
Instrument 4
DB-ALC1

0.079 std
alouis

vial # 2



#	Compound	Area	RT
1	Ethanol	521	1.002
2	n-Propanol	1451	1.654

Totals:



Ethanol 0.078 g/100ml



n-Propanol 1.000 g/100ml

AL
20080204

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 2:22:43 PM 0.158 std
Instrument 4 alouis
DB-ALC1

vial # 3



#	Compound	Area	RT
1	Ethanol	1067	1.002
2	n-Propanol	1478	1.654

Totals:



Ethanol 0.158 g/100ml



n-Propanol 1.000 g/100ml

AL
20080204

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 2:26:04 PM
Instrument 4
DB-ALC1

0.316 std
alouis

vial # 4



#	Compound	Area	RT
1	Ethanol	2551	1.002
2	n-Propanol	1722	1.654

Totals:



Ethanol 0.317 g/100ml



n-Propanol 1.000 g/100ml

Al
20080204

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 2:29:43 PM
Instrument 4
DB-ALC1

blank
alouis

vial # 5



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1576	1.654

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

AR
20080204

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M
1/30/2008 2:42:17 PM
Instrument 4
DB-ALC1

0.04 vols
alouis

vial # 6



#	Compound	Area	RT
1	Methanol	138	0.800
2	Ethanol	296	1.004
3	Isopropanol	583	1.225
4	Acetone	1249	1.475
5	n-Propanol	1651	1.654

Totals:



Methanol 39.734 mg/dL



Ethanol 40.039 mg/dL

AR
20080204



Isopropanol 41.298 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M



Acetone 42.753 mg/dL



n-Propanol 1.000 mg/dL

AL
2008 0204

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M
1/30/2008 2:45:32 PM
Instrument 4
DB-ALC1

0.08 vols
alouis

vial # 7



#	Compound	Area	RT
1	Methanol	323	0.800
2	Ethanol	687	1.004
3	Isopropanol	1306	1.225
4	Acetone	2685	1.475
5	n-Propanol	1884	1.654

Totals:



Methanol 80.296 mg/dL



Ethanol 80.205 mg/dL

AL
20080204



Isopropanol 79.755 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M



Acetone 79.176 mg/dL



n-Propanol 1.000 mg/dL

AK
20080204

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 2:57:54 PM
Instrument 4
DB-ALC1

0.02 std
alouis

vial # 8



#	Compound	Area	RT
1	Ethanol	148	1.002
2	n-Propanol	1612	1.654

Totals:



Ethanol 0.021 g/100ml



n-Propanol 1.000 g/100ml

DL
20080204

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 3:01:12 PM
Instrument 4
DB-ALC1

blank
alouis

vial # 9



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1526	1.654

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

AR
2/20/2008

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 3:04:29 PM
Instrument 4
DB-ALC1

0.04 con al
alouis

vial # 10



#	Compound	Area	RT
1	Ethanol	310	1.002
2	n-Propanol	1713	1.653

Totals:



Ethanol 0.040 g/100ml



n-Propanol 1.000 g/100ml

AL
20080204

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 3:07:51 PM
Instrument 4
DB-ALC1

0.10 con al
alouis

vial # 11



#	Compound	Area	RT
1	Ethanol	769	1.002
2	n-Propanol	1682	1.654

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

AL
20080204

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 3:11:04 PM
Instrument 4
DB-ALC1

0.20 con al
alouis

vial # 12



#	Compound	Area	RT
1	Ethanol	1528	1.002
2	n-Propanol	1668	1.654

Totals:



Ethanol 0.197 g/100ml



n-Propanol 1.000 g/100ml

AR
20080204

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 3:14:16 PM
Instrument 4
DB-ALC1

blank
alouis

vial # 13



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1623	1.654

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

AR
20080204

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 3:19:51 PM
Instrument 4
DB-ALC1

08006a al
alouis

vial # 14



#	Compound	Area	RT
1	Ethanol	723	1.002
2	n-Propanol	1546	1.654

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

Al
20080204

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 3:23:06 PM
Instrument 4
DB-ALC1

08006b al
alouis

vial # 15



#	Compound	Area	RT
1	Ethanol	686	1.002
2	n-Propanol	1475	1.654

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

AL
20080204

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 3:26:25 PM
Instrument 4
DB-ALC1

08006c al
alouis

vial # 16



#	Compound	Area	RT
1	Ethanol	745	1.002
2	n-Propanol	1590	1.653

Totals:



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

AL
20080204

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 3:29:44 PM
Instrument 4
DB-ALC1

08006d al
alouis

vial # 17



#	Compound	Area	RT
1	Ethanol	762	1.002
2	n-Propanol	1622	1.654

Totals:



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

AK
20080204

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 3:33:01 PM
Instrument 4
DB-ALC1

08006e al
alouis

vial # 18



#	Compound	Area	RT
1	Ethanol	787	1.002
2	n-Propanol	1670	1.653

Totals:



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

AR
20080224

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 3:36:17 PM
Instrument 4
DB-ALC1

0.10 con al
alouis

vial # 19



#	Compound	Area	RT
1	Ethanol	780	1.002
2	n-Propanol	1711	1.653

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

AR
20080204

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
1/30/2008 3:39:32 PM
Instrument 4
DB-ALC1

blank
alouis

vial # 20



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1605	1.654

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

AR
20080204

Sequence Parameters:

Operator:	Justin Knoy
Data File Naming:	Auto
Data Directory:	D:\HPCHEM\1\DATA\
Data Subdirectory:	080207JK
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	BLDALCO	1	Sample		
2	Vial 2	0.079 CAL	BLDALCO	1	Calib		
3	Vial 3	0.158 CAL	BLDALCO	1	Calib		
4	Vial 4	0.316 CAL	BLDALCO	1	Calib		
5	Vial 5	BLANK	BLDALCO	1	Ctrl Samp		
6	Vial 6	0.04 MIX	VOL	1	Calib		
7	Vial 7	0.08 MIX	VOL	1	Calib		
8	Vial 8	0.02 STD	BLDALCO	1	Sample		
9	Vial 9	0.04 CTRL JK	BLDALCO	1	Ctrl Samp	A050530 Exp 7/2011	
10	Vial 10	0.10 CTRL JK	BLDALCO	1	Ctrl Samp	A050528 Exp 7/2011	
11	Vial 11	0.20 CTRL JK	BLDALCO	1	Ctrl Samp	A050527 Exp 7/2011	
12	Vial 12	BLANK	BLDALCO	1	Sample		
13	Vial 13	SIM 08006-1	BLDALCO	1	Sample		
14	Vial 14	SIM 08006-2	BLDALCO	1	Sample		
15	Vial 15	SIM 08006-3	BLDALCO	1	Sample		
16	Vial 16	SIM 08006-4	BLDALCO	1	Sample		
17	Vial 17	SIM 08006-5	BLDALCO	1	Sample		
18	Vial 18	0.10 CTRL JK	BLDALCO	1	Ctrl Samp		
19	Vial 19	blank	BLDALCO	1	Sample		

Sequence Table (Back Injector):

No entries - empty table!

JK 2-8-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 3:01:47 PM
Instrument 4
DB-ALC1

BLANK
Justin Knoy

vial # 1



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1512	1.656

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

JK 2-8-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 3:05:06 PM
Instrument 4
DB-ALC1

0.079 CAL
Justin Knoy

vial # 2



#	Compound	Area	RT
1	Ethanol	526	1.003
2	n-Propanol	1531	1.656

Totals:



Ethanol 0.080 g/100ml



n-Propanol 1.000 g/100ml

2-8-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 3:08:19 PM
Instrument 4
DB-ALC1

0.158 CAL
Justin Knoy

vial # 3



#	Compound	Area	RT
1	Ethanol	1040	1.003
2	n-Propanol	1532	1.655

Totals:



Ethanol 0.159 g/100ml



n-Propanol 1.000 g/100ml

2-8-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 3:11:40 PM
Instrument 4
DB-ALC1

0.316 CAL
Justin Knoy

vial # 4



#	Compound	Area	RT
1	Ethanol	2149	1.003
2	n-Propanol	1579	1.655

Totals:



Ethanol 0.316 g/100ml



n-Propanol 1.000 g/100ml

JK 2-8-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 3:15:00 PM
Instrument 4
DB-ALC1

BLANK
Justin Knoy

vial # 5



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1589	1.655

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

JK 2-8-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M
2/7/2008 3:27:34 PM
Instrument 4
DB-ALC1

0.04 MIX
Justin Knoy

vial # 6



#	Compound	Area	RT
1	Methanol	138	0.801
2	Ethanol	296	1.005
3	Isopropanol	582	1.226
4	Acetone	1230	1.477
5	n-Propanol	1719	1.656

Totals:



Methanol 40.085 mg/dL



Ethanol 39.931 mg/dL



Isopropanol 40.248 mg/dL

JK 2-8-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M



Acetone 40.202 mg/dL



n-Propanol 1.000 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M
2/7/2008 3:30:49 PM
Instrument 4
DB-ALC1

0.08 MIX
Justin Knoy

vial # 7



#	Compound	Area	RT
1	Methanol	276	0.801
2	Ethanol	598	1.005
3	Isopropanol	1170	1.226
4	Acetone	2464	1.477
5	n-Propanol	1736	1.657
Totals:			



Methanol 79.890 mg/dL



Ethanol 79.997 mg/dL



Isopropanol 79.936 mg/dL

JK 2-8-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M



Acetone 79.891 mg/dL



n-Propanol 1.000 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 3:43:23 PM
Instrument 4
DB-ALC1

0.02 STD
Justin Knoy

vial # 8



#	Compound	Area	RT
1	Ethanol	133	1.003
2	n-Propanol	1529	1.656
Totals:			



Ethanol 0.020 g/100ml



n-Propanol 1.000 g/100ml

JK 2-8-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 3:46:39 PM
Instrument 4
DB-ALC1

0.04 CTRL JK
Justin Knoy

vial # 9



#	Compound	Area	RT
1	Ethanol	264	1.003
2	n-Propanol	1532	1.656
Totals:			



Ethanol 0.040 g/100ml



n-Propanol 1.000 g/100ml

JK 2-8-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 3:49:56 PM
Instrument 4
DB-ALC1

0.10 CTRL JK
Justin Knoy

vial # 10



#	Compound	Area	RT
1	Ethanol	664	1.003
2	n-Propanol	1544	1.656

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

K 2-8-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 3:53:09 PM
Instrument 4
DB-ALC1

0.20 CTRL JK
Justin Knoy

vial # 11



#	Compound	Area	RT
1	Ethanol	1402	1.003
2	n-Propanol	1602	1.656

Totals:



Ethanol 0.203 g/100ml



n-Propanol 1.000 g/100ml

JK 2-8-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 3:56:21 PM
Instrument 4
DB-ALC1

BLANK
Justin Knoy

vial # 12



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1552	1.656

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

JK 2.8.08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 3:59:43 PM
Instrument 4
DB-ALC1

SIM 08006-1
Justin Knoy

vial # 13



#	Compound	Area	RT
1	Ethanol	680	1.003
2	n-Propanol	1547	1.656

Totals:



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

JK 2-8-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 4:05:11 PM
Instrument 4
DB-ALC1

SIM 08006-2
Justin Knoy

vial # 14



#	Compound	Area	RT
1	Ethanol	686	1.003
2	n-Propanol	1549	1.656

Totals:



Ethanol 0.103 g/100ml



n-Propanol 1.000 g/100ml

JK 2-8-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 4:08:26 PM
Instrument 4
DB-ALC1

SIM 08006-3
Justin Knoy

vial # 15



#	Compound	Area	RT
1	Ethanol	665	1.003
2	n-Propanol	1506	1.655

Totals:



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

JK 2.8.08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 4:11:45 PM
Instrument 4
DB-ALC1

SIM 08006-4
Justin Knoy

vial # 16



#	Compound	Area	RT
1	Ethanol	673	1.003
2	n-Propanol	1525	1.655
Totals:			



Ethanol 0.102 g/100ml



n-Propanol 1.000 g/100ml

JK 2-8-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 4:15:04 PM
Instrument 4
DB-ALC1

SIM 08006-5
Justin Knoy

vial # 17



#	Compound	Area	RT
1	Ethanol	693	1.003
2	n-Propanol	1552	1.655
Totals:			



Ethanol 0.103 g/100ml



n-Propanol 1.000 g/100ml

JK 2.8.08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 4:18:19 PM
Instrument 4
DB-ALC1

0.10 CTRL JK
Justin Knoy

vial # 18



#	Compound	Area	RT
1	Ethanol	696	1.003
2	n-Propanol	1616	1.655

Totals:



Ethanol 0.100 g/100ml



n-Propanol 1.000 g/100ml

JK 2-8-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 4:21:35 PM
Instrument 4
DB-ALC1

blank
Justin Knoy

vial # 19



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1638	1.655

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

2-8-08

Sequence Parameters:

Operator:	Estuardo J. Miranda
Data File Naming:	Auto
Data Directory:	D:\HPCHEM\1\DATA\
Data Subdirectory:	080207EM
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none

Sequence Comment:
Restek Control 0.04 g/100mL; Lot# 050530; Exp 07/2011
Restek Control 0.10 g/100mL; Lot# 050528; Exp 07/2011
Restek Control 0.20 g/100mL; Lot# 050527; Exp 07/2011

EM

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	BLDALCO	1	Sample		
2	Vial 2	0.079 CAL-1-EM	BLDALCO	1	Calib		
3	Vial 3	0.158 CAL-2-EM	BLDALCO	1	Calib		
4	Vial 4	0.316 CAL-3 EM	BLDALCO	1	Calib		
5	Vial 5	Neg Control-EM	BLDALCO	1	Ctrl Samp		
6	Vial 6	40 mg/dL Vol-Mix	VOL	1	Calib		
7	Vial 7	79 mg/dL Vol-Mix	VOL	1	Calib		
8	Vial 8	0.02 STD-EM Std	BLDALCO	1	Sample		
9	Vial 9	0.040-EM Control	BLDALCO	1	Ctrl Samp		
10	Vial 10	0.100-EM Control	BLDALCO	1	Ctrl Samp		
11	Vial 11	0.200-EM Control	BLDALCO	1	Ctrl Samp		
12	Vial 12	BLANK	BLDALCO	1	Sample		
13	Vial 13	Sim Sol 08006-1	BLDALCO	1	Sample		
14	Vial 14	Sim Sol 08006-2	BLDALCO	1	Sample		
15	Vial 15	Sim Sol 08006-3	BLDALCO	1	Sample		
16	Vial 16	Sim Sol 08006-4	BLDALCO	1	Sample		
17	Vial 17	Sim Sol 08006-5	BLDALCO	1	Sample		
18	Vial 18	0.100 EM Control	BLDALCO	1	Ctrl Samp		
19	Vial 19	BLANK	BLDALCO	1	Sample		

Calibration Part:

Line	Location	SampleName	Method	CalLev	Update RF	Update RT	Interval
2	Vial 2	0.079 CAL-1-EM	BLDALCO	1	Replace	Average	
3	Vial 3	0.158 CAL-2-EM	BLDALCO	2	Replace	Average	
4	Vial 4	0.316 CAL-3 EM	BLDALCO	3	Replace	Average	
6	Vial 6	40 mg/dL Vol-Mix	VOL	1	Replace	Average	
7	Vial 7	79 mg/dL Vol-Mix	VOL	2	Replace	Average	

Sequence Table (Back Injector):

Sequence: D:\HPCHEM\1\SEQUENCE\EM_SIM.S

No entries - empty table!

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 5:01:59 PM
Instrument 4
DB-ALC1

BLANK
Estuardo J. Miranda EJM

vial # 1



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1514	1.656

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 5:05:17 PM
Instrument 4
DB-ALC1

0.079 CAL-1-EM
Estuardo J. Miranda

EM

vial # 2



#	Compound	Area	RT
1	Ethanol	535	1.003
2	n-Propanol	1544	1.656

Totals:



Ethanol 0.080 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 5:08:30 PM
Instrument 4
DB-ALC1

0.158 CAL-2-EM
Estuardo J. Miranda

EM

vial # 3



#	Compound	Area	RT
1	Ethanol	1068	1.003
2	n-Propanol	1558	1.656

Totals:



Ethanol 0.158 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 5:12:05 PM
Instrument 4
DB-ALC1

0.316 CAL-3 EM
Estuardo J. Miranda

vial # 4



#	Compound	Area	RT
1	Ethanol	2136	1.003
2	n-Propanol	1574	1.656
Totals:			



Ethanol 0.315 g/100ml

EM



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 5:15:25 PM
Instrument 4
DB-ALC1

Neg Control-EM
Estuardo J. Miranda

vial # 5



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1640	1.656

Totals:



Ethanol 0.000 g/100ml

EM



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M
2/7/2008 5:27:55 PM
Instrument 4
DB-ALC1

40 mg/dL Vol-Mix
Estuardo J. Miranda

vial # 6



#	Compound	Area	RT
1	Methanol	142	0.801
2	Ethanol	307	1.005
3	Isopropanol	601	1.226
4	Acetone	1282	1.477
5	n-Propanol	1765	1.657 ✓
Totals:			



Methanol 40.365 mg/dL



Ethanol 40.246 mg/dL



Isopropanol 40.292 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M



Acetone 40.621 mg/dL

EM



n-Propanol 1.000 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M
2/7/2008 5:31:10 PM
Instrument 4
DB-ALC1

79 mg/dL Vol-Mix
Estuardo J. Miranda

vial # 7



#	Compound	Area	RT
1	Methanol	277	0.801
2	Ethanol	600	1.005
3	Isopropanol	1175	1.226
4	Acetone	2519	1.477
5	n-Propanol	1750	1.657

Totals:



Methanol 79.759 mg/dL



Ethanol 79.824 mg/dL



Isopropanol 79.818 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M



Acetone 79.883 mg/dL



n-Propanol 1.000 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 5:43:43 PM
Instrument 4
DB-ALC1

0.02 STD-EM Std
Estuardo J. Miranda

vial # 8



#	Compound	Area	RT
1	Ethanol	135	1.003
2	n-Propanol	1536	1.656

Totals:



Ethanol 0.019 g/100ml

EM



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 5:47:00 PM
Instrument 4
DB-ALC1

0.040-EM Control
Estuardo J. Miranda

vial # 9



#	Compound	Area	RT
1	Ethanol	262	1.003
2	n-Propanol	1524	1.656

Totals:



Ethanol 0.039 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 5:50:16 PM
Instrument 4
DB-ALC1

0.100-EM Control
Estuardo J. Miranda

vial # 10



#	Compound	Area	RT
1	Ethanol	664	1.003
2	n-Propanol	1534	1.656

Totals:



Ethanol 0.100 g/100ml

EM



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 5:53:37 PM
Instrument 4
DB-ALC1

0.200-EM Control
Estuardo J. Miranda

vial # 11



#	Compound	Area	RT
1	Ethanol	1333	1.003
2	n-Propanol	1537	1.656

Totals:



Ethanol 0.201 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 5:57:00 PM
Instrument 4
DB-ALC1

BLANK
Estuardo J. Miranda

vial # 12



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1555	1.656
Totals:			



Ethanol 0.000 g/100ml

EM



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 6:00:21 PM
Instrument 4
DB-ALC1

Sim Sol 08006-1
Estuardo J. Miranda

vial # 13



#	Compound	Area	RT
1	Ethanol	669	1.003
2	n-Propanol	1519	1.656 ✓

Totals:



Ethanol 0.102 g/100ml

EM



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 6:05:35 PM
Instrument 4
DB-ALC1

Sim Sol 08006-2
Estuardo J. Miranda

vial # 14



#	Compound	Area	RT
1	Ethanol	795	1.003
2	n-Propanol	1752	1.656

Totals:



Ethanol 0.105 g/100ml

EM



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 6:09:03 PM
Instrument 4
DB-ALC1

Sim Sol 08006-3
Estuardo J. Miranda

vial # 15



#	Compound	Area	RT
1	Ethanol	667	1.003
2	n-Propanol	1500	1.656 ✓

Totals:



Ethanol 0.103 g/100ml

EM



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 6:12:22 PM
Instrument 4
DB-ALC1

Sim Sol 08006-4
Estuardo J. Miranda

vial # 16



#	Compound	Area	RT
1	Ethanol	685	1.004
2	n-Propanol	1530	1.657

Totals:



Ethanol 0.103 g/100ml

EM



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 6:15:41 PM
Instrument 4
DB-ALC1

Sim Sol 08006-5
Estuardo J. Miranda

vial # 17



#	Compound	Area	RT
1	Ethanol	678	1.003
2	n-Propanol	1523	1.656
Totals:			



Ethanol 0.103 g/100ml

EM



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 6:18:56 PM
Instrument 4
DB-ALC1

0.100 EM Control
Estuardo J. Miranda

vial # 18



#	Compound	Area	RT
1	Ethanol	684	1.003
2	n-Propanol	1577	1.656

Totals:



Ethanol 0.100 g/100ml

EM



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/7/2008 6:22:11 PM
Instrument 4
DB-ALC1

BLANK
Estuardo J. Miranda

vial # 19



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1910	1.656
Totals:			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

Sequence Parameters:

Operator:	Brianna Peterson
Data File Naming:	Auto
Data Directory:	D:\HPCHEM\1\DATA\
Data Subdirectory:	080208BP
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	BLDALCO	1	Sample		
2	Vial 2	0.079 STD	BLDALCO	1	Calib		
3	Vial 3	0.158 STD	BLDALCO	1	Calib		
4	Vial 4	0.316 STD	BLDALCO	1	Calib		
5	Vial 5	BLANK BP	BLDALCO	1	Ctrl Samp		
6	Vial 6	0.04 MIX	VOL	1	Calib		
7	Vial 7	0.08 MIX	VOL	1	Calib		
8	Vial 8	0.02 STD	BLDALCO	1	Sample		
9	Vial 9	0.04 CTRL BP	BLDALCO	1	Ctrl Samp	Lot # = A050530	Exp: 07/2011
10	Vial 10	0.10 CTRL BP	BLDALCO	1	Ctrl Samp	Lot # = A050528	Exp: 07/2011
11	Vial 11	0.20 CTRL BP	BLDALCO	1	Ctrl Samp	Lot # = A050527	Exp: 07/2011
12	Vial 12	BLANK	BLDALCO	1	Sample		
13	Vial 13	SIM 08006-1 BP	BLDALCO	1	Sample		
14	Vial 14	SIM 08006-2 BP	BLDALCO	1	Sample		
15	Vial 15	SIM 08006-3 BP	BLDALCO	1	Sample		
16	Vial 16	SIM 08006-4 BP	BLDALCO	1	Sample		
17	Vial 17	SIM 08006-5 BP	BLDALCO	1	Sample		
18	Vial 18	0.10 CTRL BP	BLDALCO	1	Ctrl Samp		
19	Vial 19	BLANK	BLDALCO	1	Sample		

BP

Calibration Part:

Line	Location	SampleName	Method	CalLev	Update RF	Update RT	Interval
2	Vial 2	0.079 STD	BLDALCO	1	Replace	Average	
3	Vial 3	0.158 STD	BLDALCO	2	Replace	Average	
4	Vial 4	0.316 STD	BLDALCO	3	Replace	Average	
6	Vial 6	0.04 MIX	VOL	1	Replace	Average	
7	Vial 7	0.08 MIX	VOL	2	Replace	Average	

Sequence Table (Back Injector):

No entries - empty table!

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 11:15:45 AM
Instrument 4
DB-ALC1

BLANK
Brianna Peterson

vial # 1



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1641	1.657

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 11:19:04 AM
Instrument 4
DB-ALC1

0.079 STD
Brianna Peterson

vial # 2



#	Compound	Area	RT
1	Ethanol	552	1.004
2	n-Propanol	1596	1.658

Totals:



Ethanol 0.078 g/100ml



n-Propanol 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 11:22:17 AM
Instrument 4
DB-ALC1

0.158 STD
Brianna Peterson

vial # 3



#	Compound	Area	RT
1	Ethanol	1113	1.005
2	n-Propanol	1618	1.658

Totals:



Ethanol 0.154 g/100ml



n-Propanol 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 11:25:38 AM
Instrument 4
DB-ALC1

0.316 STD
Brianna Peterson

vial # 4



#	Compound	Area	RT
1	Ethanol	2295	1.004
2	n-Propanol	1662	1.658

Totals:



Ethanol 0.316 g/100ml



n-Propanol 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 11:28:57 AM
Instrument 4
DB-ALC1

BLANK BP
Brianna Peterson

vial # 5



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1555	1.657

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M
2/8/2008 11:41:24 AM
Instrument 4
DB-ALC1

0.04 MIX
Brianna Peterson

vial # 6



#	Compound	Area	RT
1	Methanol	136	0.802
2	Ethanol	292	1.006
3	Isopropanol	573	1.227
4	Acetone	1201	1.478
5	n-Propanol	1672	1.658

Totals:



Methanol 38.881 mg/dL



Ethanol 39.130 mg/dL



Isopropanol 39.933 mg/dL

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M



Acetone 40.475 mg/dL



n-Propanol 1.000 mg/dL

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M
2/8/2008 11:44:40 AM
Instrument 4
DB-ALC1

0.08 MIX
Brianna Peterson

vial # 7



#	Compound	Area	RT
1	Methanol	290	0.802
2	Ethanol	620	1.006
3	Isopropanol	1191	1.227
4	Acetone	2457	1.479
5	n-Propanol	1755	1.658
Totals:			



Methanol 80.222 mg/dL



Ethanol 80.139 mg/dL



Isopropanol 79.873 mg/dL

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M



Acetone 79.658 mg/dL



n-Propanol 1.000 mg/dL

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 11:57:05 AM
Instrument 4
DB-ALC1

0.02 STD
Brianna Peterson

vial # 8



#	Compound	Area	RT
1	Ethanol	138	1.004
2	n-Propanol	1540	1.657

Totals:



Ethanol 0.020 g/100ml



n-Propanol 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 12:00:23 PM
Instrument 4
DB-ALC1

0.04 CTRL BP
Brianna Peterson

vial # 9



#	Compound	Area	RT
1	Ethanol	270	1.004
2	n-Propanol	1530	1.657

Totals:



Ethanol 0.040 g/100ml



n-Propanol 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 12:03:39 PM
Instrument 4
DB-ALC1

0.10 CTRL BP
Brianna Peterson

vial # 10



#	Compound	Area	RT
1	Ethanol	672	1.004
2	n-Propanol	1526	1.657

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 12:06:52 PM
Instrument 4
DB-ALC1

0.20 CTRL BP
Brianna Peterson

vial # 11



#	Compound	Area	RT
1	Ethanol	1363	1.004
2	n-Propanol	1543	1.657

Totals:



Ethanol 0.202 g/100ml



n-Propanol 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 12:10:04 PM
Instrument 4
DB-ALC1

BLANK
Brianna Peterson

vial # 12



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1633	1.657

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 12:13:27 PM
Instrument 4
DB-ALC1

SIM 08006-1 BP
Brianna Peterson

vial # 13



#	Compound	Area	RT
1	Ethanol	828	1.004
2	n-Propanol	1767	1.657

Totals:



Ethanol 0.107 g/100ml



n-Propanol 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 12:18:46 PM
Instrument 4
DB-ALC1

SIM 08006-2 BP
Brianna Peterson

vial # 14



#	Compound	Area	RT
1	Ethanol	708	1.004
2	n-Propanol	1553	1.657

Totals:



Ethanol 0.104 g/100ml



n-Propanol 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 12:22:03 PM
Instrument 4
DB-ALC1

SIM 08006-3 BP
Brianna Peterson

vial # 15



#	Compound	Area	RT
1	Ethanol	670	1.004
2	n-Propanol	1478	1.657

Totals:



Ethanol 0.104 g/100ml



n-Propanol 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 12:25:22 PM
Instrument 4
DB-ALC1

SIM 08006-4 BP
Brianna Peterson

vial # 16



#	Compound	Area	RT
1	Ethanol	674	1.004
2	n-Propanol	1488	1.657

Totals:



Ethanol 0.104 g/100ml



n-Propanol 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 12:28:40 PM
Instrument 4
DB-ALC1

SIM 08006-5 BP
Brianna Peterson

vial # 17



#	Compound	Area	RT
1	Ethanol	707	1.004
2	n-Propanol	1555	1.657

Totals:



Ethanol 0.104 g/100ml



n-Propanol 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 12:31:56 PM
Instrument 4
DB-ALC1

0.10 CTRL BP
Brianna Peterson

vial # 18



#	Compound	Area	RT
1	Ethanol	708	1.004
2	n-Propanol	1598	1.657

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 12:35:12 PM
Instrument 4
DB-ALC1

BLANK
Brianna Peterson

vial # 19



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1648	1.657

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 10:31:21 AM
Instrument 4
DB-ALC1

BLANK
BRITTANY BALL

vial # 1



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1539	1.657
Totals:			





Ethanol 0.000 g/100ml

n-Propanol 1.000 g/100ml

Wrong operator listed in sequence parameters, sequence aborted and restarted from beginning

BP 02-08-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 10:34:38 AM
Instrument 4
DB-ALC1

0.079 STD
BRITTANY BALL

vial # 2



#	Compound	Area	RT
1	Ethanol	516	1.004
2	n-Propanol	1483	1.657
Totals:			





Ethanol 0.078 g/100ml

n-Propanol 1.000 g/100ml

Wrong operator listed in sequence parameters
Sequence aborted and started from beginning
BP 02-08-08

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/8/2008 10:38:00 AM
Instrument 4
DB-ALC1

0.158 STD
BRITTANY BALL

vial # 3



#	Compound	Area	RT
1	Ethanol	1057	1.004
2	n-Propanol	1521	1.657
Totals:			



Ethanol 0.155 g/100ml



n-Propanol 1.000 g/100ml

Wrong operator listed in sequence parameters
Sequence aborted and restarted from beginning
BP 02-08-08

Sequence Parameters:

Operator:	Christie Mitchell
Data File Naming:	Auto
Data Directory:	D:\HPCHEM\1\DATA\
Data Subdirectory:	080211CM
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile	Lot#	Exp.
1	Vial 1	BLANK	BLDALCO	1	Sample				
2	Vial 2	0.079 CAL	BLDALCO	1	Calib				
3	Vial 3	0.158 CAL	BLDALCO	1	Calib				
4	Vial 4	0.316 CAL	BLDALCO	1	Calib				
5	Vial 5	BLANK	BLDALCO	1	Sample				
6	Vial 6	0.04 MIX	VOL	1	Calib				
7	Vial 7	0.08 MIX	VOL	1	Calib			0.04 Cntl A050530	7/2011
8	Vial 8	0.02 STD	BLDALCO	1	Sample				
9	Vial 9	BLANK	BLDALCO	1	Ctrl Samp				
10	Vial 10	0.04 CONTROL cm	BLDALCO	1	Ctrl Samp			0.10 Cntl A050528	7/2011
11	Vial 11	0.10 CONTROL cm	BLDALCO	1	Ctrl Samp				
12	Vial 12	0.20 CONTROL cm	BLDALCO	1	Ctrl Samp			0.20 Cntl A050527	7/2011
13	Vial 13	BLANK	BLDALCO	1	Sample				
14	Vial 14	SIM 08006-1 CM	BLDALCO	1	Sample				
15	Vial 15	SIM 08006-2 CM	BLDALCO	1	Sample				
16	Vial 16	SIM 08006-3 CM	BLDALCO	1	Sample				
17	Vial 17	SIM 08006-4 CM	BLDALCO	1	Sample				
18	Vial 18	SIM 08006-5 CM	BLDALCO	1	Sample				
19	Vial 19	0.10 CONTROL cm	BLDALCO	1	Ctrl Samp				
20	Vial 20	BLANK	BLDALCO	1	Sample				

Sequence Table (Back Injector):

No entries - empty table!

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 7:55:46 AM
Instrument 4
DB-ALC1

BLANK
Christie Mitchell

vial # 1



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1515	1.663

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 7:59:04 AM
Instrument 4
DB-ALC1

0.079 CAL
Christie Mitchell

vial # 2



#	Compound	Area	RT
1	Ethanol	489	1.007
2	n-Propanol	1414	1.663

Totals:



Ethanol 0.079 g/100ml



n-Propanol 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 8:02:16 AM
Instrument 4
DB-ALC1

0.158 CAL
Christie Mitchell

vial # 3



#	Compound	Area	RT
1	Ethanol	1003	1.007
2	n-Propanol	1446	1.663

Totals:



Ethanol 0.158 g/100ml



n-Propanol 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 8:05:29 AM
Instrument 4
DB-ALC1

0.316 CAL
Christie Mitchell

vial # 4



#	Compound	Area	RT
1	Ethanol	2207	1.007
2	n-Propanol	1566	1.663

Totals:



Ethanol 0.317 g/100ml



n-Propanol 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 8:08:48 AM
Instrument 4
DB-ALC1

BLANK
Christie Mitchell

vial # 5



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1494	1.663

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M
2/11/2008 8:21:21 AM
Instrument 4
DB-ALC1

0.04 MIX
Christie Mitchell

vial # 6



#	Compound	Area	RT
1	Methanol	130	0.804
2	Ethanol	282	1.009
3	Isopropanol	558	1.231
4	Acetone	1206	1.483
5	n-Propanol	1618	1.663

Totals:



Methanol 39.346 mg/dL



Ethanol 39.618 mg/dL



Isopropanol 40.411 mg/dL

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M



Acetone 41.718 mg/dL



n-Propanol 1.000 mg/dL

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M
2/11/2008 8:24:34 AM
Instrument 4
DB-ALC1

0.08 MIX
Christie Mitchell

vial # 7



#	Compound	Area	RT
1	Methanol	299	0.804
2	Ethanol	634	1.009
3	Isopropanol	1206	1.231
4	Acetone	2503	1.483
5	n-Propanol	1779	1.663
6		23	2.170
Totals:			



Methanol 80.515 mg/dL



Ethanol 80.300 mg/dL



Isopropanol 79.777 mg/dL

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M



Acetone 79.205 mg/dL



n-Propanol 1.000 mg/dL

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 8:37:06 AM
Instrument 4
DB-ALC1

0.02 STD
Christie Mitchell

vial # 8



#	Compound	Area	RT
1	Ethanol	140	1.007
2	n-Propanol	1552	1.662

Totals:



Ethanol 0.021 g/100ml



n-Propanol 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 8:40:24 AM
Instrument 4
DB-ALC1

BLANK
Christie Mitchell

vial # 9



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1630	1.662
Totals:			



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 8:43:39 AM
Instrument 4
DB-ALC1

0.04 CONTROL cm
Christie Mitchell

vial # 10



#	Compound	Area	RT
1	Ethanol	251	1.007
2	n-Propanol	1441	1.662

Totals:



Ethanol 0.040 g/100ml



n-Propanol 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 8:46:52 AM
Instrument 4
DB-ALC1

0.10 CONTROL cm
Christie Mitchell

vial # 11



#	Compound	Area	RT
1	Ethanol	636	1.007
2	n-Propanol	1454	1.662

Totals:



Ethanol 0.099 g/100ml



n-Propanol 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 8:50:05 AM
Instrument 4
DB-ALC1

0.20 CONTROL cm
Christie Mitchell

vial # 12



#	Compound	Area	RT
1	Ethanol	1263	1.007
2	n-Propanol	1445	1.663

Totals:



Ethanol 0.197 g/100ml



n-Propanol 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 8:53:17 AM
Instrument 4
DB-ALC1

BLANK
Christie Mitchell

vial # 13



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1538	1.663

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 8:58:52 AM
Instrument 4
DB-ALC1

SIM 08006-1 CM
Christie Mitchell

vial # 14



#	Compound	Area	RT
1	Ethanol	636	1.007
2	n-Propanol	1421	1.663

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 9:02:12 AM
Instrument 4
DB-ALC1

SIM 08006-2 CM
Christie Mitchell

vial # 15



#	Compound	Area	RT
1	Ethanol	633	1.007
2	n-Propanol	1413	1.663

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 9:05:31 AM
Instrument 4
DB-ALC1

SIM 08006-3 CM
Christie Mitchell

vial # 16



#	Compound	Area	RT
1	Ethanol	631	1.007
2	n-Propanol	1411	1.663

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 9:08:48 AM
Instrument 4
DB-ALC1

SIM 08006-4 CM
Christie Mitchell

vial # 17



#	Compound	Area	RT
1	Ethanol	717	1.007
2	n-Propanol	1562	1.662

Totals:



Ethanol 0.104 g/100ml



n-Propanol 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 9:12:05 AM
Instrument 4
DB-ALC1

SIM 08006-5 CM
Christie Mitchell

vial # 18



#	Compound	Area	RT
1	Ethanol	636	1.007
2	n-Propanol	1427	1.662

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 9:15:21 AM
Instrument 4
DB-ALC1

0.10 CONTROL cm
Christie Mitchell

vial # 19



#	Compound	Area	RT
1	Ethanol	719	1.007
2	n-Propanol	1607	1.662

Totals:



Ethanol 0.101 g/100ml



n-Propanol 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/11/2008 9:18:35 AM
Instrument 4
DB-ALC1

BLANK
Christie Mitchell

vial # 20



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1506	1.663

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

CM

Sequence Parameters:

Operator:	A. Black
Data File Naming:	Auto
Data Directory:	D:\HPCHEM\1\DATA\
Data Subdirectory:	080206AB
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	BLDALCO	1	Sample		
2	Vial 2	0.079 CAL-AB	BLDALCO	1	Calib		
3	Vial 3	0.158 CAL-AB	BLDALCO	1	Calib		
4	Vial 4	0.316 CAL-AB	BLDALCO	1	Calib		
5	Vial 5	BLANK AB Control	BLDALCO	1	Ctrl Samp		
6	Vial 6	0.02 STD AB	BLDALCO	1	Sample		
7	Vial 7	40 MIX-CAL-AB	VOL	1	Calib		
8	Vial 8	80 MIX-CAL-AB	VOL	1	Calib		
9	Vial 9	0.04 AB Control	BLDALCO	1	Ctrl Samp	A050530	7/2011
10	Vial 10	0.10 AB Control	BLDALCO	1	Ctrl Samp	A050528	7/2011
11	Vial 11	0.20 AB Control	BLDALCO	1	Ctrl Samp	A050527	7/2011
12	Vial 12	BLANK	BLDALCO	1	Sample		
13	Vial 13	SIM 08006-1	BLDALCO	1	Sample		
14	Vial 14	SIM 08006-2	BLDALCO	1	Sample		
15	Vial 15	SIM 08006-3	BLDALCO	1	Sample		
16	Vial 16	SIM 08006-4	BLDALCO	1	Sample		
17	Vial 17	SIM 08006-5	BLDALCO	1	Sample		
18	Vial 18	0.10 AB Control	BLDALCO	1	Ctrl Samp		
19	Vial 19	BLANK	BLDALCO	1	Sample		

Sequence Table (Back Injector):

No entries - empty table!

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/6/2008 9:42:49 AM
Instrument 4
DB-ALC1

BLANK
A. Black

vial # 1



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1508	1.655

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

05

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/6/2008 9:46:07 AM
Instrument 4
DB-ALC1

0.079 CAL-AB
A. Black

vial # 2



#	Compound	Area	RT
1	Ethanol	502	1.002
2	n-Propanol	1516	1.654

Totals:



Ethanol 0.079 g/100ml



n-Propanol 1.000 g/100ml

AB

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/6/2008 9:49:22 AM
Instrument 4
DB-ALC1

0.158 CAL-AB
A. Black

vial # 3



#	Compound	Area	RT
1	Ethanol	985	1.003
2	n-Propanol	1492	1.655

Totals:



Ethanol 0.158 g/100ml



n-Propanol 1.000 g/100ml

AB

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/6/2008 9:52:45 AM
Instrument 4
DB-ALC1

0.316 CAL-AB
A. Black

vial # 4



#	Compound	Area	RT
1	Ethanol	2040	1.003
2	n-Propanol	1539	1.655

Totals:



Ethanol 0.316 g/100ml



n-Propanol 1.000 g/100ml

ab

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/6/2008 9:56:03 AM
Instrument 4
DB-ALC1

BLANK AB Control
A. Black

vial # 5



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1653	1.655

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

AB

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/6/2008 9:59:22 AM
Instrument 4
DB-ALC1

0.02 STD AB
A. Black

vial # 6



#	Compound	Area	RT
1	Ethanol	135	1.003
2	n-Propanol	1571	1.655

Totals:



Ethanol 0.021 g/100ml



n-Propanol 1.000 g/100ml

AB

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M
2/6/2008 10:11:47 AM
Instrument 4
DB-ALC1

40 MIX-CAL-AB
A. Black

vial # 7



#	Compound	Area	RT
1	Methanol	137	0.801
2	Ethanol	296	1.005
3	Isopropanol	581	1.226
4	Acetone	1247	1.476
5	n-Propanol	1729	1.656

Totals:



Methanol 39.755 mg/dL



Ethanol 40.012 mg/dL



Isopropanol 40.646 mg/dL

AB

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M



Acetone 41.535 mg/dL



n-Propanol 1.000 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M
2/6/2008 10:15:01 AM
Instrument 4
DB-ALC1

80 MIX-CAL-AB
A. Black

vial # 8



#	Compound	Area	RT
1	Methanol	276	0.801
2	Ethanol	596	1.005
3	Isopropanol	1164	1.226
4	Acetone	2492	1.476
5	n-Propanol	1740	1.655

Totals:



Methanol 80.003 mg/dL



Ethanol 80.009 mg/dL



Isopropanol 79.938 mg/dL

OB

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\VOL.M



Acetone 79.907 mg/dL



n-Propanol 1.000 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/6/2008 10:27:34 AM
Instrument 4
DB-ALC1

0.04 AB Control
A. Black

vial # 9



#	Compound	Area	RT
1	Ethanol	276	1.003
2	n-Propanol	1595	1.655

Totals:



Ethanol 0.041 g/100ml



n-Propanol 1.000 g/100ml

OB

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/6/2008 10:30:51 AM
Instrument 4
DB-ALC1

0.10 AB Control
A. Black

vial # 10



#	Compound	Area	RT
1	Ethanol	653	1.003
2	n-Propanol	1522	1.655

Totals:



Ethanol 0.103 g/100ml



n-Propanol 1.000 g/100ml

AB

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/6/2008 10:34:07 AM
Instrument 4
DB-ALC1

0.20 AB Control
A. Black

vial # 11



#	Compound	Area	RT
1	Ethanol	1288	1.003
2	n-Propanol	1499	1.655

Totals:



Ethanol 0.205 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/6/2008 10:37:28 AM
Instrument 4
DB-ALC1

BLANK
A. Black

vial # 12



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1504	1.655

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml

DB

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/6/2008 10:40:41 AM
Instrument 4
DB-ALC1

SIM 08006-1
A. Black

vial # 13



#	Compound	Area	RT
1	Ethanol	659	1.002
2	n-Propanol	1509	1.654
Totals:			



Ethanol 0.104 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/6/2008 10:44:02 AM
Instrument 4
DB-ALC1

SIM 08006-2
A. Black

vial # 14



#	Compound	Area	RT
1	Ethanol	651	1.002
2	n-Propanol	1488	1.655

Totals:



Ethanol 0.104 g/100ml



n-Propanol 1.000 g/100ml

AB

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/6/2008 10:49:31 AM
Instrument 4
DB-ALC1

SIM 08006-3
A. Black

vial # 15



#	Compound	Area	RT
1	Ethanol	663	1.003
2	n-Propanol	1512	1.655

Totals:



Ethanol 0.105 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/6/2008 10:52:47 AM
Instrument 4
DB-ALC1

SIM 08006-4
A. Black

vial # 16



#	Compound	Area	RT
1	Ethanol	644	1.002
2	n-Propanol	1477	1.654

Totals:



Ethanol 0.104 g/100ml



n-Propanol 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/6/2008 10:56:06 AM
Instrument 4
DB-ALC1

SIM 08006-5
A. Black

vial # 17



#	Compound	Area	RT
1	Ethanol	669	1.002
2	n-Propanol	1523	1.655

Totals:



Ethanol 0.105 g/100ml



n-Propanol 1.000 g/100ml

AB

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/6/2008 10:59:22 AM
Instrument 4
DB-ALC1

0.10 AB Control
A. Black

vial # 18



#	Compound	Area	RT
1	Ethanol	681	1.003
2	n-Propanol	1581	1.654

Totals:



Ethanol 0.103 g/100ml



n-Propanol 1.000 g/100ml

AB

WASHINGTON STATE TOXICOLOGY LABORATORY

D:\HPCHEM\1\METHODS\BLDALCO.M
2/6/2008 11:02:40 AM
Instrument 4
DB-ALC1

BLANK
A. Black

vial # 19



#	Compound	Area	RT
1	Ethanol	0	0.000
2	n-Propanol	1615	1.655

Totals:



Ethanol 0.000 g/100ml



n-Propanol 1.000 g/100ml