WASHINGTON STATE TOXICOLOGY LABORATORY
SIMULATOR SOLUTION DATA ENTRY REVIEW

Reviewer/ s: Denton / Gullberg Date: 6/5/2008

Location: Tox-Lab - Seattle Solution Batch Number: 08026

	YES	NO	N/A
Preparation date precedes all analysis dates:	☒	☐	☐
Declarations signed and properly dated:	☒	☐	☐
Data entry corresponds to all chromatograms:	☒	☐	☐
All signatures present on Analysis sheet:	☒	☐	☐
Avg. solution concentration correct?:	☒	☐	☐
Standard deviation correct:	☒	☐	☐
Range correct if applicable:	☒	☐	☐
Equivalent vapor concentration correct?	☒	☐	☐
Blank Chromatograms included in file:	☒	☐	☐
External Control information correct: (lot # present and future date)	☒	☐	☐
Complies with accuracy and precision requirements established by the State Toxicologist:	☒	☐	☐
CV% Correct?	☒	☐	☐
Reviewed for outliers per policy and none found?	☐	☒	☐

Comments:

ONE OUTLIER (0.107) FOUND AND REMOVED PER POLICY

Reviewer Signature: R. L. Gullberg Date: 6-11-08

Reviewer Signature: K.L. D[illegible] Date: 6/11/2008

Rev. Date: 2/14/2008

Batch Worksheet Check Off

Please check the data entered into the worksheet is correct and that the date to the right of your name is the date that you tested the solution and then sign the worksheet.

Please initial below to affirm that you have:

1 – Initialed your chromatograms
2 – Checked your data
3 – Checked the date to the right of your name on the worksheet
4 – Signed the worksheet.

	Initials	Date
Amanda Black	AB	6-9-08
Asa Louis	AL	20080609
Brian Capron		
Brianna Peterson	BP	6-10-08
Brianne Akins		
Brittany Ball	BB	6/11/08
Christie Mitchell	CM	6/9/08
Christopher Johnston		
Erin Kolbrich		
Estuardo Miranda	EM	6-10-08
Gwynyth Scherperel		
Justin Knoy	JK	6-9-08
Lisa Noble		
Melissa Pemberton		
Naziha Nuwayhid		
Rebecca Flaherty	RF	6/9/2008
Sarah Swenson		

WASHINGTON STATE TOXICOLOGY LABORATORY
FORENSIC LABORATORY SERVICES BUREAU
WASHINGTON STATE PATROL
2203 AIRPORT WAY S, SUITE 360
SEATTLE, WASHINGTON 98134-2027
(206) 262-6100 FAX (206) 262-6145

SIMULATOR EXTERNAL STANDARD SOLUTION DATABASE

Preparation and Certification of 0.08 g/210L Simulator External Standard Solution

Batch number **08026** Date prepared: 05/27/2008

Preparation: **66.5** mL of absolute ethyl alcohol diluted to **52** Liters with water

Concentration of ethanol (g/100mL) measured by gas chromatography:

	Analyst 1	Analyst 2	Analyst 3	Analyst 4	Analyst 5	Analyst 6	Analyst 7	Analyst 8	Analyst 9	Analyst 10
1	0.099	0.102	0.101	0.101	0.099	0.101	0.101	0.101		
2	0.099	0.101	0.101	0.101	0.098	0.102	0.101	0.101		
3	0.099	0.102	0.102	0.100	0.098	0.101	0.101	0.101		
4	0.099		0.101	0.101	0.098	0.101	0.101	0.101		
5	0.099	0.102	0.102	0.102	0.098	0.101	0.101	0.103		
Ctrl	0.097	0.098	0.099	0.095	0.096	0.099	0.099	0.099		

Statistics:

Avg. solution concent.: 0.1006 g/100 mL

SD: 0.00131

Range (3.3XSD): 0.0963 to 0.1049

Precision CV (%): 1.3044 %

External Control:

Lot #: A056938 Exp date: MM 04 / YYYY 2012

Target concentration: 0.10 g/100mL

Equivalent vapor concent.: 0.0818 g/210L

Analyst	Name	Signature	Date Tested
1	Rebecca Flaherty		05/27/2008
2	Justin L Knoy		05/28/2008
3	Christie Mitchell		05/29/2008
4	Asa Louis		05/29/2008
5	Amanda Black		06/02/2008
6	Brittany Ball	Brittany Ball	06/02/2008
7	Estuardo J. Miranda		06/02/2008
8	Brianna Peterson	Brianna Peterson	06/02/2008
9			
10			

Prepared by: Rebecca Flaherty according to the approved protocol.

Final review by: BC

ver 12/19/2007

WASHINGTON STATE TOXICOLOGY LABORATORY
FORENSIC LABORATORY SERVICES BUREAU
WASHINGTON STATE PATROL
2203 AIRPORT WAY S, SUITE 360
SEATTLE, WASHINGTON 98134-2027
(206) 262-6100 FAX (206) 262-6145

SIMULATOR EXTERNAL STANDARD SOLUTION DATABASE

Preparation and Certification of 0.08 g/210L Simulator External Standard Solution

Batch number **08026** Date prepared: 05/27/2008

Preparation: **66.5** mL of absolute ethyl alcohol diluted to **52** Liters with water

Concentration of ethanol (g/100mL) measured by gas chromatography:

	Analyst 1	Analyst 2	Analyst 3	Analyst 4	Analyst 5	Analyst 6	Analyst 7	Analyst 8	Analyst 9	Analyst 10
1	0.099	0.102	0.101	0.101	0.099	0.101	0.101	0.101		
2	0.099	0.101	0.101	0.101	0.098	0.102	0.101	0.101		
3	0.099	0.102	0.102	0.100	0.098	0.101	0.101	0.101		
4	0.099	~~0.107~~ 10	0.101	0.101	0.098	0.101	0.101	0.101		
5	0.099	0.102	0.102	0.102	0.098	0.101	0.101	0.103		
Ctrl	0.097	0.098	0.099	0.095	0.096	0.099	0.099	0.099		

OUTLIER - REMOVED PER POLICY - NEW DATABASE PRINTED W/ OUTLIER REMOVED.

Statistics:

Avg. solution concent.: 0.1008 g/100 mL

SD: 0.00164

Range (3.3XSD): 0.0954 to 0.1062

Precision CV (%): 1.6317 %

External Control:

Lot #: A056938 Exp date: MM 04 / YYYY 2012

Target concentration: 0.10 g/100mL

Equivalent vapor concent.: 0.0820 g/210L

Analyst	Name	Signature	Date Tested
1	Rebecca Flaherty		05/27/2008
2	Justin L Knoy		05/28/2008
3	Christie Mitchell		05/29/2008
4	Asa Louis		05/29/2008
5	Amanda Black		06/02/2008
6	Brittany Ball		06/02/2008
7	Estuardo J. Miranda		06/02/2008
8	Brianna Peterson		06/02/2008
9			
10			

Prepared by: Rebecca Flaherty according to the approved protocol.

Final review by:_____

ver 12/19/2007

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION FOR LOT 08026

I, Rebecca Flaherty, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degrees in Biochemistry and Psychobiology and MS degree in Forensic Science.

The simulator solution, Lot Number 08026, was prepared in the Washington State Toxicology Laboratory on 5/27/2008. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/27/2009.

Seattle, WA

6/9/2008
Rebecca Flaherty Date
Forensic Toxicologist

RF/ik
RFSIMSOL

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION FOR LOT 08026

I, Justin L. Knoy, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Biology, and MS degree in Forensic Science.

The simulator solution, Lot Number 08026, was prepared in the Washington State Toxicology Laboratory on 5/27/2008. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/27/2009.

Seattle, WA

6-9-08

Justin L. Knoy Date
Forensic Toxicologist

JLK/ik
JKSIMSOL

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION FOR LOT 08026

I, Christie Mitchell, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BA degree in Chemistry and MFS degree in Forensic Science.

The simulator solution, Lot Number 08026, was prepared in the Washington State Toxicology Laboratory on 5/27/2008. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/27/2009.

Seattle, WA

Christie Mitchell 6/9/2008
Christie Mitchell Date
Forensic Toxicologist

CM/ik
CMSIMSOL

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION FOR LOT 08026

I, Asa J. Louis, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: B.S. degree in Biochemistry and eight years of toxicology experience.

The simulator solution, Lot Number 08026, was prepared in the Washington State Toxicology Laboratory on 5/27/2008. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/27/2009.

Seattle, WA

20080609

Asa J. Louis Date
Forensic Toxicologist

AJL/ik
AJLSIMSOL

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION FOR LOT 08026

I, Amanda M. Black, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degrees in Chemistry and Veterinary Science.

The simulator solution, Lot Number 08026, was prepared in the Washington State Toxicology Laboratory on 5/27/2008. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/27/2009.

Seattle, WA

6-9-08

Amanda M. Black Date
Forensic Toxicologist

AB/ik
ABSIMSOL

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION FOR LOT 08026

I, Brittany Ball, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Biology and a Masters in Forensic Science.

The simulator solution, Lot Number 08026, was prepared in the Washington State Toxicology Laboratory on 5/27/2008. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/27/2009

Seattle, WA

Brittany Ball 6/9/08
Brittany Ball Date
Forensic Toxicologist

BB/ik
BBSIMSOL

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION FOR LOT 08026

I, Estuardo J. Miranda, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: Bachelor of Science in Chemistry, Master of Science in Zoology and nine years experience in Forensic Toxicology.

The simulator solution, Lot Number 08026, was prepared in the Washington State Toxicology Laboratory on 5/27/2008. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/27/2009.

Seattle, WA

E Miranda 6-10-08

Estuardo J. Miranda Date
Forensic Toxicologist

EM/ik
EMSIMSOL

CHRISTINE O. GREGOIRE
Governor

JOHN R. BATISTE
Chief

STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY
2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2927 • (206) 262-6100 • FAX (206) 262-6145

DATAMASTER 0.08 SIMULATOR SOLUTION CERTIFICATION FOR LOT 08026

I, Brianna Peterson, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and a part of my responsibilities includes preparing and testing the alcohol solutions for the DataMaster breath test instrument.

I possess the following qualifications: BS degree in Chemistry, MS degree in Forensic Science, Ph.D. degree in Toxicology, and two years of experience in forensic toxicology.

The simulator solution, Lot Number 08026, was prepared in the Washington State Toxicology Laboratory on 5/27/2008. I examined and tested this solution. It was found to conform to those standards established by the state toxicologist for the certification of simulator solution. It should not be used for evidential breath tests after 5/27/2009.

Seattle, WA

Brianna Peterson 6/10/08
Brianna Peterson, Ph.D. Date
Forensic Toxicologist

BP/ik
BPSIMSOL

Sequence: C:\HPCHEM\1\SEQUENCE\RFSIM3.S

Sequence Parameters:

Operator:	Rebecca Flaherty
Data File Naming:	Auto
Data Directory:	C:\HPCHEM\2\DATA\
Data Subdirectory:	080527RF
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	BLDALCO3	1	Sample		
2	Vial 2	0.079 CAL	BLDALCO3	1	Calib		
3	Vial 3	0.158 CAL	BLDALCO3	1	Calib		
4	Vial 4	0.316 CAL	BLDALCO3	1	Calib		
5	Vial 5	Blank-RF	BLDALCO3	1	Ctrl Samp		
6	Vial 6	0.04 Mix	VOL	1	Calib		
7	Vial 7	0.08 Mix	VOL	1	Calib		
8	Vial 8	0.02	BLDALCO3	1	Sample		
9	Vial 9	0.04 Ctrl - rf	BLDALCO3	1	Ctrl Samp	LOT # A056758	EXP 03/2012
10	Vial 10	0.10 Ctrl - rf	BLDALCO3	1	Ctrl Samp	LOT# A056938	EXP 04/2012
11	Vial 11	0.20 Ctrl - rf	BLDALCO3	1	Ctrl Samp	LOT# A055525	EXP 02/2012
12	Vial 12	Blank	BLDALCO3	1	Sample		
13	Vial 13	SIM08026-1	BLDALCO3	1	Sample		
14	Vial 14	SIM08026-2	BLDALCO3	1	Sample		
15	Vial 15	SIM08026-3	BLDALCO3	1	Sample		
16	Vial 16	SIM08026-4	BLDALCO3	1	Sample		
17	Vial 17	SIM08026-5	BLDALCO3	1	Sample		
18	Vial 18	0.10 Ctrl-RF	BLDALCO3	1	Ctrl Samp		
19	Vial 19	BLANK	BLDALCO3	1	Sample		

Calibration Part:

Line	Location	SampleName	Method	CalLev	Update RF	Update RT	Interval
2	Vial 2	0.079 CAL	BLDALCO3	1	Replace	Average	
3	Vial 3	0.158 CAL	BLDALCO3	2	Replace	Average	
4	Vial 4	0.316 CAL	BLDALCO3	3	Replace	Average	
6	Vial 6	0.04 Mix	VOL	1	Replace	Average	
7	Vial 7	0.08 Mix	VOL	2	Replace	Average	

Sequence Table (Back Injector):

No entries - empty table!

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/27/2008 1:32:22 PM
Instrument 3
db-alc2

BLANK
Rebecca Flaherty

vial # 1



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1574	1.825

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/27/2008 1:35:29 PM
Instrument 3
db-alc2

0.079 CAL
Rebecca Flaherty

vial # 2



#	Compound	Area	RT
1	ETHANOL	576	1.064
2	n-PROPANOL	1577	1.825

Totals:



ETHANOL 0.077 g/100ml



n-PROPANOL 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/27/2008 1:38:37 PM
Instrument 3
db-alc2

0.158 CAL
Rebecca Flaherty

vial # 3



#	Compound	Area	RT
1	ETHANOL	1196	1.064
2	n-PROPANOL	1551	1.824

Totals:



ETHANOL 0.163 g/100ml



n-PROPANOL 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/27/2008 1:41:44 PM
Instrument 3
db-alc2

0.316 CAL
Rebecca Flaherty

vial # 4



#	Compound	Area	RT
1	ETHANOL	2296	1.064
2	n-PROPANOL	1557	1.824

Totals:



ETHANOL 0.314 g/100ml



n-PROPANOL 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/27/2008 1:44:51 PM
Instrument 3
db-alc2

Blank-RF
Rebecca Flaherty

vial # 5



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1561	1.824

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M
5/27/2008 1:48:01 PM
Instrument 3
db-alc2

0.04 Mix
Rebecca Flaherty

vial # 6



#	Compound	Area	RT
1	METHANOL	132	0.851
2	ETHANOL	278	1.064
3	ACETONE	1152	1.180
4	ISOPROPANOL	574	1.249
5	n-PROPANOL	1558	1.825

Totals:



METHANOL 40.339 mg/dL



ETHANOL 40.198 mg/dL



ACETONE 39.585 mg/dL

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M



ISOPROPANOL 39.805 mg/dL



n-PROPANOL 1.000 mg/dL

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M
5/27/2008 1:51:08 PM
Instrument 3
db-alc2

0.08 Mix
Rebecca Flaherty

vial # 7



#	Compound	Area	RT
1	METHANOL	261	0.851
2	ETHANOL	550	1.063
3	ACETONE	2301	1.179
4	ISOPROPANOL	1144	1.248
5	n-PROPANOL	1544	1.823

Totals:



METHANOL 79.990 mg/dL



ETHANOL 79.978 mg/dL



ACETONE 80.105 mg/dL

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M



ISOPROPANOL 80.079 mg/dL



n-PROPANOL 1.000 mg/dL

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/27/2008 1:54:21 PM
Instrument 3
db-alc2

0.02
Rebecca Flaherty

vial # 8



#	Compound	Area	RT
1	ETHANOL	141	1.064
2	n-PROPANOL	1562	1.825

Totals:



ETHANOL 0.018 g/100ml



n-PROPANOL 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/27/2008 1:57:28 PM
Instrument 3
db-alc2

0.04 Ctrl - rf
Rebecca Flaherty

vial # 9



#	Compound	Area	RT
1	ETHANOL	277	1.064
2	n-PROPANOL	1550	1.824
Totals:			



ETHANOL 0.037 g/100ml



n-PROPANOL 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/27/2008 2:00:35 PM
Instrument 3
db-alc2

0.10 Ctrl - rf
Rebecca Flaherty

vial # 10



#	Compound	Area	RT
1	ETHANOL	713	1.064
2	n-PROPANOL	1559	1.824

Totals:



ETHANOL 0.096 g/100ml



n-PROPANOL 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/27/2008 2:03:43 PM
Instrument 3
db-alc2

0.20 Ctrl - rf
Rebecca Flaherty

vial # 11



#	Compound	Area	RT
1	ETHANOL	1420	1.065
2	n-PROPANOL	1546	1.826

Totals:



ETHANOL 0.195 g/100ml



n-PROPANOL 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/27/2008 2:06:50 PM
Instrument 3
db-alc2

Blank
Rebecca Flaherty

vial # 12



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1563	1.825

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/27/2008 2:09:57 PM
Instrument 3
db-alc2

SIM08026-1
Rebecca Flaherty

vial # 13



#	Compound	Area	RT
1	ETHANOL	726	1.064
2	n-PROPANOL	1554	1.825

Totals:



ETHANOL 0.099 g/100ml



n-PROPANOL 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/27/2008 2:13:04 PM
Instrument 3
db-alc2

SIM08026-2
Rebecca Flaherty

vial # 14



#	Compound	Area	RT
1	ETHANOL	727	1.064
2	n-PROPANOL	1553	1.825

Totals:



ETHANOL 0.099 g/100ml



n-PROPANOL 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/27/2008 2:16:12 PM
Instrument 3
db-alc2

SIM08026-3
Rebecca Flaherty

vial # 15



#	Compound	Area	RT
1	ETHANOL	734	1.065
2	n-PROPANOL	1569	1.825

Totals:



ETHANOL 0.099 g/100ml



n-PROPANOL 1.000 g/100ml

LF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/27/2008 2:19:19 PM
Instrument 3
db-alc2

SIM08026-4
Rebecca Flaherty

vial # 16



#	Compound	Area	RT
1	ETHANOL	732	1.064
2	n-PROPANOL	1563	1.824

Totals:



ETHANOL 0.099 g/100ml



n-PROPANOL 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/27/2008 2:22:26 PM
Instrument 3
db-alc2

SIM08026-5
Rebecca Flaherty

vial # 17



#	Compound	Area	RT
1	ETHANOL	744	1.064
2	n-PROPANOL	1590	1.825
Totals:			



ETHANOL 0.099 g/100ml



n-PROPANOL 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/27/2008 2:25:33 PM
Instrument 3
db-alc2

0.10 Ctrl-RF
Rebecca Flaherty

vial # 18



#	Compound	Area	RT
1	ETHANOL	737	1.064
2	n-PROPANOL	1612	1.824

Totals:



ETHANOL 0.097 g/100ml



n-PROPANOL 1.000 g/100ml

RF

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/27/2008 2:28:40 PM
Instrument 3
db-alc2

BLANK
Rebecca Flaherty

vial # 19



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1624	1.824

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

RF

Sequence Parameters:

Operator:	Justin Knoy
Data File Naming:	Prefix/Counter
Signal 1 Prefix:	SIG1
Counter:	0001
Signal 2 Prefix:	SIG2
Counter:	0001
Data Directory:	C:\HPCHEM\2\DATA\
Data Subdirectory:	080528JK
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	BLDALCO3	1	Sample		
2	Vial 2	0.079 CAL	BLDALCO3	1	Calib		
3	Vial 3	0.158 CAL	BLDALCO3	1	Calib		
4	Vial 4	0.316 CAL	BLDALCO3	1	Calib		
5	Vial 5	BLANK	BLDALCO3	1	Ctrl Samp		
6	Vial 6	0.04 MIX	VOL3	1	Calib		
7	Vial 7	0.08 MIX	VOL3	1	Calib		
8	Vial 8	0.02 STD	BLDALCO3	1	Sample		
9	Vial 9	0.04 CTRL JK	BLDALCO3	1	Ctrl Samp	A056758	3/2012
10	Vial 10	0.10 CTRL JK	BLDALCO3	1	Ctrl Samp	A056938	4/2012
11	Vial 11	0.20 CTRL JK	BLDALCO3	1	Ctrl Samp	A055525	2/2012
12	Vial 12	BLANK	BLDALCO3	1	Sample		
13	Vial 13	08026-1	BLDALCO3	1	Sample		
14	Vial 14	08026-2	BLDALCO3	1	Sample		
15	Vial 15	08026-3	BLDALCO3	1	Sample		
16	Vial 16	08026-4	BLDALCO3	1	Sample		
17	Vial 17	08026-5	BLDALCO3	1	Sample		
18	Vial 18	0.10 CTRL JK	BLDALCO3	1	Ctrl Samp		
19	Vial 19	BLANK	BLDALCO3	1	Sample		

Calibration Part:

Line	Location	SampleName	Method	CalLev	Update RF	Update RT	Interval
2	Vial 2	0.079 CAL	BLDALCO3	1	Replace	Replace	
3	Vial 3	0.158 CAL	BLDALCO3	2	Replace	Average	
4	Vial 4	0.316 CAL	BLDALCO3	3	Replace	Average	
6	Vial 6	0.04 MIX	VOL3	1	Replace	Replace	
7	Vial 7	0.08 MIX	VOL3	2	Replace	Average	

Sequence Table (Back Injector):

No entries - empty table!

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/28/2008 3:04:21 PM
Instrument 3
db-alc2

BLANK
Justin Knoy

vial # 1



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1573	1.826
Totals:			



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/28/2008 3:07:28 PM
Instrument 3
db-alc2

0.079 CAL
Justin Knoy

vial # 2



#	Compound	Area	RT
1	ETHANOL	575	1.065
2	n-PROPANOL	1562	1.826
Totals:			



ETHANOL 0.079 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/28/2008 3:10:35 PM
Instrument 3
db-alc2

0.158 CAL
Justin Knoy

vial # 3



#	Compound	Area	RT
1	ETHANOL	1152	1.065
2	n-PROPANOL	1557	1.825

Totals:



ETHANOL 0.158 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/28/2008 3:13:43 PM
Instrument 3
db-alc2

0.316 CAL
Justin Knoy

vial # 4



#	Compound	Area	RT
1	ETHANOL	2300	1.065
2	n-PROPANOL	1573	1.827

Totals:



ETHANOL 0.315 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/28/2008 3:16:50 PM
Instrument 3
db-alc2

BLANK
Justin Knoy

vial # 5



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1590	1.827

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL3.M
5/28/2008 3:19:59 PM
Instrument 3
db-alc2

0.04 MIX
Justin Knoy

vial # 6



#	Compound	Area	RT
1	METHANOL	140	0.852
2	ETHANOL	293	1.065
3	ACETONE	1166	1.181
4	ISOPROPANOL	596	1.250
5	n-PROPANOL	1640	1.827

Totals:



METHANOL 40.345 mg/dL



ETHANOL 40.281 mg/dL



ACETONE 39.574 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL3.M



ISOPROPANOL 39.995 mg/dL



n-PROPANOL 1.000 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL3.M
5/28/2008 3:23:06 PM
Instrument 3
db-alc2

0.08 MIX
Justin Knoy

vial # 7



#	Compound	Area	RT
1	METHANOL	263	0.851
2	ETHANOL	554	1.064
3	ACETONE	2309	1.180
4	ISOPROPANOL	1154	1.249
5	n-PROPANOL	1556	1.825

Totals:



METHANOL 79.824 mg/dL



ETHANOL 79.981 mg/dL



ACETONE 80.560 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL3.M



ISOPROPANOL 80.258 mg/dL



n-PROPANOL 1.000 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/28/2008 3:26:17 PM
Instrument 3
db-alc2

0.02 STD
Justin Knoy

vial # 8



#	Compound	Area	RT
1	ETHANOL	147	1.066
2	n-PROPANOL	1614	1.827

Totals:



ETHANOL 0.019 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/28/2008 3:29:25 PM
Instrument 3
db-alc2

0.04 CTRL JK
Justin Knoy

vial # 9



#	Compound	Area	RT
1	ETHANOL	282	1.065
2	n-PROPANOL	1562	1.826
Totals:			



ETHANOL 0.038 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/28/2008 3:32:32 PM
Instrument 3
db-alc2

0.10 CTRL JK
Justin Knoy

vial # 10



#	Compound	Area	RT
1	ETHANOL	754	1.065
2	n-PROPANOL	1636	1.827
Totals:			



ETHANOL 0.099 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/28/2008 3:35:39 PM
Instrument 3
db-alc2

0.20 CTRL JK
Justin Knoy

vial # 11



#	Compound	Area	RT
1	ETHANOL	1445	1.065
2	n-PROPANOL	1573	1.825
Totals:			



ETHANOL 0.198 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/28/2008 3:38:46 PM
Instrument 3
db-alc2

BLANK
Justin Knoy

vial # 12



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1551	1.826
Totals:			



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/28/2008 3:41:54 PM
Instrument 3
db-alc2

08026-1
Justin Knoy

vial # 13



#	Compound	Area	RT
1	ETHANOL	774	1.065
2	n-PROPANOL	1638	1.827
Totals:			



ETHANOL 0.102 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/28/2008 3:45:01 PM
Instrument 3
db-alc2

08026-2
Justin Knoy

vial # 14



#	Compound	Area	RT
1	ETHANOL	732	1.065
2	n-PROPANOL	1554	1.827
Totals:			



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/28/2008 3:48:08 PM
Instrument 3
db-alc2

08026-3
Justin Knoy

vial # 15



#	Compound	Area	RT
1	ETHANOL	784	1.065
2	n-PROPANOL	1655	1.827

Totals:



ETHANOL 0.102 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/28/2008 3:51:15 PM
Instrument 3
db-alc2

08026-4
Justin Knoy

vial # 16



#	Compound	Area	RT
1	ETHANOL	777	1.064
2	n-PROPANOL	1564	1.826

Totals:



ETHANOL 0.107 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/28/2008 3:54:23 PM
Instrument 3
db-alc2

08026-5
Justin Knoy

vial # 17



#	Compound	Area	RT
1	ETHANOL	751	1.064
2	n-PROPANOL	1589	1.824
Totals:			



ETHANOL 0.102 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/28/2008 3:57:30 PM
Instrument 3
db-alc2

0.10 CTRL JK
Justin Knoy

vial # 18



#	Compound	Area	RT
1	ETHANOL	750	1.064
2	n-PROPANOL	1640	1.824
Totals:			



ETHANOL 0.098 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/28/2008 4:00:37 PM
Instrument 3
db-alc2

BLANK
Justin Knoy

vial # 19



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1732	1.827

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

Sequence Parameters:

Operator:	Christie Mitchell
Data File Naming:	Prefix/Counter
Signal 1 Prefix:	SIG1
Counter:	0001
Signal 2 Prefix:	SIG2
Counter:	0001
Data Directory:	C:\HPCHEM\2\DATA\
Data Subdirectory:	080529CM
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile	Lot#	Exp
1	Vial 1	BLANK	BLDALCO3	1	Sample				
2	Vial 2	0.079 CAL	BLDALCO3	1	Calib				
3	Vial 3	0.158 CAL	BLDALCO3	1	Calib				
4	Vial 4	0.316 CAL	BLDALCO3	1	Calib				
5	Vial 5	BLANK-CM	BLDALCO3	1	Ctrl Samp				
6	Vial 6	0.04 MIX	VOL	1	Calib				
7	Vial 7	0.08 MIX	VOL	1	Calib				
8	Vial 8	0.02 STD	BLDALCO3	1	Sample				
9	Vial 9	0.04 CONTROL-CM	BLDALCO3	1	Ctrl Samp			0.04 CM	A056758 3/2012
10	Vial 10	0.10 CONTROL-CM	BLDALCO3	1	Ctrl Samp			0.10 CM	A056938 4/2012
11	Vial 11	0.20 CONTROL-CM	BLDALCO3	1	Ctrl Samp			0.20 CM	A055525 2/2012
12	Vial 12	BLANK	BLDALCO3	1	Sample				
13	Vial 13	SIM 08026-1	BLDALCO3	1	Sample				
14	Vial 14	SIM 08026-2	BLDALCO3	1	Sample				
15	Vial 15	SIM 08026-3	BLDALCO3	1	Sample				
16	Vial 16	SIM 08026-4	BLDALCO3	1	Sample				
17	Vial 17	SIM 08026-5	BLDALCO3	1	Sample				
18	Vial 18	0.10 CONTROL-CM	BLDALCO3	1	Ctrl Samp				
19	Vial 19	BLANK	BLDALCO3	1	Sample				

Calibration Part:

Line	Location	SampleName	Method	CalLev	Update RF	Update RT	Interval
2	Vial 2	0.079 CAL	BLDALCO3	1	Replace	Replace	
3	Vial 3	0.158 CAL	BLDALCO3	2	Replace	Average	
4	Vial 4	0.316 CAL	BLDALCO3	3	Replace	Average	
6	Vial 6	0.04 MIX	VOL	1	Average	Replace	
7	Vial 7	0.08 MIX	VOL	2	Average	Average	

Sequence Table (Back Injector):

No entries - empty table!

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 8:23:08 AM
Instrument 3
db-alc2

BLANK
Christie Mitchell

vial # 1



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1541	1.824

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 8:26:15 AM
Instrument 3
db-alc2

0.079 CAL
Christie Mitchell

vial # 2



#	Compound	Area	RT
1	ETHANOL	571	1.065
2	n-PROPANOL	1539	1.825

Totals:



ETHANOL 0.079 g/100ml



n-PROPANOL 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 8:29:22 AM
Instrument 3
db-alc2

0.158 CAL
Christie Mitchell

vial # 3



#	Compound	Area	RT
1	ETHANOL	1128	1.064
2	n-PROPANOL	1542	1.824

Totals:



ETHANOL 0.158 g/100ml



n-PROPANOL 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 8:32:29 AM
Instrument 3
db-alc2

0.316 CAL
Christie Mitchell

vial # 4



#	Compound	Area	RT
1	ETHANOL	2267	1.065
2	n-PROPANOL	1552	1.824

Totals:



ETHANOL 0.316 g/100ml



n-PROPANOL 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 8:35:37 AM
Instrument 3
db-alc2

BLANK-CM
Christie Mitchell

vial # 5



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1550	1.827

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M
5/29/2008 8:38:47 AM
Instrument 3
db-alc2

0.04 MIX
Christie Mitchell

vial # 6



#	Compound	Area	RT
1	METHANOL	131	0.851
2	ETHANOL	277	1.064
3	ACETONE	1147	1.180
4	ISOPROPANOL	572	1.249
5	n-PROPANOL	1551	1.824

Totals:



METHANOL 40.065 mg/dL



ETHANOL 40.126 mg/dL

CH



ACETONE 39.811 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M



ISOPROPANOL 39.894 mg/dL



n-PROPANOL 1.000 mg/dL

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M
5/29/2008 8:41:55 AM
Instrument 3
db-alc2

0.08 MIX
Christie Mitchell

vial # 7



#	Compound	Area	RT
1	METHANOL	261	0.851
2	ETHANOL	550	1.064
3	ACETONE	2291	1.180
4	ISOPROPANOL	1142	1.249
5	n-PROPANOL	1534	1.824

Totals:



METHANOL 80.499 mg/dL



ETHANOL 80.562 mg/dL

CM



ACETONE 80.293 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M



ISOPROPANOL 80.463 mg/dL



n-PROPANOL 1.000 mg/dL

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 8:45:07 AM
Instrument 3
db-alc2

0.02 STD
Christie Mitchell

vial # 8



#	Compound	Area	RT
1	ETHANOL	142	1.065
2	n-PROPANOL	1519	1.825

Totals:



ETHANOL 0.020 g/100ml



n-PROPANOL 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 8:48:14 AM
Instrument 3
db-alc2

0.04 CONTROL-CM
Christie Mitchell

vial # 9



#	Compound	Area	RT
1	ETHANOL	274	1.065
2	n-PROPANOL	1521	1.826

Totals:



ETHANOL 0.038 g/100ml



n-PROPANOL 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 8:51:21 AM
Instrument 3
db-alc2

0.10 CONTROL-CM
Christie Mitchell

vial # 10



#	Compound	Area	RT
1	ETHANOL	709	1.065
2	n-PROPANOL	1545	1.825

Totals:



ETHANOL 0.099 g/100ml



n-PROPANOL 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 8:54:28 AM
Instrument 3
db-alc2

0.20 CONTROL-CM
Christie Mitchell

vial # 11



#	Compound	Area	RT
1	ETHANOL	1416	1.064
2	n-PROPANOL	1539	1.824

Totals:



ETHANOL 0.199 g/100ml



n-PROPANOL 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 8:57:36 AM
Instrument 3
db-alc2

BLANK
Christie Mitchell

vial # 12



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1561	1.826

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 9:00:43 AM
Instrument 3
db-alc2

SIM 08026-1
Christie Mitchell

vial # 13



#	Compound	Area	RT
1	ETHANOL	726	1.064
2	n-PROPANOL	1548	1.824

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 9:03:50 AM
Instrument 3
db-alc2

SIM 08026-2
Christie Mitchell

vial # 14



#	Compound	Area	RT
1	ETHANOL	727	1.064
2	n-PROPANOL	1546	1.824

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 9:06:57 AM
Instrument 3
db-alc2

SIM 08026-3
Christie Mitchell

vial # 15



#	Compound	Area	RT
1	ETHANOL	728	1.064
2	n-PROPANOL	1541	1.824
Totals:			



ETHANOL 0.102 g/100ml



n-PROPANOL 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 9:10:04 AM
Instrument 3
db-alc2

SIM 08026-4
Christie Mitchell

vial # 16



#	Compound	Area	RT
1	ETHANOL	733	1.065
2	n-PROPANOL	1557	1.826
Totals:			



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 9:13:11 AM
Instrument 3
db-alc2

SIM 08026-5
Christie Mitchell

vial # 17



#	Compound	Area	RT
1	ETHANOL	737	1.065
2	n-PROPANOL	1564	1.825

Totals:



ETHANOL 0.102 g/100ml



n-PROPANOL 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 9:16:19 AM
Instrument 3
db-alc2

0.10 CONTROL-CM
Christie Mitchell

vial # 18



#	Compound	Area	RT
1	ETHANOL	734	1.065
2	n-PROPANOL	1595	1.826

Totals:



ETHANOL 0.099 g/100ml



n-PROPANOL 1.000 g/100ml

CM

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 9:19:26 AM
Instrument 3
db-alc2

BLANK
Christie Mitchell

vial # 19



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1614	1.826

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

CM

Sequence Parameters:

Operator:	alouis
Data File Naming:	Auto
Data Directory:	C:\HPCHEM\2\DATA\
Data Subdirectory:	080529AL
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none

Sequence Comment:
0.04 con a056758 exp 03/2012
0.10 con a056938 exp 04/2012
0.20 con a055525 exp 02/2012

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	blank	BLDALCO3	1	Sample		
2	Vial 2	0.079 std	BLDALCO3	1	Calib		
3	Vial 3	0.158 std	BLDALCO3	1	Calib		
4	Vial 4	0.316 std	BLDALCO3	1	Calib		
5	Vial 5	blank	BLDALCO3	1	Ctrl Samp		
6	Vial 6	0.04 vol	VOL	1	Calib		
7	Vial 7	0.08 vol	VOL	1	Calib		
8	Vial 8	0.02	BLDALCO3	1	Sample		
9	Vial 9	0.04 con al	BLDALCO3	1	Ctrl Samp		
10	Vial 10	0.10 con al	BLDALCO3	1	Ctrl Samp		
11	Vial 11	0.20 con al	BLDALCO3	1	Ctrl Samp		
12	Vial 12	blank	BLDALCO3	1	Sample		
13	Vial 13	08026a	BLDALCO3	1	Sample		
14	Vial 14	08026b	BLDALCO3	1	Sample		
15	Vial 15	08026c	BLDALCO3	1	Sample		
16	Vial 16	08026d	BLDALCO3	1	Sample		
17	Vial 17	08026e	BLDALCO3	1	Sample		
18	Vial 18	0.10 con al	BLDALCO3	1	Ctrl Samp		
19	Vial 19	blank	BLDALCO3	1	Ctrl Samp		

Sequence Table (Back Injector):

No entries - empty table!

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 9:58:19 AM
Instrument 3
db-alc2

blank
alouis

vial # 1



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1657	1.827

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

AC
20080529

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 10:01:26 AM
Instrument 3
db-alc2

0.079 std
alouis

vial # 2



#	Compound	Area	RT
1	ETHANOL	602	1.065
2	n-PROPANOL	1611	1.826

Totals:



ETHANOL 0.080 g/100ml



n-PROPANOL 1.000 g/100ml

AL
20080529

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 10:04:33 AM
Instrument 3
db-alc2

0.158 std
alouis

vial # 3



#	Compound	Area	RT
1	ETHANOL	1150	1.065
2	n-PROPANOL	1574	1.825

Totals:



ETHANOL 0.158 g/100ml



n-PROPANOL 1.000 g/100ml

A
20080529

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 10:07:40 AM
Instrument 3
db-alc2

0.316 std
alouis

vial # 4



#	Compound	Area	RT
1	ETHANOL	2500	1.065
2	n-PROPANOL	1685	1.826

Totals:



ETHANOL 0.317 g/100ml



n-PROPANOL 1.000 g/100ml

AL 20080529

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 10:10:48 AM
Instrument 3
db-alc2

blank
alouis

vial # 5



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1725	1.828

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

N
20080528

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M
5/29/2008 10:13:58 AM
Instrument 3
db-alc2

0.04 vol
alouis

vial # 6



#	Compound	Area	RT
1	METHANOL	149	0.852
2	ETHANOL	308	1.064
3	ACETONE	1185	1.180
4	ISOPROPANOL	615	1.250
5	n-PROPANOL	1695	1.826

Totals:



METHANOL 41.030 mg/dL



ETHANOL 40.578 mg/dL



ACETONE 38.340 mg/dL

A
20080529

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M



ISOPROPANOL 39.438 mg/dL



n-PROPANOL 1.000 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M
5/29/2008 10:17:06 AM
Instrument 3
db-alc2

0.08 vol
alouis

vial # 7



#	Compound	Area	RT
1	METHANOL	285	0.852
2	ETHANOL	593	1.065
3	ACETONE	2340	1.181
4	ISOPROPANOL	1203	1.251
5	n-PROPANOL	1643	1.827

Totals:



METHANOL 79.829 mg/dL



ETHANOL 79.907 mg/dL

A
290529



ACETONE 80.251 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M



ISOPROPANOL 80.122 mg/dL



n-PROPANOL 1.000 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 10:20:18 AM 0.02
Instrument 3 alouis
db-alc2

vial # 8



#	Compound	Area	RT
1	ETHANOL	157	1.066
2	n-PROPANOL	1651	1.828

Totals:



ETHANOL 0.021 g/100ml



n-PROPANOL 1.000 g/100ml

AL
20080528

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 10:23:25 AM
Instrument 3
db-alc2

0.04 con al
alouis

vial # 9



#	Compound	Area	RT
1	ETHANOL	279	1.065
2	n-PROPANOL	1543	1.827

Totals:



ETHANOL 0.039 g/100ml



n-PROPANOL 1.000 g/100ml

AL
20080529

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 10:26:33 AM
Instrument 3
db-alc2

0.10 con al
alouis

vial # 10



#	Compound	Area	RT
1	ETHANOL	716	1.066
2	n-PROPANOL	1600	1.828

Totals:



ETHANOL 0.096 g/100ml



n-PROPANOL 1.000 g/100ml

AC
290805-29

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 10:29:40 AM
Instrument 3
db-alc2

0.20 con al
alouis

vial # 11



#	Compound	Area	RT
1	ETHANOL	1358	1.065
2	n-PROPANOL	1526	1.825

Totals:



ETHANOL 0.190 g/100ml



n-PROPANOL 1.000 g/100ml

A 20080529

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 10:32:47 AM
Instrument 3
db-alc2

blank
alouis

vial # 12



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1625	1.826

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

AL 20080529

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 10:35:54 AM
Instrument 3
db-alc2

08026a
alouis

vial # 13



#	Compound	Area	RT
1	ETHANOL	768	1.065
2	n-PROPANOL	1620	1.827

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

AL 20080529

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 10:39:02 AM
Instrument 3
db-alc2

08026b
alouis

vial # 14



#	Compound	Area	RT
1	ETHANOL	775	1.065
2	n-PROPANOL	1634	1.826

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

AR
20080529

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 10:42:09 AM 08026c
Instrument 3 alouis
db-alc2

vial # 15



#	Compound	Area	RT
1	ETHANOL	724	1.064
2	n-PROPANOL	1541	1.824

Totals:



ETHANOL 0.100 g/100ml



n-PROPANOL 1.000 g/100ml

AL
20080529

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 10:45:16 AM
Instrument 3
db-alc2

08026d
alouis

vial # 16



#	Compound	Area	RT
1	ETHANOL	788	1.065
2	n-PROPANOL	1671	1.827

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

AL 20080529

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 10:48:23 AM
Instrument 3
db-alc2

08026e
alouis

vial # 17



#	Compound	Area	RT
1	ETHANOL	878	1.065
2	n-PROPANOL	1839	1.827

Totals:



ETHANOL 0.102 g/100ml



n-PROPANOL 1.000 g/100ml

AL
20080529

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 10:51:30 AM
Instrument 3
db-alc2

0.10 con al
alouis

vial # 18



#	Compound	Area	RT
1	ETHANOL	707	1.064
2	n-PROPANOL	1592	1.824

Totals:



ETHANOL 0.095 g/100ml



n-PROPANOL 1.000 g/100ml

R
200805-29

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
5/29/2008 10:54:37 AM
Instrument 3
db-alc2

blank
alouis

vial # 19



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1615	1.825

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

AL
20080529

Sequence Parameters:

Operator:	A. Black
Data File Naming:	Prefix/Counter
Signal 1 Prefix:	SIG1
Counter:	0001
Signal 2 Prefix:	SIG2
Counter:	0001
Data Directory:	C:\HPCHEM\2\DATA\
Data Subdirectory:	080602AB
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	BLDALCO3	1	Sample		
2	Vial 2	0.079 CAL	BLDALCO3	1	Calib		
3	Vial 3	0.158 CAL	BLDALCO3	1	Calib		
4	Vial 4	0.316 CAL	BLDALCO3	1	Calib		
5	Vial 5	BLANK CTL	BLDALCO3	1	Ctrl Samp		
6	Vial 6	0.02 STD	BLDALCO3	1	Sample		
7	Vial 7	0.04 MIX	VOL	1	Calib		
8	Vial 8	0.08 MIX	VOL	1	Calib		
9	Vial 9	0.04 CONTROL-AB	BLDALCO3	1	Ctrl Samp	A056758	3/2012
10	Vial 10	0.10 CONTROL-AB	BLDALCO3	1	Ctrl Samp	A056938	4/2012
11	Vial 11	0.20 CONTROL-AB	BLDALCO3	1	Ctrl Samp	A055525	2/2012
12	Vial 12	BLANK	BLDALCO3	1	Sample		
13	Vial 13	SIM 08026-1	BLDALCO3	1	Sample		
14	Vial 14	SIM 08026-2	BLDALCO3	1	Sample		
15	Vial 15	SIM 08026-3	BLDALCO3	1	Sample		
16	Vial 16	SIM 08026-4	BLDALCO3	1	Sample		
17	Vial 17	SIM 08026-5	BLDALCO3	1	Sample		
18	Vial 18	0.10 CONTROL-AB	BLDALCO3	1	Ctrl Samp		
19	Vial 19	BLANK	BLDALCO3	1	Sample		

Calibration Part:

Line	Location	SampleName	Method	CalLev	Update RF	Update RT	Interval
2	Vial 2	0.079 CAL	BLDALCO3	1	Replace	Replace	
3	Vial 3	0.158 CAL	BLDALCO3	2	Replace	Average	
4	Vial 4	0.316 CAL	BLDALCO3	3	Replace	Average	
7	Vial 7	0.04 MIX	VOL	1	Replace	Replace	
8	Vial 8	0.08 MIX	VOL	2	Replace	Average	

Sequence Table (Back Injector):

No entries - empty table!

OB

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 8:18:17 AM
Instrument 3
db-alc2

BLANK
A. Black

vial # 1



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1565	1.825

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

OB

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 8:21:24 AM
Instrument 3
db-alc2

0.079 CAL
A. Black

vial # 2



#	Compound	Area	RT
1	ETHANOL	628	1.065
2	n-PROPANOL	1672	1.825

Totals:



ETHANOL 0.082 g/100ml



n-PROPANOL 1.000 g/100ml

06

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 8:24:31 AM
Instrument 3
db-alc2

0.158 CAL
A. Black

vial # 3



#	Compound	Area	RT
1	ETHANOL	1242	1.065
2	n-PROPANOL	1671	1.826

Totals:



ETHANOL 0.161 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 8:27:38 AM
Instrument 3
db-alc2

0.316 CAL
A. Black

vial # 4



#	Compound	Area	RT
1	ETHANOL	2436	1.065
2	n-PROPANOL	1643	1.825

Totals:



ETHANOL 0.316 g/100ml



n-PROPANOL 1.000 g/100ml

OB

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 8:30:45 AM
Instrument 3
db-alc2

BLANK CTL
A. Black

vial # 5



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1578	1.824

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

GB

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 8:33:52 AM
Instrument 3
db-alc2

0.02 STD
A. Black

vial # 6



#	Compound	Area	RT
1	ETHANOL	141	1.064
2	n-PROPANOL	1580	1.824

Totals:



ETHANOL 0.019 g/100ml



n-PROPANOL 1.000 g/100ml

AB

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M
6/2/2008 8:37:02 AM
Instrument 3
db-alc2

0.04 MIX
A. Black

vial # 7



#	Compound	Area	RT
1	METHANOL	130	0.851
2	ETHANOL	271	1.063
3	ACETONE	1141	1.179
4	ISOPROPANOL	564	1.247
5	n-PROPANOL	1563	1.823

Totals:



METHANOL 39.671 mg/dL



ETHANOL 39.684 mg/dL



ACETONE 41.114 mg/dL

OB

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M



ISOPROPANOL 40.238 mg/dL



n-PROPANOL 1.000 mg/dL

OB

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M
6/2/2008 8:40:09 AM
Instrument 3
db-alc2

0.08 MIX
A. Black

vial # 8



#	Compound	Area	RT
1	METHANOL	261	0.851
2	ETHANOL	549	1.064
3	ACETONE	2271	1.179
4	ISOPROPANOL	1132	1.248
5	n-PROPANOL	1565	1.823

Totals:



METHANOL 80.006 mg/dL



ETHANOL 80.168 mg/dL



ACETONE 79.925 mg/dL

OS

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M



ISOPROPANOL 80.038 mg/dL



n-PROPANOL 1.000 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 8:43:20 AM
Instrument 3
db-alc2

0.04 CONTROL-AB
A. Black

vial # 9



#	Compound	Area	RT
1	ETHANOL	276	1.064
2	n-PROPANOL	1553	1.825

Totals:



ETHANOL 0.037 g/100ml



n-PROPANOL 1.000 g/100ml

OB

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 8:46:27 AM
Instrument 3
db-alc2

0.10 CONTROL-AB
A. Black

vial # 10



#	Compound	Area	RT
1	ETHANOL	715	1.064
2	n-PROPANOL	1581	1.825

Totals:



ETHANOL 0.096 g/100ml



n-PROPANOL 1.000 g/100ml

AB

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 8:49:34 AM
Instrument 3
db-alc2

0.20 CONTROL-AB
A. Black

vial # 11



#	Compound	Area	RT
1	ETHANOL	1414	1.064
2	n-PROPANOL	1563	1.824

Totals:



ETHANOL 0.193 g/100ml



n-PROPANOL 1.000 g/100ml

OB

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 8:52:41 AM
Instrument 3
db-alc2

BLANK
A. Black

vial # 12



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1562	1.824

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 8:55:48 AM
Instrument 3
db-alc2

SIM 08026-1
A. Black

vial # 13



#	Compound	Area	RT
1	ETHANOL	743	1.064
2	n-PROPANOL	1600	1.826

Totals:



ETHANOL 0.099 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 8:58:55 AM
Instrument 3
db-alc2

SIM 08026-2
A. Black

vial # 14



#	Compound	Area	RT
1	ETHANOL	736	1.064
2	n-PROPANOL	1597	1.824

Totals:



ETHANOL 0.098 g/100ml



n-PROPANOL 1.000 g/100ml

AB

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 9:02:02 AM
Instrument 3
db-alc2

SIM 08026-3
A. Black

vial # 15



#	Compound	Area	RT
1	ETHANOL	727	1.064
2	n-PROPANOL	1581	1.823

Totals:



ETHANOL 0.098 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 9:05:09 AM
Instrument 3
db-alc2

SIM 08026-4
A. Black

vial # 16



#	Compound	Area	RT
1	ETHANOL	729	1.064
2	n-PROPANOL	1578	1.824

Totals:



ETHANOL 0.098 g/100ml



n-PROPANOL 1.000 g/100ml

QB

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 9:08:17 AM
Instrument 3
db-alc2

SIM 08026-5
A. Black

vial # 17



#	Compound	Area	RT
1	ETHANOL	737	1.064
2	n-PROPANOL	1594	1.824

Totals:



ETHANOL 0.098 g/100ml



n-PROPANOL 1.000 g/100ml

CB

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 9:11:24 AM
Instrument 3
db-alc2

0.10 CONTROL-AB
A. Black

vial # 18



#	Compound	Area	RT
1	ETHANOL	745	1.064
2	n-PROPANOL	1649	1.824

Totals:



ETHANOL 0.096 g/100ml



n-PROPANOL 1.000 g/100ml

CB

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 9:14:31 AM
Instrument 3
db-alc2

BLANK
A. Black

vial # 19



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1648	1.824

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

03

Sequence Parameters:

Operator:	BRITTANY BALL
Data File Naming:	Prefix/Counter
Signal 1 Prefix:	SIG1
Counter:	0001
Signal 2 Prefix:	SIG2
Counter:	0001
Data Directory:	C:\HPCHEM\2\DATA\
Data Subdirectory:	080602BB
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	BLDALCO3	1	Sample		
2	Vial 2	0.079 CAL	BLDALCO3	1	Calib		
3	Vial 3	0.158 CAL	BLDALCO3	1	Calib		
4	Vial 4	0.316 CAL	BLDALCO3	1	Calib		
5	Vial 5	BLANK	BLDALCO3	1	Ctrl Samp		
6	Vial 6	0.04 MIX	VOL	1	Calib		
7	Vial 7	0.08 MIX	VOL	1	Calib		
8	Vial 8	0.02 STD	BLDALCO3	1	Sample		
9	Vial 9	0.04 CONTROL-BB	BLDALCO3	1	Ctrl Samp	A056758	exp 3/2012
10	Vial 10	0.10 CONTROL-BB	BLDALCO3	1	Ctrl Samp	A056938	exp 4/2012
11	Vial 11	0.20 CONTROL-BB	BLDALCO3	1	Ctrl Samp	A055525	exp 2/2012
12	Vial 12	BLANK	BLDALCO3	1	Sample		
13	Vial 13	sim08026-1bb	BLDALCO3	1	Sample		
14	Vial 14	sim08026-2bb	BLDALCO3	1	Sample		
15	Vial 15	sim08026-3bb	BLDALCO3	1	Sample		
16	Vial 16	sim08026-4bb	BLDALCO3	1	Sample		
17	Vial 17	sim08026-5bb	BLDALCO3	1	Sample		
18	Vial 18	0.10 CONTROL-BB	BLDALCO3	1	Ctrl Samp		
19	Vial 19	BLANK	BLDALCO3	1	Sample		

Calibration Part:

Line	Location	SampleName	Method	CalLev	Update RF	Update RT	Interval
2	Vial 2	0.079 CAL	BLDALCO3	1	Replace	Replace	
3	Vial 3	0.158 CAL	BLDALCO3	2	Replace	Average	
4	Vial 4	0.316 CAL	BLDALCO3	3	Replace	Average	
6	Vial 6	0.04 MIX	VOL	1	Replace	Replace	
7	Vial 7	0.08 MIX	VOL	2	Replace	Average	

BB 6/2/08

Sequence Table (Back Injector):

No entries - empty table!

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 10:01:17 AM
Instrument 3
db-alc2

BLANK
BRITTANY BALL

vial # 1



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1542	1.826

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

BB
6/2/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 10:04:24 AM
Instrument 3
db-alc2

0.079 CAL
BRITTANY BALL

vial # 2



#	Compound	Area	RT
1	ETHANOL	573	1.064
2	n-PROPANOL	1517	1.824

Totals:



ETHANOL 0.080 g/100ml



n-PROPANOL 1.000 g/100ml

BB
6/2/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 10:07:31 AM
Instrument 3
db-alc2

0.158 CAL
BRITTANY BALL

vial # 3



#	Compound	Area	RT
1	ETHANOL	1133	1.064
2	n-PROPANOL	1512	1.825

Totals:



ETHANOL 0.159 g/100ml



n-PROPANOL 1.000 g/100ml

BB
6/2/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 10:10:38 AM
Instrument 3
db-alc2

0.316 CAL
BRITTANY BALL

vial # 4



#	Compound	Area	RT
1	ETHANOL	2378	1.065
2	n-PROPANOL	1576	1.826

Totals:



ETHANOL 0.316 g/100ml



n-PROPANOL 1.000 g/100ml

BB
6/2/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 10:13:45 AM
Instrument 3
db-alc2

BLANK
BRITTANY BALL

vial # 5



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1546	1.826

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

BB
6/2/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M
6/2/2008 10:16:55 AM
Instrument 3
db-alc2

0.04 MIX
BRITTANY BALL

vial # 6



#	Compound	Area	RT
1	METHANOL	130	0.850
2	ETHANOL	273	1.063
3	ACETONE	1114	1.179
4	ISOPROPANOL	565	1.247
5	n-PROPANOL	1502	1.823

Totals:



METHANOL 41.124 mg/dL



ETHANOL 41.012 mg/dL

BB
6/2/08



ACETONE 40.581 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M



ISOPROPANOL 41.086 mg/dL



n-PROPANOL 1.000 mg/dL

BB
6/2/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M
6/2/2008 10:20:02 AM
Instrument 3
db-alc2

0.08 MIX
BRITTANY BALL

vial # 7



#	Compound	Area	RT
1	METHANOL	260	0.850
2	ETHANOL	548	1.063
3	ACETONE	2246	1.179
4	ISOPROPANOL	1139	1.247
5	n-PROPANOL	1502	1.822

Totals:



METHANOL 79.984 mg/dL



ETHANOL 80.034 mg/dL

BB
6/2/08



ACETONE 80.111 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M



ISOPROPANOL 80.098 mg/dL



n-PROPANOL 1.000 mg/dL

BB
6/2/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 10:23:13 AM
Instrument 3
db-alc2

0.02 STD
BRITTANY BALL

vial # 8



#	Compound	Area	RT
1	ETHANOL	140	1.064
2	n-PROPANOL	1497	1.824

Totals:



ETHANOL 0.020 g/100ml



n-PROPANOL 1.000 g/100ml

BB
6/2/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 10:26:21 AM
Instrument 3
db-alc2

0.04 CONTROL-BB
BRITTANY BALL

vial # 9



#	Compound	Area	RT
1	ETHANOL	300	1.065
2	n-PROPANOL	1584	1.827

Totals:



ETHANOL 0.040 g/100ml



n-PROPANOL 1.000 g/100ml

BB
6/2/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 10:29:28 AM
Instrument 3
db-alc2

0.10 CONTROL-BB
BRITTANY BALL

vial # 10



#	Compound	Area	RT
1	ETHANOL	768	1.065
2	n-PROPANOL	1611	1.826

Totals:



ETHANOL 0.100 g/100ml



n-PROPANOL 1.000 g/100ml

BB
6/2/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 10:32:35 AM
Instrument 3
db-alc2

0.20 CONTROL-BB
BRITTANY BALL

vial # 11



#	Compound	Area	RT
1	ETHANOL	1423	1.064
2	n-PROPANOL	1504	1.824

Totals:



ETHANOL 0.198 g/100ml



n-PROPANOL 1.000 g/100ml

BB
6/2/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 10:35:43 AM
Instrument 3
db-alc2

BLANK
BRITTANY BALL

vial # 12



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1522	1.825

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

BB
6/2/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 10:38:50 AM
Instrument 3
db-alc2

sim08026-1bb
BRITTANY BALL

vial # 13



#	Compound	Area	RT
1	ETHANOL	729	1.064
2	n-PROPANOL	1520	1.825

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

BB
6/2/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 10:41:57 AM
Instrument 3
db-alc2

sim08026-2bb
BRITTANY BALL

vial # 14



#	Compound	Area	RT
1	ETHANOL	789	1.065
2	n-PROPANOL	1626	1.826

Totals:



ETHANOL 0.102 g/100ml



n-PROPANOL 1.000 g/100ml

BB 6/2/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 10:45:04 AM
Instrument 3
db-alc2

sim08026-3bb
BRITTANY BALL

vial # 15



#	Compound	Area	RT
1	ETHANOL	764	1.064
2	n-PROPANOL	1580	1.826

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

BB 6/2/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 10:48:12 AM
Instrument 3
db-alc2

sim08026-4bb
BRITTANY BALL

vial # 16



#	Compound	Area	RT
1	ETHANOL	755	1.065
2	n-PROPANOL	1565	1.826

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

BB
6/12/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 10:51:19 AM
Instrument 3
db-alc2

sim08026-5bb
BRITTANY BALL

vial # 17



#	Compound	Area	RT
1	ETHANOL	740	1.065
2	n-PROPANOL	1539	1.827

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

BB
6/2/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 10:54:26 AM
Instrument 3
db-alc2

0.10 CONTROL-BB
BRITTANY BALL

vial # 18



#	Compound	Area	RT
1	ETHANOL	739	1.065
2	n-PROPANOL	1569	1.825

Totals:



ETHANOL 0.099 g/100ml



n-PROPANOL 1.000 g/100ml

BB
6/2/08

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 10:57:33 AM
Instrument 3
db-alc2

BLANK
BRITTANY BALL

vial # 19



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1582	1.824

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

BB
6/2/08

EM

Sequence Parameters:

Operator:	Estuardo J. Miranda
Data File Naming:	Auto
Data Directory:	C:\HPCHEM\2\DATA\
Data Subdirectory:	080602EM
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	BLDALCO3	1	Sample		
2	Vial 2	0.079 CAL	BLDALCO3	1	Calib		
3	Vial 3	0.158 CAL	BLDALCO3	1	Calib		
4	Vial 4	0.316 CAL	BLDALCO3	1	Calib		
5	Vial 5	NEGATIVE EM	BLDALCO3	1	Ctrl Samp		
6	Vial 6	0.04 MIX	VOL3	1	Calib		
7	Vial 7	0.08 MIX	VOL3	1	Calib		
8	Vial 8	0.020 ~~Control~~ Std. EM	BLDALCO3	1	Sample		
9	Vial 9	0.040 Control EM	BLDALCO3	1	Ctrl Samp		
10	Vial 10	0.100 Control EM	BLDALCO3	1	Ctrl Samp		
11	Vial 11	0.200 Control EM	BLDALCO3	1	Ctrl Samp		
12	Vial 12	Blank	BLDALCO3	1	Sample		
13	Vial 13	Sim Sol 08026-1	BLDALCO3	1	Sample		
14	Vial 14	Sim Sol 08026-2	BLDALCO3	1	Sample		
15	Vial 15	Sim Sol 08026-3	BLDALCO3	1	Sample		
16	Vial 16	Sim Sol 08026-4	BLDALCO3	1	Sample		
17	Vial 17	Sim Sol 08026-5	BLDALCO3	1	Sample		
18	Vial 18	0.100 Control EM	BLDALCO3	1	Ctrl Samp		
19	Vial 19	Blank	BLDALCO3	1	Sample		

0.04 Control
Lot# A056758
exp: 03/2012

0.10 Control
Lot# A056938
exp: 04/2012

0.20 Control
Lot# A055525
exp: 02/2012

Calibration Part:

Line	Location	SampleName	Method	CalLev	Update RF	Update RT	Interval
2	Vial 2	0.079 CAL	BLDALCO3	1	Replace	Average	
3	Vial 3	0.158 CAL	BLDALCO3	2	Replace	Average	
4	Vial 4	0.316 CAL	BLDALCO3	3	Replace	Average	
6	Vial 6	0.04 MIX	VOL3	1	Replace	Average	
7	Vial 7	0.08 MIX	VOL3	2	Replace	Average	

Sequence Table (Back Injector):

No entries - empty table!

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 12:33:58 PM
Instrument 3
db-alc2

BLANK
Estuardo J. Miranda

EM

vial # 1



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1507	1.824

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 12:37:06 PM
Instrument 3
db-alc2

0.079 CAL
Estuardo J. Miranda EM

vial # 2



#	Compound	Area	RT
1	ETHANOL	579	1.064
2	n-PROPANOL	1528	1.824

Totals:



ETHANOL 0.079 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 12:40:13 PM
Instrument 3
db-alc2

0.158 CAL
Estuardo J. Miranda EM

vial # 3



#	Compound	Area	RT
1	ETHANOL	1128	1.064
2	n-PROPANOL	1509	1.823

Totals:



ETHANOL 0.157 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 12:43:20 PM
Instrument 3
db-alc2

0.316 CAL
Estuardo J. Miranda EM

vial # 4



#	Compound	Area	RT
1	ETHANOL	2276	1.065
2	n-PROPANOL	1508	1.826

Totals:



ETHANOL 0.316 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 12:46:27 PM
Instrument 3
db-alc2

NEGATIVE EM
Estuardo J. Miranda EM

vial # 5



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1504	1.824

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL3.M
6/2/2008 12:49:37 PM
Instrument 3
db-alc2

0.04 MIX
Estuardo J. Miranda EM

vial # 6



#	Compound	Area	RT
1	METHANOL	130	0.851
2	ETHANOL	273	1.063
3	ACETONE	1119	1.179
4	ISOPROPANOL	565	1.247
5	n-PROPANOL	1504	1.823

Totals:



METHANOL 41.109 mg/dL



ETHANOL 40.947 mg/dL



ACETONE 40.517 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL3.M

EM



ISOPROPANOL 40.887 mg/dL



n-PROPANOL 1.000 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL3.M
6/2/2008 12:52:44 PM
Instrument 3
db-alc2

0.08 MIX
Estuardo J. Miranda EJM

vial # 7



#	Compound	Area	RT
1	METHANOL	260	0.850
2	ETHANOL	548	1.063
3	ACETONE	2249	1.179
4	ISOPROPANOL	1137	1.247
5	n-PROPANOL	1503	1.822

Totals:



METHANOL 80.013 mg/dL



ETHANOL 80.061 mg/dL



ACETONE 80.084 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL3.M



ISOPROPANOL 80.107 mg/dL



n-PROPANOL 1.000 mg/dL

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 12:55:55 PM
Instrument 3
db-alc2

0.020 ~~Control~~ std. EM
Estuardo J. Miranda

EM

vial # 8



#	Compound	Area	RT
1	ETHANOL	141	1.064
2	n-PROPANOL	1506	1.824

Totals:



ETHANOL 0.020 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 12:59:02 PM
Instrument 3
db-alc2

0.040 Control EM
Estuardo J. Miranda

vial # 9



#	Compound	Area	RT
1	ETHANOL	281	1.064
2	n-PROPANOL	1526	1.824

Totals:



ETHANOL 0.039 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 1:02:09 PM
Instrument 3
db-alc2

0.100 Control EM
Estuardo J. Miranda

EM

vial # 10



#	Compound	Area	RT
1	ETHANOL	709	1.064
2	n-PROPANOL	1506	1.825

Totals:



ETHANOL 0.099 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 1:05:16 PM
Instrument 3
db-alc2

0.200 Control EM
Estuardo J. Miranda

EM

vial # 11



#	Compound	Area	RT
1	ETHANOL	1420	1.063
2	n-PROPANOL	1504	1.823

Totals:



ETHANOL 0.198 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 1:08:23 PM
Instrument 3
db-alc2

Blank
Estuardo J. Miranda

vial # 12



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1507	1.823

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 1:11:30 PM
Instrument 3
db-alc2

Sim Sol 08026-1
Estuardo J. Miranda

vial # 13



#	Compound	Area	RT
1	ETHANOL	717	1.064
2	n-PROPANOL	1497	1.825

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 1:14:38 PM
Instrument 3
db-alc2

Sim Sol 08026-2
Estuardo J. Miranda

vial # 14



#	Compound	Area	RT
1	ETHANOL	726	1.064
2	n-PROPANOL	1514	1.824

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 1:17:45 PM
Instrument 3
db-alc2

Sim Sol 08026-3
Estuardo J. Miranda

vial # 15



#	Compound	Area	RT
1	ETHANOL	729	1.063
2	n-PROPANOL	1516	1.823

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 1:20:52 PM
Instrument 3
db-alc2

Sim Sol 08026-4
Estuardo J. Miranda

EM

vial # 16



#	Compound	Area	RT
1	ETHANOL	729	1.064
2	n-PROPANOL	1517	1.824

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 1:24:04 PM
Instrument 3
db-alc2

Sim Sol 08026-5
Estuardo J. Miranda

EM

vial # 17



#	Compound	Area	RT
1	ETHANOL	738	1.064
2	n-PROPANOL	1529	1.824

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 1:27:06 PM
Instrument 3
db-alc2

0.100 Control EM
Estuardo J. Miranda

EM

vial # 18



#	Compound	Area	RT
1	ETHANOL	745	1.065
2	n-PROPANOL	1579	1.825

Totals:



ETHANOL 0.099 g/100ml



n-PROPANOL 1.000 g/100ml

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 1:30:13 PM
Instrument 3
db-alc2

Blank
Estuardo J. Miranda EM

vial # 19



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1590	1.825

Totals:



ETHANOL 0.000 g/100ml ✓



n-PROPANOL 1.000 g/100ml

Sequence Parameters:

Operator:	Brianna Peterson
Data File Naming:	Auto
Data Directory:	C:\HPCHEM\2\DATA\
Data Subdirectory:	080602BP
Part of Methods to run:	According to Runtime Checklist
Barcode Reader:	not used
Shutdown Cmd/Macro:	none
Sequence Comment:	

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	Blank	BLDALCO3	1	Sample		
2	Vial 2	0.079 std	BLDALCO3	1	Calib		
3	Vial 3	0.158 std	BLDALCO3	1	Calib		
4	Vial 4	0.316 std	BLDALCO3	1	Calib		
5	Vial 5	blank BP	BLDALCO3	1	Ctrl Samp		
6	Vial 6	0.04 vol	VOL	1	Calib		
7	Vial 7	0.08 vol	VOL	1	Calib		
8	Vial 8	0.02 std	BLDALCO3	1	Sample		
9	Vial 9	0.04 con BP	BLDALCO3	1	Ctrl Samp		
10	Vial 10	0.10 con BP	BLDALCO3	1	Ctrl Samp		
11	Vial 11	0.20 con BP	BLDALCO3	1	Ctrl Samp		
12	Vial 12	blank	BLDALCO3	1	Sample		
13	Vial 13	SIM 08026-1	BLDALCO3	1	Sample		
14	Vial 14	SIM 08026-2	BLDALCO3	1	Sample		
15	Vial 15	SIM 08026-3	BLDALCO3	1	Sample		
16	Vial 16	SIM 08026-4	BLDALCO3	1	Sample		
17	Vial 17	SIM 08026-5	BLDALCO3	1	Sample		
18	Vial 18	0.10 con BP	BLDALCO3	1	Ctrl Samp		
19	Vial 19	Blank	BLDALCO3	1	Sample		

0.04 Control
Lot# A056758
exp: 03/2012

0.10 Control
Lot# A056938
exp: 04/2012

0.20 Control
Lot# A055525
exp: 02/2012

Sequence Table (Back Injector):

No entries - empty table!

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 1:48:13 PM
Instrument 3
db-alc2

Blank
Brianna Peterson

vial # 1



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1490	1.825

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 1:51:21 PM
Instrument 3
db-alc2

0.079 std
Brianna Peterson

vial # 2



#	Compound	Area	RT
1	ETHANOL	569	1.064
2	n-PROPANOL	1491	1.825

Totals:



ETHANOL 0.080 g/100ml



n-PROPANOL 1.000 g/100ml

BR

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 1:54:28 PM
Instrument 3
db-alc2

0.158 std
Brianna Peterson

vial # 3



#	Compound	Area	RT
1	ETHANOL	1130	1.065
2	n-PROPANOL	1494	1.825

Totals:



ETHANOL 0.158 g/100ml



n-PROPANOL 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 1:57:35 PM
Instrument 3
db-alc2

0.316 std
Brianna Peterson

vial # 4



#	Compound	Area	RT
1	ETHANOL	2261	1.064
2	n-PROPANOL	1489	1.824

Totals:



ETHANOL 0.316 g/100ml



n-PROPANOL 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 2:00:42 PM
Instrument 3
db-alc2

blank BP
Brianna Peterson

vial # 5



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1481	1.825

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M
6/2/2008 2:03:50 PM
Instrument 3
db-alc2

0.04 vol
Brianna Peterson

vial # 6



#	Compound	Area	RT
1	METHANOL	137	0.851
2	ETHANOL	285	1.064
3	ACETONE	1127	1.180
4	ISOPROPANOL	581	1.249
5	n-PROPANOL	1538	1.825

Totals:



METHANOL 40.725 mg/dL



ETHANOL 40.421 mg/dL



ACETONE 39.460 mg/dL

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M



ISOPROPANOL 39.890 mg/dL



n-PROPANOL 1.000 mg/dL

BJ

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M
6/2/2008 2:06:58 PM
Instrument 3
db-alc2

0.08 vol
Brianna Peterson

vial # 7



#	Compound	Area	RT
1	METHANOL	258	0.850
2	ETHANOL	544	1.063
3	ACETONE	2218	1.178
4	ISOPROPANOL	1130	1.247
5	n-PROPANOL	1470	1.822

Totals:



METHANOL 79.835 mg/dL



ETHANOL 79.972 mg/dL



ACETONE 80.392 mg/dL

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\VOL.M



ISOPROPANOL 80.238 mg/dL



n-PROPANOL 1.000 mg/dL

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 2:10:06 PM
Instrument 3
db-alc2

0.02 std
Brianna Peterson

vial # 8



#	Compound	Area	RT
1	ETHANOL	141	1.063
2	n-PROPANOL	1490	1.823

Totals:



ETHANOL 0.020 g/100ml



n-PROPANOL 1.000 g/100ml

Bol

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 2:13:13 PM
Instrument 3
db-alc2

0.04 con BP
Brianna Peterson

vial # 9



#	Compound	Area	RT
1	ETHANOL	280	1.064
2	n-PROPANOL	1500	1.824

Totals:



ETHANOL 0.039 g/100ml



n-PROPANOL 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 2:16:22 PM
Instrument 3
db-alc2

0.10 con BP
Brianna Peterson

vial # 10



#	Compound	Area	RT
1	ETHANOL	729	1.064
2	n-PROPANOL	1525	1.826

Totals:



ETHANOL 0.100 g/100ml



n-PROPANOL 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 2:19:28 PM
Instrument 3
db-alc2

0.20 con BP
Brianna Peterson

vial # 11



#	Compound	Area	RT
1	ETHANOL	1413	1.064
2	n-PROPANOL	1485	1.823

Totals:



ETHANOL 0.198 g/100ml



n-PROPANOL 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 2:22:35 PM
Instrument 3
db-alc2

blank
Brianna Peterson

vial # 12



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1502	1.826

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 2:25:42 PM
Instrument 3
db-alc2

SIM 08026-1
Brianna Peterson

vial # 13



#	Compound	Area	RT
1	ETHANOL	724	1.064
2	n-PROPANOL	1496	1.823

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 2:28:50 PM
Instrument 3
db-alc2

SIM 08026-2
Brianna Peterson

vial # 14



#	Compound	Area	RT
1	ETHANOL	729	1.064
2	n-PROPANOL	1504	1.824

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 2:31:57 PM
Instrument 3
db-alc2

SIM 08026-3
Brianna Peterson

vial # 15



#	Compound	Area	RT
1	ETHANOL	725	1.064
2	n-PROPANOL	1496	1.823

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 2:35:04 PM
Instrument 3
db-alc2

SIM 08026-4
Brianna Peterson

vial # 16



#	Compound	Area	RT
1	ETHANOL	722	1.064
2	n-PROPANOL	1486	1.824

Totals:



ETHANOL 0.101 g/100ml



n-PROPANOL 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 2:38:11 PM
Instrument 3
db-alc2

SIM 08026-5
Brianna Peterson

vial # 17



#	Compound	Area	RT
1	ETHANOL	745	1.064
2	n-PROPANOL	1510	1.825

Totals:



ETHANOL 0.103 g/100ml



n-PROPANOL 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 2:41:18 PM
Instrument 3
db-alc2

0.10 con BP
Brianna Peterson

vial # 18



#	Compound	Area	RT
1	ETHANOL	743	1.064
2	n-PROPANOL	1567	1.824

Totals:



ETHANOL 0.099 g/100ml



n-PROPANOL 1.000 g/100ml

BP

WASHINGTON STATE TOXICOLOGY LABORATORY

C:\HPCHEM\2\METHODS\BLDALCO3.M
6/2/2008 2:44:26 PM
Instrument 3
db-alc2

Blank
Brianna Peterson

vial # 19



#	Compound	Area	RT
1	ETHANOL	0	0.000
2	n-PROPANOL	1607	1.825

Totals:



ETHANOL 0.000 g/100ml



n-PROPANOL 1.000 g/100ml

BP