



**QUALITY ASSURANCE PROCEDURE SOLUTION TEST REPORT**

**BATCH REPORT: 16017**

**CUSTOMER INFORMATION**

Washington State Patrol – Breath Test Program  
811 East Roanoke SEATTLE, WA 98102

**TESTING PROCEDURE USED:** TLD Technical Manual, Chapter 4.0 Certification of Simulator Solutions; Headspace-Gas Chromatography.

**TESTING ITEM INFORMATION**

TARGET VAPOR CONCENTRATION: 0.04 g/210L  
DATE PREPARED: 05/20/2016  
BATCH UNITS: g/100mL

IDENTITY: QAP Solution  
PREPARED BY: Christopher S. Johnston

	CSJ	AC	KH
1	0.050	0.050	0.050
2	0.050	0.050	0.050
3	0.050	0.050	0.050
4	0.050	0.050	0.049
5	0.050	0.049	0.049
C	0.100	0.102	0.099

**ETHANOL CONTROL INFORMATION**


LOT NUMBER: FN08051301 EXPIRATION: 10/2018 CONCENTRATION: 0.10 g/100mL

**RESULTS OF TESTING**

AVERAGE SOLUTION CONCENTRATION: 0.0498 g/100mL PRECISION CV (%): 0.83  
STANDARD DEVIATION: 0.00041 NUMBER OF TESTS: 15




EQUIVALENT VAPOR CONCENTRATION: **0.0405 g/210L**  
EXPANDED UNCERTAINTY: ± 0.0012 (k=2, 95.45% confidence interval)

**WASHINGTON STATE PATROL – TOXICOLOGY LABORATORY DIVISION**

  
\_\_\_\_\_  
Lisa Noble Forensic Scientist Supervisor

6/29/16  
DATE REPORT ISSUED

THIS TESTING WAS PERFORMED BY:

ANALYST	NAME	SIGNATURE	DATE TESTED
CSJ	Christopher S. Johnston		05/20/2016
AC	Amanda Chandler		05/24/2016
KH	Katie Harris		05/27/2016

Washington State Patrol - Toxicology Laboratory Division  
QAP Test Report Calculation Record

QAP Solution Batch #: 16017

Date Prepared: 5/20/2016

Analyst:	CSJ	AC	KH
Date Tested:	5/20/2016	5/24/2016	5/27/2016
Instrument:	HSGC #1	HSGC #1	HSGC #1
1	0.050	0.050	0.050
2	0.050	0.050	0.050
3	0.050	0.050	0.050
4	0.050	0.050	0.049
5	0.050	0.049	0.049
C	0.100	0.102	0.099

$CV^2_{COA}$	$CV^2_{QAP\ Solution}$	$CV^2_{Control}$	$CV^2_{Part\ Coef}$
0.0000084100	0.0000046082	0.0000772618	0.0001016326

Ethanol Control Lot #: FN08051301  
Control Uncertainty (%): 0.29

Average Solution Concentration: 0.0498 g/100mL  
Standard Deviation: 0.00041 g/100mL  
Precision CV (%): 0.83  
Equivalent Vapor Concentration: 0.0405 g/210L  
Combined Standard Uncertainty ( $\pm$ ): 0.0006 g/210L  
Expanded Uncertainty ( $\pm$ ): 0.0012 coverage factor (k) =2 (95.45% level of confidence)

Calculations performed by: Lisa Noble [Signature] 6/13/16  
Name Signature Date

Calculations verified by: Amanda M. Black [Signature] 6-27-16  
Name Signature Date

Method: Hard calculation

Tech. review performed by: Lisa Noble [Signature] 6/13/16  
Name Signature Date

20

**SIMULATOR SOLUTION DATA ENTRY REVIEW**

Reviewer/s: Amanda H. Black

Date: 6-27-14

Location: WSP-FLSB Seattle, WA

Solution Batch Number: 16017

	YES	NO	N/A
Analysis dates do not precede preparation date:	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Declarations signed and properly dated:	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Data entry corresponds to all chromatograms:	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
All signatures present on Test Report:	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Average solution concentration correct:	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Standard deviation correct:	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
CV (%) correct:	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Equivalent vapor concentration correct:	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
All chromatograms and sequences included in file:	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Ethanol control information present: (lot # present & used within expiration)	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
Complies with accuracy and precision requirements established by the State Toxicologist:	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

Comments:

Reviewer Signature:



Date:

6-27-14



**SOLUTION CERTIFICATE REVIEW**

Please check that the data on your chromatograms is the data entered into the Test Report, that the date to the right of your name is the date that you tested the solution, and then sign the Test Report.

Please initial and date below to affirm that you have:

- 1) Checked your data
- 2) Checked the date to the right of your name on the Test Report
- 3) Signed the Test Report

	Initials	Date
<b>Amanda Chandler</b>	AC	6/13/14
<b>Andrew Gingras</b>		
<b>Asa Louis</b>		
<b>Brittany Thomas</b>		
<b>Christie Mitchell-Mata</b>		
<b>Christopher Johnston</b>	CJ	6/13/14
<b>David Nguyen</b>		
<b>Dawn Sklerov</b>		
<b>Elizabeth Wehner</b>		
<b>Justin Knoy</b>		
<b>Katie Harris</b>	KH	6/13/16
<b>Lyndsey Lowe</b>		
<b>Naziha Nuwayhid</b>		
<b>Rebecca Flaherty</b>		

Batch # 16017 for 6/13/16

JAY INSLEE  
Governor



JOHN R. BATISTE  
Chief

STATE OF WASHINGTON  
WASHINGTON STATE PATROL  
WASHINGTON STATE TOXICOLOGY LABORATORY

2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2027 • (206) 262-6100 • FAX (206) 262-6145

**0.04 g/210 L QUALITY ASSURANCE PROCEDURE SOLUTION  
CERTIFICATION FOR LOT 16017**


I, Christopher S. Johnston, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and my responsibilities include the preparation and certification of alcohol solutions for use with evidential breath test instruments.

I possess the following qualifications: BS degree in Biochemistry.

The quality assurance procedure (QAP) solution, Lot Number 16017, was prepared in the Washington State Toxicology Laboratory on 5/20/2016. I tested this solution and it was found to conform to those standards established by the State Toxicologist for the certification of simulator solution. It shall not be used to perform a quality assurance procedure after 5/20/2017.

Seattle, WA

  
\_\_\_\_\_  
Christopher S. Johnston                      6/13/2016  
Forensic Scientist                              Date



JAY INSLEE  
Governor



JOHN R. BATISTE  
Chief

STATE OF WASHINGTON  
WASHINGTON STATE PATROL  
WASHINGTON STATE TOXICOLOGY LABORATORY

2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2027 • (206) 262-6100 • FAX (206) 262-6145

**0.04 g/210 L QUALITY ASSURANCE PROCEDURE SOLUTION  
CERTIFICATION FOR LOT 16017**

I, Amanda Chandler, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and my responsibilities include the preparation and certification of alcohol solutions for use with evidential breath test instruments.

I possess the following qualifications: MS degree in Forensic Toxicology.

The quality assurance procedure (QAP) solution, Lot Number 16017, was prepared in the Washington State Toxicology Laboratory on 5/20/2016. I tested this solution and it was found to conform to those standards established by the State Toxicologist for the certification of simulator solution. It shall not be used to perform a quality assurance procedure after 5/20/2017.

Seattle, WA

 4/13/16

Amanda Chandler  
Forensic Scientist

Date



JAY INSLEE  
Governor



JOHN R. BATISTE  
Chief

STATE OF WASHINGTON  
WASHINGTON STATE PATROL  
WASHINGTON STATE TOXICOLOGY LABORATORY

2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2027 • (206) 262-6100 • FAX (206) 262-6145

**0.04 g/210 L QUALITY ASSURANCE PROCEDURE SOLUTION  
CERTIFICATION FOR LOT 16017**

I, Katie Harris, do certify under penalty of perjury that:

I am employed by the Washington State Toxicology Laboratory, and my responsibilities include the preparation and certification of alcohol solutions for use with evidential breath test instruments.

I possess the following qualifications: BS in Biochemistry and MS degree in Forensic Science.

The quality assurance procedure (QAP) solution, Lot Number 16017, was prepared in the Washington State Toxicology Laboratory on 5/20/2016. I tested this solution and it was found to conform to those standards established by the State Toxicologist for the certification of simulator solution. It shall not be used to perform a quality assurance procedure after 5/20/2017.

Seattle, WA

Katie Harris      6/13/16

Katie Harris

Date

Forensic Scientist



FILE A COPY IN THE BATCH FILE FOR EACH SOLUTION LISTED ON THE WORKSHEET

Preparation Date: 5/20/16 Expiration Date: 5/20/17 Initials of Preparer: CJ

Lot # of 200-proof Ethanol used in preparation: 2EA0437

Date the 200-proof Ethanol bottle was opened: 4/7/16

After opening, each bottle of 200-proof Ethanol is approved for use for 6 months unless an extension is approved by the State Toxicologist. This timeframe applies to the 200-proof Ethanol only, not to simulator solutions which have a 1 year expiration.

Environmental conditions verified as acceptable:

Simulator Solution	Volume of Ethanol (mL)	Volume of Deionized Water (L)		Batch Number
QAP 0.04	11.2	18	<input checked="" type="checkbox"/>	<u>16017</u>
QAP 0.08	22.4	18	<input checked="" type="checkbox"/>	<u>16018</u>
<del>QAP 0.08</del> - QAP 0.10	<del>22.4</del> 28.1	18	<input checked="" type="checkbox"/>	<u>16019</u>
QAP 0.15	42.1	18	<input checked="" type="checkbox"/>	<u>16020</u>
QAP 0.20	56.1	18	<input type="checkbox"/>	
ESS	66.5	52	<input type="checkbox"/>	

Stir bar is rotating

Stirred for minimum 30 minutes; 2 hours for ESS

Spigot purged

Aliquot taken

Batch labeled, packaged and sealed

5/20/16  
Date

If different ethanol lot numbers are used in the preparation of solutions, record them and the corresponding solution batch numbers in the comments section.

Comments:

For operational needs I asked Chris to make two batches of 0.08 QAP.  
In 5/23/16

[Signature]  
Analyst Signature

5/20/16  
Date

*[Handwritten mark]*



Sequence Parameters:

Operator: Chris Johnston  
 Data File Naming: Prefix/Counter  
 Signal 1 Prefix: SIG1  
 Counter: 0001  
 Signal 2 Prefix: SIG2  
 Counter: 0001  
 Data Directory: C:\HPCHEM\1\DATA\  
 Data Subdirectory: 160520CJ  
 Part of Methods to run: According to Runtime Checklist  
 Barcode Reader: not used  
 Shutdown Cmd/Macro: none

Sequence Comment:

Ethanol Calibrator 1, E0416-01 - Exp. 10/01/2016  
 Ethanol Calibrator 2, E0416-02 - Exp. 10/01/2016  
 Ethanol Calibrator 3, E0416-03 - Exp. 10/01/2016  
 CTRL1 (0.04g/100mL), Lot # FN05011301 - Exp. 05/2018  
 CTRL2 (0.10g/100mL), Lot # FN08051301 - Exp. 10/2018  
 CTRL3 (0.20g/100mL), Lot # FN03211401 - Exp. 06/2019  
 Internal Standard Lot#P0316 - Exp. 06/29/2016

Calibration vials 1-9 filed with 16017.

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	SIMALC1	1	Sample		
2	Vial 2	0.079 CAL 1	SIMALC1	1	Calib		
3	Vial 3	0.158 CAL 2	SIMALC1	1	Calib		
4	Vial 4	0.316 CAL 3	SIMALC1	1	Calib		
5	Vial 5	NEG CTRL	SIMALC1	1	Ctrl Samp		
6	Vial 6	0.04 CTRL	SIMALC1	1	Ctrl Samp		
7	Vial 7	0.10 CTRL	SIMALC1	1	Ctrl Samp		
8	Vial 8	0.20 CTRL	SIMALC1	1	Ctrl Samp		
9	Vial 9	NEG CTRL	SIMALC1	1	Ctrl Samp		
10	Vial 10	16017-1	SIMALC1	1	Sample		
11	Vial 11	16017-2	SIMALC1	1	Sample		
12	Vial 12	16017-3	SIMALC1	1	Sample		
13	Vial 13	16017-4	SIMALC1	1	Sample		
14	Vial 14	16017-5	SIMALC1	1	Sample		
15	Vial 15	0.10 CTRL	SIMALC1	1	Ctrl Samp		
16	Vial 16	NEG CTRL	SIMALC1	1	Ctrl Samp		
17	Vial 17	16018-1	SIMALC1	1	Sample		
18	Vial 18	16018-2	SIMALC1	1	Sample		
19	Vial 19	16018-3	SIMALC1	1	Sample		
20	Vial 20	16018-4	SIMALC1	1	Sample		
21	Vial 21	16018-5	SIMALC1	1	Sample		
22	Vial 22	0.10 CTRL	SIMALC1	1	Ctrl Samp		
23	Vial 23	NEG CTRL	SIMALC1	1	Ctrl Samp		
24	Vial 24	16019-1	SIMALC1	1	Sample		
25	Vial 25	16019-2	SIMALC1	1	Sample		
26	Vial 26	16019-3	SIMALC1	1	Sample		

16017  
 Jn 6/13/16

W

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
27	Vial 27	16019-4	SIMALC1	1	Sample		
28	Vial 28	16019-5	SIMALC1	1	Sample		
29	Vial 29	0.10 CTRL	SIMALC1	1	Ctrl Samp		
30	Vial 30	NEG CTRL	SIMALC1	1	Ctrl Samp		
31	Vial 31	16020-1	SIMALC1	1	Sample		
32	Vial 32	16020-2	SIMALC1	1	Sample		
33	Vial 33	16020-3	SIMALC1	1	Sample		
34	Vial 34	16020-4	SIMALC1	1	Sample		
35	Vial 35	16020-5	SIMALC1	1	Sample		
36	Vial 36	0.10 CTRL	SIMALC1	1	Ctrl Samp		
37	Vial 37	NEG CTRL	SIMALC1	1	Ctrl Samp		

Calibration Part:

Line	Location	SampleName	Method	CalLev	Update	RF	Update	RT	Interval
2	Vial 2	0.079 CAL 1	SIMALC1	1	Replace		Replace		
3	Vial 3	0.158 CAL 2	SIMALC1	2	Replace		Replace		
4	Vial 4	0.316 CAL 3	SIMALC1	3	Replace		Replace		

Sequence Table (Back Injector):

No entries - empty table!

16017

*Jn 6/13/16*

*W*

=====  
Calibration Table  
=====

EtOH Calibration

Calib. Data Modified : 5/23/2016 2:21:09 PM

Calculate : Internal Standard  
Based on : Peak Area

Rel. Reference Window : 5.000 %  
Abs. Reference Window : 0.050 min  
Rel. Non-ref. Window : 5.000 %  
Abs. Non-ref. Window : 0.050 min  
Multiplier : 1.0000  
Dilution : 1.0000  
Sample Amount : 0.00000  
Use Multiplier & Dilution Factor with ISTDs  
Uncalibrated Peaks : not reported  
Partial Calibration : No recalibration if peaks missing

Curve Type : Linear  
Origin : Included  
Weight : Equal

Recalibration Settings:  
Average Response : No Update  
Average Retention Time: No Update

Calibration Report Options :  
Printout of recalibrations within a sequence:  
Normal Report after Recalibration

Sample ISTD Information:

ISTD #	ISTD Amount [g/100mL]	Name
5	1.20000e-2	n-Propanol

Signal 1: FID1 A,


RetTime [min]	Lvl Sig	Amount [g/100mL]	Area	Amt/Area	Ref	Grp Name
1.094	1 1	7.91100e-2	982.52124	8.05173e-5	5	Ethanol
		2 1.59090e-1	1948.51196	8.16469e-5		
		3 3.15200e-1	4099.33789	7.68905e-5		
1.757	1 1	1.20000e-2	2563.04150	4.68194e-6	I5	n-Propanol
		2 1.20000e-2	2551.61938	4.70290e-6		
		3 1.20000e-2	2641.86401	4.54225e-6		

16017  
5/23/16


=====  
Peak Sum Table  
=====

\*\*\*No Entries in table\*\*\*  
=====

=====  
Calibration Curves  
=====



Ethanol at exp. RT: 1.094  
FID1 A,  
Correlation: 0.99990  
Residual Std. Dev.: 0.01128  
Formula:  $y = mx + b$   
m: 5.90646e-2  
b: -6.30026e-3  
x: Amount Ratio  
y: Area Ratio




n-Propanol at exp. RT: 1.757  
FID1 A,  
Correlation: 1.00000  
Residual Std. Dev.: 0.00000  
Formula:  $y = mx + b$   
m: 1.00000  
b: 0.00000  
x: Amount Ratio  
y: Area Ratio

16017


*Jnv/13/16*

Washington State Patrol Toxicology Laboratory  
2203 Airport Way S Seattle, WA 98134


Inj. Date: 5/20/2016 2:23:25 PM      Sample Name: BLANK  
Instrument: HSGC#1      Operator: Chris Johnston  
Column: DB-ALC1      Location: Vial 1  
Method: C:\HPCHEM\1\METHODS\BLDALCO1.M  
Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	0	0.000
2	n-Propanol	0	0.000



Ethanol      0.000 g/100mL




n-Propanol      0.000 g/100mL

*fr*


W

Washington State Patrol Toxicology Laboratory  
2203 Airport Way S Seattle, WA 98134


Inj. Date: 5/20/2016 2:26:44 PM      Sample Name: 0.079 CAL 1  
Instrument: HSGC#1      Operator: Chris Johnston  
Column: DB-ALC1      Location: Vial 2  
Method: C:\HPCHEM\1\METHODS\BLDALCO1.M  
Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	983	1.092
2	n-Propanol	2563	1.754



Ethanol      0.079 g/100mL




n-Propanol      0.012 g/100mL

*Handwritten signature*


W

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134


Inj. Date: 5/20/2016 2:30:01 PM      Sample Name: 0.158 CAL 2  
 Instrument: HSGC#1      Operator: Chris Johnston  
 Column: DB-ALC1      Location: Vial 3  
 Method: C:\HPCHEM\1\METHODS\BLDALCO1.M  
 Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	1949	1.092
2	n-Propanol	2552	1.754



Ethanol      0.156 g/100mL



n-Propanol      0.012 g/100mL

*Handwritten signature*

*Handwritten 'W'*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/20/2016 2:33:18 PM

Sample Name: 0.316 CAL 3

Instrument: HSGC#1


Operator: Chris Johnston

Column: DB-ALC1


Location: Vial 4

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	4099	1.094
2	n-Propanol	2642	1.757



Ethanol 0.317 g/100mL



n-Propanol 0.012 g/100mL

*h*


*w*




Washington State Patrol Toxicology Laboratory  
2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/20/2016 2:36:31 PM  
Instrument: HSGC#1  
Column: DB-ALC1  
Method: C:\HPCHEM\1\METHODS\SIMALC1.M  
Sample Info: 16017


Sample Name: NEG CTRL  
Operator: Chris Johnston  
Location: Vial 5



#	Compound	Peak Area	RT (min)
1	Ethanol	0	0.000
2	n-Propanol	2634	1.756



Ethanol 0.000 g/100mL



n-Propanol 0.012 g/100mL


*Handwritten signature*

*Handwritten mark*


Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/20/2016 2:39:45 PM  
 Instrument: HSGC#1  
 Column: DB-ALC1  
 Method: C:\HPCHEM\1\METHODS\SIMALC1.M  
 Sample Info: 16017


Sample Name: 0.04 CTRL  
 Operator: Chris Johnston  
 Location: Vial 6



#	Compound	Peak Area	RT (min)
1	Ethanol	497	1.093
2	n-Propanol	2539	1.755



Ethanol 0.041 g/100mL



n-Propanol 0.012 g/100mL


*[Handwritten signature]*

*[Handwritten mark]*


Washington State Patrol Toxicology Laboratory  
2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/20/2016 2:42:58 PM  
Instrument: HSGC#1  
Column: DB-ALC1  
Method: C:\HPCHEM\1\METHODS\SIMALC1.M  
Sample Info: 16017


Sample Name: 0.10 CTRL  
Operator: Chris Johnston  
Location: Vial 7



#	Compound	Peak Area	RT (min)
1	Ethanol	1228	1.092
2	n-Propanol	2521	1.754



Ethanol 0.100 g/100mL



n-Propanol 0.012 g/100mL

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/20/2016 2:46:12 PM

Sample Name: 0.20 CTRL

Instrument: HSGC#1


Operator: Chris Johnston

Column: DB-ALC1


Location: Vial 8

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	2466	1.093
2	n-Propanol	2534	1.756



Ethanol 0.199 g/100mL



n-Propanol 0.012 g/100mL


*Handwritten signature*

*Handwritten mark*


Washington State Patrol Toxicology Laboratory  
2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/20/2016 2:49:25 PM  
Instrument: HSGC#1  
Column: DB-ALC1  
Method: C:\HPCHEM\1\METHODS\SIMALC1.M  
Sample Info: 16017


Sample Name: NEG CTRL  
Operator: Chris Johnston  
Location: Vial 9



#	Compound	Peak Area	RT (min)
1	Ethanol	0	0.000
2	n-Propanol	2615	1.757



Ethanol 0.000 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*Handwritten mark*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/20/2016 2:52:38 PM

Sample Name: 16017-1

Instrument: HSGC#1


Operator: Chris Johnston

Column: DB-ALC1


Location: Vial 10

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info:



#	Compound	Peak Area	RT (min)
1	Ethanol	609	1.090
2	n-Propanol	2529	1.753



Ethanol 0.050 g/100mL



n-Propanol 0.012 g/100mL

*fr*

*W*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/20/2016 2:55:52 PM

Sample Name: 16017-2

Instrument: HSGC#1


Operator: Chris Johnston

Column: DB-ALC1


Location: Vial 11

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info:



#	Compound	Peak Area	RT (min)
1	Ethanol	614	1.091
2	n-Propanol	2538	1.754



Ethanol 0.050 g/100mL



n-Propanol 0.012 g/100mL

*fr*

*w*

Inj. Date: 5/20/2016 2:59:05 PM

Sample Name: 16017-3

Instrument: HSGC#1


Operator: Chris Johnston

Column: DB-ALC1


Location: Vial 12

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info:



#	Compound	Peak Area	RT (min)
1	Ethanol	608	1.091
2	n-Propanol	2517	1.754



Ethanol 0.050 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten initials*

*Handwritten mark*



Washington State Patrol Toxicology Laboratory  
2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/20/2016 3:02:18 PM

Sample Name: 16017-4

Instrument: HSGC#1


Operator: Chris Johnston

Column: DB-ALC1


Location: Vial 13

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info:



#	Compound	Peak Area	RT (min)
1	Ethanol	609	1.088
2	n-Propanol	2530	1.752



Ethanol 0.050 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

W

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/20/2016 3:05:32 PM

Sample Name: 16017-5

Instrument: HSGC#1


Operator: Chris Johnston

Column: DB-ALC1


Location: Vial 14

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info:



#	Compound	Peak Area	RT (min)
1	Ethanol	602	1.088
2	n-Propanol	2498	1.751



Ethanol 0.050 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*Handwritten mark*

Inj. Date: 5/20/2016 3:08:45 PM

Sample Name: 0.10 CTRL

Instrument: HSGC#1


Operator: Chris Johnston

Column: DB-ALC1


Location: Vial 15

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	1226	1.092
2	n-Propanol	2532	1.754



Ethanol 0.100 g/100mL



n-Propanol 0.012 g/100mL

*fr*

*W*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/20/2016 3:11:58 PM

Sample Name: NEG CTRL

Instrument: HSGC#1


Operator: Chris Johnston

Column: DB-ALC1


Location: Vial 16

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	0	0.000
2	n-Propanol	2535	1.755



Ethanol 0.000 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*Handwritten mark*

Sequence Parameters:

Operator: Amanda Chandler  
 Data File Naming: Prefix/Counter  
 Signal 1 Prefix: SIG1  
 Counter: 0001  
 Signal 2 Prefix: SIG2  
 Counter: 0001  
 Data Directory: C:\HPCHEM\1\DATA\  
 Data Subdirectory: 160524AC  
 Part of Methods to run: According to Runtime Checklist  
 Barcode Reader: not used  
 Shutdown Cmd/Macro: none

Sequence Comment:

Ethanol Calibrator 1, E0416-01 - Exp. 10/01/2016  
 Ethanol Calibrator 2, E0416-02 - Exp. 10/01/2016  
 Ethanol Calibrator 3, E0416-03 - Exp. 10/01/2016  
 CTRL1 (0.04g/100mL), Lot # FN05011301 - Exp. 05/2018  
 CTRL2 (0.10g/100mL), Lot # FN08051301 - Exp. 10/2018  
 CTRL3 (0.20g/100mL), Lot # FN03211401 - Exp. 06/2019  
 Internal Standard Lot#P0316 - Exp. 06/29/2016

Calibration vials 1-9 filed with 16017.

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	SIMALC1	1	Sample		
2	Vial 2	0.079 CAL 1	SIMALC1	1	Calib		
3	Vial 3	0.158 CAL 2	SIMALC1	1	Calib		
4	Vial 4	0.316 CAL 3	SIMALC1	1	Calib		
5	Vial 5	NEG CTRL	SIMALC1	1	Ctrl Samp		
6	Vial 6	0.04 CTRL	SIMALC1	1	Ctrl Samp		
7	Vial 7	0.10 CTRL	SIMALC1	1	Ctrl Samp		
8	Vial 8	0.20 CTRL	SIMALC1	1	Ctrl Samp		
9	Vial 9	NEG CTRL	SIMALC1	1	Ctrl Samp		
10	Vial 10	16017-1	SIMALC1	1	Sample		
11	Vial 11	16017-2	SIMALC1	1	Sample		
12	Vial 12	16017-3	SIMALC1	1	Sample		
13	Vial 13	16017-4	SIMALC1	1	Sample		
14	Vial 14	16017-5	SIMALC1	1	Sample		
15	Vial 15	0.10 CTRL	SIMALC1	1	Ctrl Samp		
16	Vial 16	NEG CTRL	SIMALC1	1	Ctrl Samp		
17	Vial 17	16018-1	SIMALC1	1	Sample		
18	Vial 18	16018-2	SIMALC1	1	Sample		
19	Vial 19	16018-3	SIMALC1	1	Sample		
20	Vial 20	16018-4	SIMALC1	1	Sample		
21	Vial 21	16018-5	SIMALC1	1	Sample		
22	Vial 22	0.10 CTRL	SIMALC1	1	Ctrl Samp		
23	Vial 23	NEG CTRL	SIMALC1	1	Ctrl Samp		
24	Vial 24	16019-1	SIMALC1	1	Sample		
25	Vial 25	16019-2	SIMALC1	1	Sample		
26	Vial 26	16019-3	SIMALC1	1	Sample		

16017  
 Feb/13/16

R

Sequence: C:\HPCHEM\1\SEQUENCE\ACQAP1.S

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
27	Vial 27	16019-4	SIMALC1	1	Sample		
28	Vial 28	16019-5	SIMALC1	1	Sample		
29	Vial 29	0.10 CTRL	SIMALC1	1	Ctrl Samp		
30	Vial 30	NEG CTRL	SIMALC1	1	Ctrl Samp		
31	Vial 31	16020-1	SIMALC1	1	Sample		
32	Vial 32	16020-2	SIMALC1	1	Sample		
33	Vial 33	16020-3	SIMALC1	1	Sample		
34	Vial 34	16020-4	SIMALC1	1	Sample		
35	Vial 35	16020-5	SIMALC1	1	Sample		
36	Vial 36	0.10 CTRL	SIMALC1	1	Ctrl Samp		
37	Vial 37	NEG CTRL	SIMALC1	1	Ctrl Samp		

Calibration Part:

Line	Location	SampleName	Method	CalLev	Update	RF	Update	RT	Interval
2	Vial 2	0.079 CAL 1	SIMALC1	1	Replace		Replace		
3	Vial 3	0.158 CAL 2	SIMALC1	2	Replace		Replace		
4	Vial 4	0.316 CAL 3	SIMALC1	3	Replace		Replace		

Sequence Table (Back Injector):

No entries - empty table!

16017  
In 01/13/16

AZ

=====  
Calibration Table  
=====

Calib. Data Modified : Tuesday, May 24, 2016 12:46:25 PM  
Calculate : Internal Standard  
Based on : Peak Area  
Rel. Reference Window : 5.000 %  
Abs. Reference Window : 0.050 min  
Rel. Non-ref. Window : 5.000 %  
Abs. Non-ref. Window : 0.050 min  
Multiplier : 1.0000  
Dilution : 1.0000  
Sample Amount : 0.00000  
Use Multiplier & Dilution Factor with ISTDs  
Uncalibrated Peaks : not reported  
Partial Calibration : No recalibration if peaks missing  
Curve Type : Linear  
Origin : Included  
Weight : Equal

Recalibration Settings:  
Average Response : No Update  
Average Retention Time: No Update

Calibration Report Options :  
Printout of recalibrations within a sequence:  
Normal Report after Recalibration

Sample ISTD Information:

ISTD #	ISTD Amount [g/100mL]	Name
1	1.20000e-2	n-Propanol

Signal 1: FID1 A,

RetTime [min]	Lvl Sig	Amount [g/100mL]	Area	Amt/Area	Ref	Grp Name
1.094	1 1	7.91100e-2	985.80859	8.02488e-5	1	Ethanol
	2	1.59090e-1	2057.29932	7.73295e-5		
	3	3.15200e-1	4016.58643	7.84746e-5		
1.757	1 1	1.20000e-2	2536.36426	4.73118e-6	I1	n-Propanol
	2	1.20000e-2	2642.72852	4.54076e-6		
	3	1.20000e-2	2586.19165	4.64003e-6		

16017


*Ln6/13/16*

=====  
Peak Sum Table  
=====


\*\*\*No Entries in table\*\*\*  
=====

*AZ*

=====  
Calibration Curves  
=====



Ethanol at exp. RT: 1.094  
FID1 A,  
Correlation: 0.99999  
Residual Std. Dev.: 0.00315  
Formula:  $y = mx + b$   
m: 5.91178e-2  
b: -1.51989e-3  
x: Amount Ratio  
y: Area Ratio



n-Propanol at exp. RT: 1.757  
FID1 A,  
Correlation: 1.00000  
Residual Std. Dev.: 0.00000  
Formula:  $y = mx + b$   
m: 1.00000  
b: 0.00000  
x: Amount Ratio  
y: Area Ratio

16017  
*2/13/16*


*AR*




Washington State Patrol Toxicology Laboratory  
2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/24/2016 12:34:20 PM  
Instrument: HSGC#1  
Column: DB-ALC1  
Method: C:\HPCHEM\1\METHODS\SIMALC1.M  
Sample Info: 16017


Sample Name: BLANK  
Operator: Amanda Chandler  
Location: Vial 1



#	Compound	Peak Area	RT (min)
1	Ethanol	0	0.000
2	n-Propanol	0	0.000



Ethanol 0.000 g/100mL



n-Propanol 0.000 g/100mL

*Handwritten signature*

*Handwritten initials 'AR'*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/24/2016 12:37:38 PM

Sample Name: 0.079 CAL 1

Instrument: HSGC#1


Operator: Amanda Chandler

Column: DB-ALC1


Location: Vial 2

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	986	1.091
2	n-Propanol	2536	1.754



Ethanol 0.079 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*Handwritten signature*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/24/2016 12:40:55 PM

Sample Name: 0.158 CAL 2

Instrument: HSGC#1


Operator: Amanda Chandler

Column: DB-ALC1


Location: Vial 3

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	2057	1.094
2	n-Propanol	2643	1.757



Ethanol 0.158 g/100mL



n-Propanol 0.012 g/100mL


*Handwritten signature*

*Handwritten signature*


Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/24/2016 12:44:12 PM  
 Instrument: HSGC#1  
 Column: DB-ALC1  
 Method: C:\HPCHEM\1\METHODS\SIMALC1.M  
 Sample Info: 16017


Sample Name: 0.316 CAL 3  
 Operator: Amanda Chandler  
 Location: Vial 4



#	Compound	Peak Area	RT (min)
1	Ethanol	4017	1.094
2	n-Propanol	2586	1.757



Ethanol 0.316 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*Handwritten signature*

Washington State Patrol Toxicology Laboratory  
2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/24/2016 12:47:25 PM

Sample Name: NEG CTRL

Instrument: HSGC#1


Operator: Amanda Chandler

Column: DB-ALC1


Location: Vial 5

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	0	0.000
2	n-Propanol	2576	1.757



Ethanol 0.000 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*Handwritten signature*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/24/2016 12:50:39 PM

Sample Name: 0.04 CTRL

Instrument: HSGC#1


Operator: Amanda Chandler

Column: DB-ALC1


Location: Vial 6

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	495	1.091
2	n-Propanol	2545	1.755



Ethanol 0.040 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*Handwritten initials*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/24/2016 12:53:52 PM

Sample Name: 0.10 CTRL

Instrument: HSGC#1


Operator: Amanda Chandler

Column: DB-ALC1


Location: Vial 7

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	1430	1.093
2	n-Propanol	2863	1.757



Ethanol 0.102 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*Handwritten initials AR*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/24/2016 12:57:05 PM

Sample Name: 0.20 CTRL

Instrument: HSGC#1


Operator: Amanda Chandler

Column: DB-ALC1


Location: Vial 8

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	2716	1.094
2	n-Propanol	2789	1.757



Ethanol 0.198 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*Handwritten initials*



Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/24/2016 1:00:18 PM

Sample Name: NEG CTRL

Instrument: HSGC#1


Operator: Amanda Chandler

Column: DB-ALC1


Location: Vial 9

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	0	0.000
2	n-Propanol	2778	1.756



Ethanol 0.000 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*Handwritten initials 'AC'*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/24/2016 1:03:32 PM

Sample Name: 16017-1

Instrument: HSGC#1


Operator: Amanda Chandler

Column: DB-ALC1


Location: Vial 10

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info:



#	Compound	Peak Area	RT (min)
1	Ethanol	621	1.091
2	n-Propanol	2556	1.754



Ethanol 0.050 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*Handwritten signature*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/24/2016 1:06:45 PM

Sample Name: 16017-2

Instrument: HSGC#1


Operator: Amanda Chandler

Column: DB-ALC1


Location: Vial 11

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info:



#	Compound	Peak Area	RT (min)
1	Ethanol	631	1.093
2	n-Propanol	2586	1.756



Ethanol 0.050 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*Handwritten initials AR*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/24/2016 1:09:58 PM

Sample Name: 16017-3

Instrument: HSGC#1


Operator: Amanda Chandler

Column: DB-ALC1


Location: Vial 12

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info:



#	Compound	Peak Area	RT (min)
1	Ethanol	622	1.092
2	n-Propanol	2560	1.756



Ethanol 0.050 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten initials*

*Handwritten initials AR*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/24/2016 1:13:11 PM

Sample Name: 16017-4

Instrument: HSGC#1


Operator: Amanda Chandler

Column: DB-ALC1


Location: Vial 13

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info:



#	Compound	Peak Area	RT (min)
1	Ethanol	615	1.090
2	n-Propanol	2528	1.753



Ethanol 0.050 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*Handwritten initials AR*

Washington State Patrol Toxicology Laboratory  
2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/24/2016 1:16:24 PM

Sample Name: 16017-5

Instrument: HSGC#1


Operator: Amanda Chandler

Column: DB-ALC1


Location: Vial 14

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info:



#	Compound	Peak Area	RT (min)
1	Ethanol	623	1.091
2	n-Propanol	2578	1.754



Ethanol 0.049 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*Handwritten signature*


Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/24/2016 1:19:38 PM  
 Instrument: HSGC#1


Sample Name: 0.10 CTRL  
 Operator: Amanda Chandler  
 Location: Vial 15

Column: DB-ALC1  
 Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	1321	1.093
2	n-Propanol	2651	1.756



Ethanol 0.102 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten initials*

*Handwritten initials*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/24/2016 1:22:51 PM

Sample Name: NEG CTRL

Instrument: HSGC#1


Operator: Amanda Chandler

Column: DB-ALC1


Location: Vial 16

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	0	0.000
2	n-Propanol	2583	1.756



Ethanol 0.000 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*Handwritten initials 'AC'*



Sequence Parameters:

Operator: Katie Harris  
 Data File Naming: Prefix/Counter  
 Signal 1 Prefix: SIG1  
 Counter: 0001  
 Signal 2 Prefix: SIG2  
 Counter: 0001  
 Data Directory: C:\HPCHEM\1\DATA\  
 Data Subdirectory: 160527KH  
 Part of Methods to run: According to Runtime Checklist  
 Barcode Reader: not used  
 Shutdown Cmd/Macro: none

Sequence Comment:

Ethanol Calibrator 1, E0416-01 - Exp. 10/01/2016  
 Ethanol Calibrator 2, E0416-02 - Exp. 10/01/2016  
 Ethanol Calibrator 3, E0416-03 - Exp. 10/01/2016  
 CTRL1 (0.04g/100mL), Lot # FN05011301 - Exp. 05/2018  
 CTRL2 (0.10g/100mL), Lot # FN08051301 - Exp. 10/2018  
 CTRL3 (0.20g/100mL), Lot # FN03211401 - Exp. 06/2019  
 Internal Standard Lot#P0316 - Exp. 06/29/2016

Calibration vials 1-9 filed with 16017.

Sequence Table (Front Injector):

Method and Injection Info Part:

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
1	Vial 1	BLANK	SIMALC1	1	Sample		
2	Vial 2	0.079 CAL 1	SIMALC1	1	Calib		
3	Vial 3	0.158 CAL 2	SIMALC1	1	Calib		
4	Vial 4	0.316 CAL 3	SIMALC1	1	Calib		
5	Vial 5	NEG CTRL	SIMALC1	1	Ctrl Samp		
6	Vial 6	0.04 CTRL	SIMALC1	1	Ctrl Samp		
7	Vial 7	0.10 CTRL	SIMALC1	1	Ctrl Samp		
8	Vial 8	0.20 CTRL	SIMALC1	1	Ctrl Samp		
9	Vial 9	NEG CTRL	SIMALC1	1	Ctrl Samp		
10	Vial 10	16017-1	SIMALC1	1	Sample		
11	Vial 11	16017-2	SIMALC1	1	Sample		
12	Vial 12	16017-3	SIMALC1	1	Sample		
13	Vial 13	16017-4	SIMALC1	1	Sample		
14	Vial 14	16017-5	SIMALC1	1	Sample		
15	Vial 15	0.10 CTRL	SIMALC1	1	Ctrl Samp		
16	Vial 16	NEG CTRL	SIMALC1	1	Ctrl Samp		
17	Vial 17	16018-1	SIMALC1	1	Sample		
18	Vial 18	16018-2	SIMALC1	1	Sample		
19	Vial 19	16018-3	SIMALC1	1	Sample		
20	Vial 20	16018-4	SIMALC1	1	Sample		
21	Vial 21	16018-5	SIMALC1	1	Sample		
22	Vial 22	0.10 CTRL	SIMALC1	1	Ctrl Samp		
23	Vial 23	NEG CTRL	SIMALC1	1	Ctrl Samp		
24	Vial 24	16019-1	SIMALC1	1	Sample		
25	Vial 25	16019-2	SIMALC1	1	Sample		
26	Vial 26	16019-3	SIMALC1	1	Sample		

16017

*fn0321*

*KH*

Line	Location	SampleName	Method	Inj	SampleType	InjVolume	DataFile
27	Vial 27	16019-4	SIMALC1	1	Sample		
28	Vial 28	16019-5	SIMALC1	1	Sample		
29	Vial 29	0.10 CTRL	SIMALC1	1	Ctrl Samp		
30	Vial 30	NEG CTRL	SIMALC1	1	Ctrl Samp		
31	Vial 31	16020-1	SIMALC1	1	Sample		
32	Vial 32	16020-2	SIMALC1	1	Sample		
33	Vial 33	16020-3	SIMALC1	1	Sample		
34	Vial 34	16020-4	SIMALC1	1	Sample		
35	Vial 35	16020-5	SIMALC1	1	Sample		
36	Vial 36	0.10 CTRL	SIMALC1	1	Ctrl Samp		
37	Vial 37	NEG CTRL	SIMALC1	1	Ctrl Samp		

Calibration Part:

Line	Location	SampleName	Method	CalLev	Update	RF	Update	RT	Interval
2	Vial 2	0.079 CAL 1	SIMALC1	1	Replace		Replace		
3	Vial 3	0.158 CAL 2	SIMALC1	2	Replace		Replace		
4	Vial 4	0.316 CAL 3	SIMALC1	3	Replace		Replace		

Sequence Table (Back Injector):

No entries - empty table!

16017

*Lu 6/12/16*

*KN*

=====  
Calibration Table  
=====

Calib. Data Modified : Friday, May 27, 2016 11:37:58 AM

Calculate : Internal Standard  
Based on : Peak Area

Rel. Reference Window : 5.000 %  
Abs. Reference Window : 0.050 min  
Rel. Non-ref. Window : 5.000 %  
Abs. Non-ref. Window : 0.050 min  
Multiplier : 1.0000  
Dilution : 1.0000  
Sample Amount : 0.00000  
Use Multiplier & Dilution Factor with ISTDs  
Uncalibrated Peaks : not reported  
Partial Calibration : No recalibration if peaks missing

Curve Type : Linear  
Origin : Included  
Weight : Equal

Recalibration Settings:  
Average Response : No Update  
Average Retention Time: No Update

Calibration Report Options :  
Printout of recalibrations within a sequence:  
Normal Report after Recalibration

Sample ISTD Information:

ISTD #	ISTD Amount [g/100mL]	Name
1	1.20000e-2	n-Propanol

Signal 1: FID1 A,

RetTime [min]	Lvl Sig	Amount [g/100mL]	Area	Amt/Area	Ref	Grp Name
1.094	1 1	7.91100e-2	1019.48761	7.75978e-5	1	Ethanol
		1.59090e-1	2183.18677	7.28705e-5		
		3.15200e-1	4504.18311	6.99794e-5		
1.758	1 1	1.20000e-2	2596.87256	4.62094e-6	I1	n-Propanol
		1.20000e-2	2775.89282	4.32293e-6		
		1.20000e-2	2819.96045	4.25538e-6		

16017


*Signature*

=====  
Peak Sum Table  
=====


\*\*\*No Entries in table\*\*\*  
=====

KH

=====  
Calibration Curves  
=====



Ethanol at exp. RT: 1.094  
FID1 A,  
Correlation: 0.99991  
Residual Std. Dev.: 0.01142  
Formula:  $y = mx + b$   
m: 6.08274e-2  
b: -7.21078e-3  
x: Amount Ratio  
y: Area Ratio



n-Propanol at exp. RT: 1.758  
FID1 A,  
Correlation: 1.00000  
Residual Std. Dev.: 0.00000  
Formula:  $y = mx + b$   
m: 1.00000  
b: 0.00000  
x: Amount Ratio  
y: Area Ratio

16017  
Inhibitor

KH

Washington State Patrol Toxicology Laboratory  
2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/27/2016 11:25:52 AM

Sample Name: BLANK

Instrument: HSGC#1


Operator: Katie Harris

Column: DB-ALC1


Location: Vial 1

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	0	0.000
2	n-Propanol	0	0.000



Ethanol 0.000 g/100mL



n-Propanol 0.000 g/100mL

*[Handwritten signature]*

*KH*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/27/2016 11:29:09 AM

Sample Name: 0.079 CAL 1

Instrument: HSGC#1


Operator: Katie Harris

Column: DB-ALC1


Location: Vial 2

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	1019	1.094
2	n-Propanol	2597	1.757



Ethanol 0.079 g/100mL



n-Propanol 0.012 g/100mL


*Handwritten signature*

KH


Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/27/2016 11:32:28 AM  
 Instrument: HSGC#1  
 Column: DB-ALC1  
 Method: C:\HPCHEM\1\METHODS\SIMALC1.M  
 Sample Info: 16017


Sample Name: 0.158 CAL 2  
 Operator: Katie Harris  
 Location: Vial 3



#	Compound	Peak Area	RT (min)
1	Ethanol	2183	1.094
2	n-Propanol	2776	1.757



Ethanol 0.157 g/100mL




n-Propanol 0.012 g/100mL

*Handwritten signature*


*KH*

Washington State Patrol Toxicology Laboratory  
2203 Airport Way S Seattle, WA 98134


Inj. Date: 5/27/2016 11:35:45 AM      Sample Name: 0.316 CAL 3  
Instrument: HSGC#1      Operator: Katie Harris  
Column: DB-ALC1      Location: Vial 4  
Method: C:\HPCHEM\1\METHODS\SIMALC1.M  
Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	4504	1.094
2	n-Propanol	2820	1.758



Ethanol      0.317 g/100mL



n-Propanol      0.012 g/100mL


*fr*

*KH*




Washington State Patrol Toxicology Laboratory  
2203 Airport Way S Seattle, WA 98134


Inj. Date: 5/27/2016 11:38:58 AM      Sample Name: NEG CTRL  
Instrument: HSGC#1      Operator: Katie Harris  
Column: DB-ALC1      Location: Vial 5  
Method: C:\HPCHEM\1\METHODS\SIMALC1.M  
Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	0	0.000
2	n-Propanol	2526	1.755



Ethanol      0.000 g/100mL



n-Propanol      0.012 g/100mL

*Handwritten signature*

*KH*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/27/2016 11:42:11 AM

Sample Name: 0.04 CTRL

Instrument: HSGC#1


Operator: Katie Harris

Column: DB-ALC1


Location: Vial 6

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	550	1.094
2	n-Propanol	2745	1.758



Ethanol 0.041 g/100mL




n-Propanol 0.012 g/100mL

*Handwritten signature*


*KH*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134


Inj. Date: 5/27/2016 11:45:25 AM      Sample Name: 0.10 CTRL  
 Instrument: HSGC#1      Operator: Katie Harris  
 Column: DB-ALC1      Location: Vial 7  
 Method: C:\HPCHEM\1\METHODS\SIMALC1.M  
 Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	1243	1.093
2	n-Propanol	2524	1.756



Ethanol      0.099 g/100mL



n-Propanol      0.012 g/100mL

*sh*

*KH*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/27/2016 11:48:37 AM

Sample Name: 0.20 CTRL

Instrument: HSGC#1


Operator: Katie Harris

Column: DB-ALC1


Location: Vial 8

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	2778	1.094
2	n-Propanol	2785	1.758



Ethanol 0.198 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*KH*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/27/2016 11:51:52 AM

Sample Name: NEG CTRL

Instrument: HSGC#1


Operator: Katie Harris

Column: DB-ALC1


Location: Vial 9

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	0	0.000
2	n-Propanol	2557	1.757



Ethanol 0.000 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*Handwritten initials*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/27/2016 11:55:04 AM

Sample Name: 16017-1

Instrument: HSGC#1


Operator: Katie Harris

Column: DB-ALC1


Location: Vial 10

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info:



#	Compound	Peak Area	RT (min)
1	Ethanol	680	1.094
2	n-Propanol	2744	1.758



Ethanol 0.050 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*KH*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/27/2016 11:58:17 AM

Sample Name: 16017-2

Instrument: HSGC#1


Operator: Katie Harris

Column: DB-ALC1


Location: Vial 11

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info:



#	Compound	Peak Area	RT (min)
1	Ethanol	609	1.092
2	n-Propanol	2495	1.755



Ethanol 0.050 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*


*KH*

Washington State Patrol Toxicology Laboratory  
2203 Airport Way S Seattle, WA 98134


Inj. Date: 5/27/2016 12:01:31 PM  
Instrument: HSGC#1  
Column: DB-ALC1  
Method: C:\HPCHEM\1\METHODS\SIMALC1.M

Sample Name: 16017-3  
Operator: Katie Harris  
Location: Vial 12


Sample Info:



#	Compound	Peak Area	RT (min)
1	Ethanol	611	1.091
2	n-Propanol	2501	1.754



Ethanol 0.050 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*KH*



Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/27/2016 12:04:43 PM

Sample Name: 16017-4

Instrument: HSGC#1


Operator: Katie Harris

Column: DB-ALC1


Location: Vial 13

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info:



#	Compound	Peak Area	RT (min)
1	Ethanol	615	1.090
2	n-Propanol	2532	1.754



Ethanol 0.049 g/100mL



n-Propanol 0.012 g/100mL

*fr*

KH

Washington State Patrol Toxicology Laboratory  
2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/27/2016 12:07:56 PM

Sample Name: 16017-5

Instrument: HSGC#1


Operator: Katie Harris

Column: DB-ALC1


Location: Vial 14

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info:



#	Compound	Peak Area	RT (min)
1	Ethanol	617	1.093
2	n-Propanol	2540	1.757



Ethanol 0.049 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*KH*

Washington State Patrol Toxicology Laboratory  
 2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/27/2016 12:11:09 PM

Sample Name: 0.10 CTRL

Instrument: HSGC#1


Operator: Katie Harris

Column: DB-ALC1


Location: Vial 15

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	1276	1.092
2	n-Propanol	2586	1.755



Ethanol 0.099 g/100mL



n-Propanol 0.012 g/100mL

*fr*

Washington State Patrol Toxicology Laboratory  
2203 Airport Way S Seattle, WA 98134

Inj. Date: 5/27/2016 12:14:22 PM

Sample Name: NEG CTRL

Instrument: HSGC#1


Operator: Katie Harris

Column: DB-ALC1


Location: Vial 16

Method: C:\HPCHEM\1\METHODS\SIMALC1.M


Sample Info: 16017



#	Compound	Peak Area	RT (min)
1	Ethanol	0	0.000
2	n-Propanol	2543	1.756



Ethanol 0.000 g/100mL



n-Propanol 0.012 g/100mL

*Handwritten signature*

*Handwritten initials*