

**WASHINGTON STATE PATROL
BREATH TEST SECTION
REPAIRS AND/OR ADJUSTMENTS ON DATAMASTER**

Instrument Location WSP Bellingham Serial # 949103

Complaint (s) Reported:

Breath Test Discover Report Database included wrong year.

Technician Evaluation and Action Taken:

During a recent database search it was discovered that the data that had been downloaded from the timeframe of 12/07/2010 - 12/12/2010 was placed in the year 2011. These dates should have been uploaded for the year 2010. As a result the tests had been missing for this time period when a query was run. Any breath tests run during this time period on this instrument will also reflect the incorrect year on the breath test document.

The database has been corrected to reflect the correct year include the test data from this time period. Any subsequent query will include any tests that had been previously missing.

Complete Breath Test

Technician Jim van Diest

4-19-11

Technician [Signature]

Burlinson WA

Field Lab Factory	WORK BEGAN		LOCATION OF WORK	SERVICE TIME
	DATE	TIME		
Lab	4/19/2011	11:15	Seattle Roanoke Office	-